

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX080519\
 Data File : VX011328.D
 Acq On : 05 Aug 2019 16:03
 Operator : JC/SP
 Sample : K4129-21
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 2S-E1-(30)

Quant Time: Aug 06 03:40:04 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 02 01:55:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	188996	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	273464	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	247393	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	109473	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	91559	48.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.22%	
35) Dibromofluoromethane	5.49	113	82272	46.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.16%	
50) Toluene-d8	8.71	98	324493	48.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.30%	
62) 4-Bromofluorobenzene	11.13	95	101762	42.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.58%	
Target Compounds						
8) Diethyl Ether	2.18	74	1987	1.736	ug/l	92
16) Acetone	2.43	43	10510	11.601	ug/l	98
20) Methylene Chloride	2.85	84	4667	2.352	ug/l	89
43) Isopropyl Acetate	6.45	43	8288	2.165	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX080519\
Data File : VX011328.D
Acq On : 05 Aug 2019 16:03
Operator : JC/SP
Sample : K4129-21
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
2S-E1-(30)

Quant Time: Aug 06 03:40:04 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 02 01:55:00 2019
Response via : Initial Calibration

