

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX080519\
 Data File : VX011330.D
 Acq On : 05 Aug 2019 16:50
 Operator : JC/SP
 Sample : K4130-27
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 1S-F1-(30)

Quant Time: Aug 06 03:44:38 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 02 01:55:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	191613	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	277012	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	251629	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	109483	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	93115	48.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.50%	
35) Dibromofluoromethane	5.49	113	82876	45.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.64%	
50) Toluene-d8	8.71	98	330075	48.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.70%	
62) 4-Bromofluorobenzene	11.13	95	103871	42.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.24%	
Target Compounds						
8) Diethyl Ether	2.18	74	2032	1.751	ug/l	93
16) Acetone	2.43	43	9837	10.710	ug/l	89
20) Methylene Chloride	2.85	84	4327	2.151	ug/l	92
43) Isopropyl Acetate	6.45	43	8793	2.267	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX080519\
Data File : VX011330.D
Acq On : 05 Aug 2019 16:50
Operator : JC/SP
Sample : K4130-27
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
1S-F1-(30)

Quant Time: Aug 06 03:44:38 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 02 01:55:00 2019
Response via : Initial Calibration

