

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX080520\
 Data File : VX017795.D
 Acq On : 05 Aug 2020 14:09
 Operator : JC/SP
 Sample : L3570-01RE
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-01-200804RE

Quant Time: Aug 06 03:31:56 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 24 05:48:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	568586	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	877514	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	822820	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	394756	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	386845	52.97	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.94%	
35) Dibromofluoromethane	5.47	113	294462	50.17	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.34%	
50) Toluene-d8	8.70	98	1059550	50.05	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.10%	
62) 4-Bromofluorobenzene	11.12	95	384158	45.34	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.68%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX080520\
Data File : VX017795.D
Acq On : 05 Aug 2020 14:09
Operator : JC/SP
Sample : L3570-01RE
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-01-200804RE

Quant Time: Aug 06 03:31:56 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 24 05:48:09 2020
Response via : Initial Calibration

