

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080521\
 Data File : VX023547.D
 Acq On : 04 Aug 2021 17:59
 Operator : JC/MD
 Sample : VX0805WBS01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0805WBS01

Quant Time: Aug 05 06:08:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	200800	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	326433	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	300596	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	135225	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	138626	55.344	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	110.680%
35) Dibromofluoromethane	5.397	113	110675	54.226	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	108.460%
50) Toluene-d8	8.652	98	403359	53.791	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	107.580%
62) 4-Bromofluorobenzene	11.085	95	146671	57.521	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	115.040%
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.172	85	40309	21.820	ug/l	91
3) Chloromethane	1.294	50	45814	21.289	ug/l	100
4) Vinyl Chloride	1.373	62	46817	21.492	ug/l	99
5) Bromomethane	1.599	94	33592	25.974	ug/l	98
6) Chloroethane	1.684	64	30436	22.766	ug/l	99
7) Trichlorofluoromethane	1.885	101	75184	21.422	ug/l	99
8) Diethyl Ether	2.141	74	27923	21.556	ug/l	92
9) 1,1,2-Trichlorotrifluo...	2.330	101	43522	21.889	ug/l	96
10) Methyl Iodide	2.452	142	52265	20.980	ug/l	97
11) Tert butyl alcohol	2.977	59	63572	111.101	ug/l	97
12) 1,1-Dichloroethene	2.318	96	41568	21.068	ug/l	95
13) Acrolein	2.239	56	24073	83.670	ug/l	95
14) Allyl chloride	2.666	41	75125	20.943	ug/l	93
15) Acrylonitrile	3.068	53	146556	113.424	ug/l	98
16) Acetone	2.385	43	132932	108.188	ug/l	95
17) Carbon Disulfide	2.513	76	88526	18.196	ug/l	100
18) Methyl Acetate	2.708	43	67958	23.266	ug/l	92
19) Methyl tert-butyl Ether	3.123	73	150158	21.888	ug/l	98
20) Methylene Chloride	2.794	84	50459	21.112	ug/l	# 88
21) trans-1,2-Dichloroethene	3.099	96	44539	21.250	ug/l	100
22) Diisopropyl ether	3.769	45	157410	22.230	ug/l	91
23) Vinyl Acetate	3.733	43	695761	111.308	ug/l	95
24) 1,1-Dichloroethane	3.617	63	85413	21.946	ug/l	97
25) 2-Butanone	4.568	43	210967	110.596	ug/l	93
26) 2,2-Dichloropropane	4.489	77	67557	21.224	ug/l	99
27) cis-1,2-Dichloroethene	4.501	96	52428	21.539	ug/l	96
28) Bromochloromethane	4.915	49	39911	21.973	ug/l	92
29) Tetrahydrofuran	5.019	42	139467	111.543	ug/l	89
30) Chloroform	5.104	83	88141	22.127	ug/l	99
31) Cyclohexane	5.476	56	72982	19.669	ug/l	96
32) 1,1,1-Trichloroethane	5.397	97	75510	22.180	ug/l	98
36) 1,1-Dichloropropene	5.696	75	62598	21.199	ug/l	100
37) Ethyl Acetate	4.726	43	79844	21.445	ug/l	# 95
38) Carbon Tetrachloride	5.690	117	63455	21.130	ug/l	96
39) Methylcyclohexane	7.384	83	77708	20.837	ug/l	95
40) Benzene	6.043	78	189696	21.666	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080521\
 Data File : VX023547.D
 Acq On : 04 Aug 2021 17:59
 Operator : JC/MD
 Sample : VX0805WBS01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0805WBS01

Quant Time: Aug 05 06:08:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.928	41	42437	21.587	ug/l	92
42) 1,2-Dichloroethane	6.098	62	73704	22.768	ug/l	95
43) Isopropyl Acetate	6.348	43	126725	21.965	ug/l	93
44) Trichloroethene	7.134	130	51676	21.484	ug/l	91
45) 1,2-Dichloropropane	7.439	63	50097	21.623	ug/l	99
46) Dibromomethane	7.586	93	35168	22.379	ug/l	99
47) Bromodichloromethane	7.829	83	61840	21.842	ug/l	98
48) Methyl methacrylate	7.701	41	60633	22.692	ug/l	88
49) 1,4-Dioxane	7.665	88	26839	471.721	ug/l	98
51) 4-Methyl-2-Pentanone	8.579	43	423382	119.980	ug/l	90
52) Toluene	8.720	92	122570	22.597	ug/l	99
53) t-1,3-Dichloropropene	8.982	75	69977	20.596	ug/l	96
54) cis-1,3-Dichloropropene	8.372	75	77015	21.821	ug/l	92
55) 1,1,2-Trichloroethane	9.158	97	51745	23.558	ug/l	99
56) Ethyl methacrylate	9.122	69	80532	22.873	ug/l #	83
57) 1,3-Dichloropropane	9.311	76	85405	22.883	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.244	63	186865	100.188	ug/l	97
59) 2-Hexanone	9.433	43	328421	122.560	ug/l	88
60) Dibromochloromethane	9.524	129	47683	20.793	ug/l	99
61) 1,2-Dibromoethane	9.616	107	54182	23.121	ug/l	99
64) Tetrachloroethene	9.274	164	50784	21.228	ug/l	94
65) Chlorobenzene	10.085	112	132633	21.558	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.164	131	48815	21.220	ug/l	100
67) Ethyl Benzene	10.195	91	233415	21.875	ug/l	99
68) m/p-Xylenes	10.305	106	175994	42.886	ug/l	100
69) o-Xylene	10.646	106	88732	21.809	ug/l	97
70) Styrene	10.658	104	144251	22.377	ug/l	100
71) Bromoform	10.805	173	33430	20.291	ug/l #	99
73) Isopropylbenzene	10.963	105	230176	22.558	ug/l	98
74) N-amyl acetate	10.847	43	105114	22.580	ug/l	90
75) 1,1,2,2-Tetrachloroethane	11.213	83	77609	23.349	ug/l	99
76) 1,2,3-Trichloropropane	11.243	75	87766	29.494	ug/l	90
77) Bromobenzene	11.201	156	57183	23.231	ug/l	94
78) n-propylbenzene	11.304	91	259687	22.487	ug/l	100
79) 2-Chlorotoluene	11.365	91	155762	22.572	ug/l	97
80) 1,3,5-Trimethylbenzene	11.457	105	189756	22.835	ug/l	99
81) trans-1,4-Dichloro-2-b...	11.024	75	22002	19.881	ug/l #	85
82) 4-Chlorotoluene	11.457	91	176347	22.589	ug/l	99
83) tert-Butylbenzene	11.719	119	186920	22.932	ug/l	97
84) 1,2,4-Trimethylbenzene	11.756	105	191977	23.226	ug/l	99
85) sec-Butylbenzene	11.896	105	235250	22.259	ug/l	99
86) p-Isopropyltoluene	12.012	119	192793	22.589	ug/l	98
87) 1,3-Dichlorobenzene	11.975	146	100138	21.990	ug/l	99
88) 1,4-Dichlorobenzene	12.042	146	100055	21.654	ug/l	99
89) n-Butylbenzene	12.335	91	165300	21.395	ug/l	98
90) Hexachloroethane	12.542	117	29208	20.281	ug/l	95
91) 1,2-Dichlorobenzene	12.341	146	99680	22.607	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.944	75	15586	20.628	ug/l	91
93) 1,2,4-Trichlorobenzene	13.591	180	59469	21.467	ug/l	97
94) Hexachlorobutadiene	13.725	225	26086	20.983	ug/l	100
95) Naphthalene	13.780	128	208676	21.909	ug/l	99
96) 1,2,3-Trichlorobenzene	13.962	180	59075	21.117	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080521\
Data File : VX023547.D
Acq On : 04 Aug 2021 17:59
Operator : JC/MD
Sample : VX0805WBS01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0805WBS01


Quant Time: Aug 05 06:08:39 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Aug 05 06:08:39 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.562 min Scan# 734
Delta R.T. -0.000 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59




Instrument :
MSVOA_X
ClientSampleId :
VX0805WBS01


Tgt Ion:168 Resp: 200800
Ion Ratio Lower Upper
168 100
99 57.6 45.8 68.6



#2
Dichlorodifluoromethane
Concen: 21.820 ug/l
RT: 1.172 min Scan# 14
Delta R.T. 0.006 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59

Tgt Ion: 85 Resp: 40309
Ion Ratio Lower Upper
85 100
87 28.0 16.4 49.2









#3
Chloromethane
Concen: 21.289 ug/l
RT: 1.294 min Scan# 34
Delta R.T. 0.000 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59




Instrument :
MSVOA_X
ClientSampleId :
VX0805WBS01


Tgt Ion: 50 Resp: 45814
Ion Ratio Lower Upper
50 100
52 33.2 26.4 39.6



#4
Vinyl Chloride
Concen: 21.492 ug/l
RT: 1.373 min Scan# 47
Delta R.T. -0.001 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59

Tgt Ion: 62 Resp: 46817
Ion Ratio Lower Upper
62 100
64 32.6 25.7 38.5









#5
Bromomethane
Concen: 25.974 ug/l
RT: 1.599 min Scan# 84
Delta R.T. -0.000 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59




Instrument :
MSVOA_X
ClientSampleId :
VX0805WBS01


Tgt Ion: 94 Resp: 33592
Ion Ratio Lower Upper
94 100
96 93.9 76.4 114.6



#6
Chloroethane
Concen: 22.766 ug/l
RT: 1.684 min Scan# 98
Delta R.T. 0.006 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59

Tgt Ion: 64 Resp: 30436
Ion Ratio Lower Upper
64 100
66 31.4 25.4 38.0









#7
Trichlorofluoromethane
Concen: 21.422 ug/l
RT: 1.885 min Scan# 131
Delta R.T. -0.001 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59




Instrument :
MSVOA_X
ClientSampleId :
VX0805WBS01


Tgt Ion:101 Resp: 75184
Ion Ratio Lower Upper
101 100
103 67.0 53.2 79.8



#8
Diethyl Ether
Concen: 21.556 ug/l
RT: 2.141 min Scan# 173
Delta R.T. 0.005 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59

Tgt Ion: 74 Resp: 27923
Ion Ratio Lower Upper
74 100
45 96.5 44.5 133.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 21.889 ug/l
RT: 2.330 min Scan# 204
Delta R.T. -0.001 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59




Instrument :
MSVOA_X
ClientSampleId :
VX0805WBS01


Tgt Ion:101 Resp: 43522
Ion Ratio Lower Upper
101 100
85 45.5 34.9 52.3
151 80.3 61.4 92.2



#10
Methyl Iodide
Concen: 20.980 ug/l
RT: 2.452 min Scan# 224
Delta R.T. -0.001 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59

Tgt Ion:142 Resp: 52265
Ion Ratio Lower Upper
142 100
127 43.9 33.3 49.9
141 14.3 11.4 17.0









#11
Tert butyl alcohol
Concen: 111.101 ug/l
RT: 2.977 min Scan# 310
Delta R.T. -0.006 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59




Instrument :
MSVOA_X
ClientSampleId :
VX0805WBS01


Tgt Ion: 59 Resp: 63572
Ion Ratio Lower Upper
59 100
57 10.8 7.8 11.8



#12
1,1-Dichloroethene
Concen: 21.068 ug/l
RT: 2.318 min Scan# 202
Delta R.T. -0.001 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59

Tgt Ion: 96 Resp: 41568
Ion Ratio Lower Upper
96 100
61 164.0 124.6 187.0
98 64.4 50.6 76.0









#13
Acrolein
Concen: 83.670 ug/l
RT: 2.239 min Scan# 189
Delta R.T. -0.000 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59




Instrument :
MSVOA_X
ClientSampleId :
VX0805WBS01


Tgt Ion: 56 Resp: 24073
Ion Ratio Lower Upper
56 100
55 74.6 56.4 84.6



#14
Allyl chloride
Concen: 20.943 ug/l
RT: 2.666 min Scan# 259
Delta R.T. -0.000 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59

Tgt Ion: 41 Resp: 75125
Ion Ratio Lower Upper
41 100
39 69.5 51.6 77.4
76 33.2 30.5 45.7









#15
Acrylonitrile
Concen: 113.424 ug/l
RT: 3.068 min Scan# 325
Delta R.T. -0.007 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59




Instrument :
MSVOA_X
ClientSampleId :
VX0805WBS01


Tgt Ion: 53	Resp: 146556
Ion Ratio	Lower Upper
53 100	
52 83.2	65.3 97.9
51 35.8	28.2 42.2



#16
Acetone
Concen: 108.188 ug/l
RT: 2.385 min Scan# 213
Delta R.T. -0.001 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59

Tgt Ion: 43	Resp: 132932
Ion Ratio	Lower Upper
43 100	
58 31.6	27.8 41.8









#17
Carbon Disulfide
Concen: 18.196 ug/l
RT: 2.513 min Scan# 234
Delta R.T. -0.001 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59




Instrument :
MSVOA_X
ClientSampled :
VX0805WBS01


Tgt Ion: 76 Resp: 88526
Ion Ratio Lower Upper
76 100
78 9.2 7.2 10.8



#18
Methyl Acetate
Concen: 23.266 ug/l
RT: 2.708 min Scan# 266
Delta R.T. -0.001 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59

Tgt Ion: 43 Resp: 67958
Ion Ratio Lower Upper
43 100
74 23.0 21.9 32.9









#19

Methyl tert-butyl Ether
 Concen: 21.888 ug/l
 RT: 3.123 min Scan# 334
 Delta R.T. -0.000 min
 Lab File: VX023547.D
 Acq: 04 Aug 2021 17:59

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0805WBS01




Tgt Ion: 73 Resp: 150158
 Ion Ratio Lower Upper
 73 100
 57 22.1 17.1 25.7




#20

Methylene Chloride
 Concen: 21.112 ug/l
 RT: 2.794 min Scan# 280
 Delta R.T. -0.000 min
 Lab File: VX023547.D
 Acq: 04 Aug 2021 17:59

Tgt Ion: 84 Resp: 50459
 Ion Ratio Lower Upper
 84 100
 49 129.0 88.6 132.8
 51 41.0 27.0 40.6#
 86 64.6 50.3 75.5









#21
trans-1,2-Dichloroethene
Concen: 21.250 ug/l
RT: 3.099 min Scan# 330
Delta R.T. 0.006 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59




Instrument :
MSVOA_X
ClientSampleId :
VX0805WBS01


Tgt Ion: 96 Resp: 44539
Ion Ratio Lower Upper
96 100
61 134.6 107.7 161.5
98 63.8 50.4 75.6



#22
Diisopropyl ether
Concen: 22.230 ug/l
RT: 3.769 min Scan# 440
Delta R.T. 0.000 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59

Tgt Ion: 45 Resp: 157410
Ion Ratio Lower Upper
45 100
43 81.5 58.0 87.0
87 28.8 25.4 38.0
59 10.6 10.2 15.2









#23
 Vinyl Acetate
 Concen: 111.308 ug/l
 RT: 3.733 min Scan# 434
 Delta R.T. -0.000 min
 Lab File: VX023547.D
 Acq: 04 Aug 2021 17:59




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0805WBS01


Tgt Ion: 43 Resp: 695761
 Ion Ratio Lower Upper
 43 100
 86 10.6 9.9 14.9



#24
 1,1-Dichloroethane
 Concen: 21.946 ug/l
 RT: 3.617 min Scan# 415
 Delta R.T. -0.000 min
 Lab File: VX023547.D
 Acq: 04 Aug 2021 17:59

Tgt Ion: 63 Resp: 85413
 Ion Ratio Lower Upper
 63 100
 98 6.7 3.8 11.2
 100 3.7 2.5 7.5









#25
2-Butanone
Concen: 110.596 ug/l
RT: 4.568 min Scan# 571
Delta R.T. -0.000 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59




Instrument :
MSVOA_X
ClientSampleId :
VX0805WBS01


Tgt Ion: 43 Resp: 210967
Ion Ratio Lower Upper
43 100
72 26.1 23.7 35.5



#26
2,2-Dichloropropane
Concen: 21.224 ug/l
RT: 4.489 min Scan# 558
Delta R.T. 0.006 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59

Tgt Ion: 77 Resp: 67557
Ion Ratio Lower Upper
77 100
97 23.9 11.7 35.1





#27

cis-1,2-Dichloroethene

Concen: 21.539 ug/l

RT: 4.501 min Scan# 560

Delta R.T. 0.006 min

Lab File: VX023547.D

Acq: 04 Aug 2021 17:59

Tgt Ion: 96 Resp: 52428





Ion	Ratio	Lower	Upper
96	100		
61	149.0	0.0	285.6
98	66.0	0.0	129.6

Instrument :

MSVOA_X

ClientSampleId :

VX0805WBS01



#28

Bromochloromethane

Concen: 21.973 ug/l

RT: 4.915 min Scan# 628




Delta R.T. 0.005 min


Lab File: VX023547.D

Acq: 04 Aug 2021 17:59

Tgt Ion: 49 Resp: 39911

Ion	Ratio	Lower	Upper
49	100		
129	1.9	0.0	4.0
130	77.4	68.2	102.2





#29

Tetrahydrofuran

Concen: 111.543 ug/l

RT: 5.019 min Scan# 645

Delta R.T. 0.006 min

Lab File: VX023547.D

Acq: 04 Aug 2021 17:59

Tgt Ion: 42 Resp: 139467

Ion Ratio Lower Upper

42 100

72 44.4 41.7 62.5





71 40.9 38.4 57.6

Instrument :

MSVOA_X

ClientSampleId :

VX0805WBS01



#30

Chloroform

Concen: 22.127 ug/l

RT: 5.104 min Scan# 659

Delta R.T. -0.001 min

Lab File: VX023547.D




Acq: 04 Aug 2021 17:59


Tgt Ion: 83 Resp: 88141

Ion Ratio Lower Upper

83 100


85 66.1 52.0 78.0








#31
Cyclohexane
Concen: 19.669 ug/l
RT: 5.476 min Scan# 720
Delta R.T. -0.001 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59

Instrument :
MSVOA_X
ClientSampleId :
VX0805WBS01




Tgt Ion: 56 Resp: 72982

Ion	Ratio	Lower	Upper
56	100		
69	29.9	25.3	37.9
84	85.1	71.4	107.0






#32
1,1,1-Trichloroethane
Concen: 22.180 ug/l
RT: 5.397 min Scan# 707
Delta R.T. 0.006 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59



Tgt Ion: 97 Resp: 75510

Ion	Ratio	Lower	Upper
97	100		
99	64.4	51.5	77.3
61	46.3	34.2	51.4





#33

1,2-Dichloroethane-d4

Concen: 55.344 ug/l

RT: 5.964 min Scan# 800

Delta R.T. -0.000 min

Lab File: VX023547.D

Acq: 04 Aug 2021 17:59

Tgt Ion: 65 Resp: 138626

Ion Ratio Lower Upper

65 100





67 51.9 0.0 105.4

Instrument :

MSVOA_X

ClientSampleId :

VX0805WBS01



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.769 min Scan# 932

Delta R.T. -0.000 min

Lab File: VX023547.D

Acq: 04 Aug 2021 17:59




Tgt Ion: 114 Resp: 326433


Ion Ratio Lower Upper

114 100

63 19.8 0.0 40.8

88 15.8 0.0 32.8





#35

Dibromofluoromethane

Concen: 54.226 ug/l

RT: 5.397 min Scan# 707

Delta R.T. -0.000 min

Lab File: VX023547.D

Acq: 04 Aug 2021 17:59





Tgt Ion:113 Resp: 110675

Ion Ratio Lower Upper

113 100

111 102.7 81.9 122.9

192 19.6 16.7 25.1



#36

1,1-Dichloropropene

Concen: 21.199 ug/l

RT: 5.696 min Scan# 756

Delta R.T. -0.000 min

Lab File: VX023547.D

Acq: 04 Aug 2021 17:59




Tgt Ion: 75 Resp: 62598


Ion Ratio Lower Upper

75 100

110 36.6 18.1 54.4

77 31.6 25.3 37.9









#37
Ethyl Acetate
Concen: 21.445 ug/l
RT: 4.726 min Scan# 597
Delta R.T. -0.001 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59




Instrument :
MSVOA_X
ClientSampleId :
VX0805WBS01


Tgt Ion:	43	Resp:	79844
Ion Ratio	Lower	Upper	
43	100		
61	14.2	12.8	19.2
70	10.4	10.5	15.7#



#38
Carbon Tetrachloride
Concen: 21.130 ug/l
RT: 5.690 min Scan# 755
Delta R.T. 0.006 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59

Tgt Ion:	117	Resp:	63455
Ion Ratio	Lower	Upper	
117	100		
119	99.5	76.2	114.2
121	30.3	24.8	37.2





#39

Methylcyclohexane

Concen: 20.837 ug/l

RT: 7.384 min Scan# 1033

Delta R.T. -0.001 min

Lab File: VX023547.D

Acq: 04 Aug 2021 17:59





Tgt Ion: 83 Resp: 77708

Ion Ratio Lower Upper

83 100

55 70.1 58.9 88.3

98 43.1 38.1 57.1



#40

Benzene

Concen: 21.666 ug/l

RT: 6.043 min Scan# 813

Delta R.T. 0.000 min

Lab File: VX023547.D




Acq: 04 Aug 2021 17:59


Tgt Ion: 78 Resp: 189696

Ion Ratio Lower Upper

78 100

77 22.8 18.9 28.3









#41
Methacrylonitrile
Concen: 21.587 ug/l
RT: 4.928 min Scan# 630
Delta R.T. -0.000 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59

Instrument :
MSVOA_X
ClientSampleId :
VX0805WBS01

Tgt Ion: 41 Resp: 42437




Ion	Ratio	Lower	Upper
41	100		
39	59.6	42.5	63.7
67	73.0	65.0	97.6
52	31.7	25.3	37.9




#42
1,2-Dichloroethane
Concen: 22.768 ug/l
RT: 6.098 min Scan# 822
Delta R.T. 0.000 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59

Tgt Ion: 62 Resp: 73704

Ion	Ratio	Lower	Upper
62	100		
98	8.4	0.0	20.2









#43
Isopropyl Acetate
Concen: 21.965 ug/l
RT: 6.348 min Scan# 863
Delta R.T. -0.000 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59




Instrument :
MSVOA_X
ClientSampleId :
VX0805WBS01


Tgt Ion:	43	Resp:	126725
Ion	Ratio	Lower	Upper
43	100		
61	19.0	18.2	27.4
87	13.2	12.5	18.7



#44
Trichloroethene
Concen: 21.484 ug/l
RT: 7.134 min Scan# 992
Delta R.T. 0.005 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59

Tgt Ion:	130	Resp:	51676
Ion	Ratio	Lower	Upper
130	100		
95	89.7	0.0	198.0









#45

1,2-Dichloropropane
Concen: 21.623 ug/l
RT: 7.439 min Scan# 1042
Delta R.T. -0.001 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59

Instrument :
MSVOA_X
ClientSampled :
VX0805WBS01




Tgt Ion: 63 Resp: 50097
Ion Ratio Lower Upper
63 100
65 32.9 25.7 38.5




#46

Dibromomethane
Concen: 22.379 ug/l
RT: 7.586 min Scan# 1066
Delta R.T. -0.000 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59

Tgt Ion: 93 Resp: 35168
Ion Ratio Lower Upper
93 100
95 81.9 66.8 100.2
174 102.7 83.1 124.7









#47
Bromodichloromethane
Concen: 21.842 ug/l
RT: 7.829 min Scan# 1106
Delta R.T. -0.001 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59




Instrument :
MSVOA_X
ClientSampleId :
VX0805WBS01


Tgt Ion: 83 Resp: 61840
Ion Ratio Lower Upper
83 100
85 65.5 51.4 77.0
127 9.1 6.8 10.2



#48
Methyl methacrylate
Concen: 22.692 ug/l
RT: 7.701 min Scan# 1085
Delta R.T. -0.001 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59

Tgt Ion: 41 Resp: 60633
Ion Ratio Lower Upper
41 100
69 79.3 76.0 114.0
39 60.3 44.6 67.0





#49

1,4-Dioxane

Concen: 471.721 ug/l

RT: 7.665 min Scan# 1079





Delta R.T. -0.000 min

Lab File: VX023547.D

Acq: 04 Aug 2021 17:59

Tgt Ion: 88 Resp: 26839

Ion	Ratio	Lower	Upper
88	100		
43	31.7	23.2	34.8
58	69.5	55.0	82.4



#50

Toluene-d8

Concen: 53.791 ug/l

RT: 8.652 min Scan# 1241




Delta R.T. -0.001 min


Lab File: VX023547.D

Acq: 04 Aug 2021 17:59

Tgt Ion: 98 Resp: 403359

Ion	Ratio	Lower	Upper
98	100		
100	64.3	51.9	77.9









#51
4-Methyl-2-Pentanone
Concen: 119.980 ug/l
RT: 8.579 min Scan# 1229
Delta R.T. -0.001 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59




Instrument :
MSVOA_X
ClientSampleId :
VX0805WBS01


Tgt Ion: 43 Resp: 423382
Ion Ratio Lower Upper
43 100
58 37.8 35.4 53.0



#52
Toluene
Concen: 22.597 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. -0.000 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59

Tgt Ion: 92 Resp: 122570
Ion Ratio Lower Upper
92 100
91 172.3 137.1 205.7









#53
t-1,3-Dichloropropene
Concen: 20.596 ug/l
RT: 8.982 min Scan# 1295
Delta R.T. -0.000 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59




Instrument :
MSVOA_X
ClientSampleId :
VX0805WBS01


Tgt Ion: 75 Resp: 69977
Ion Ratio Lower Upper
75 100
77 33.0 24.9 37.3



#54
cis-1,3-Dichloropropene
Concen: 21.821 ug/l
RT: 8.372 min Scan# 1195
Delta R.T. 0.000 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59

Tgt Ion: 75 Resp: 77015
Ion Ratio Lower Upper
75 100
77 29.9 25.7 38.5
39 50.9 35.1 52.7









#55
1,1,2-Trichloroethane
Concen: 23.558 ug/l
RT: 9.158 min Scan# 1324
Delta R.T. -0.001 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59




Instrument :
MSVOA_X
ClientSampleId :
VX0805WBS01


Tgt Ion:	97	Resp:	51745
Ion	Ratio	Lower	Upper
97	100		
83	85.2	69.3	103.9
85	54.9	44.3	66.5
99	64.1	50.3	75.5



#56
Ethyl methacrylate
Concen: 22.873 ug/l
RT: 9.122 min Scan# 1318
Delta R.T. -0.000 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59

Tgt Ion:	69	Resp:	80532
Ion	Ratio	Lower	Upper
69	100		
41	68.8	46.8	70.2
39	43.2	24.7	37.1









#57
1,3-Dichloropropane
Concen: 22.883 ug/l
RT: 9.311 min Scan# 1349
Delta R.T. -0.000 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59




Instrument :
MSVOA_X
ClientSampleId :
VX0805WBS01


Tgt Ion: 76 Resp: 85405
Ion Ratio Lower Upper
76 100
78 32.2 26.2 39.2



#58
2-Chloroethyl Vinyl ether
Concen: 100.188 ug/l
RT: 8.244 min Scan# 1174
Delta R.T. -0.006 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59

Tgt Ion: 63 Resp: 186865
Ion Ratio Lower Upper
63 100
106 27.7 23.5 35.3









#59
2-Hexanone
Concen: 122.560 ug/l
RT: 9.433 min Scan# 1369
Delta R.T. -0.000 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59




Instrument :
MSVOA_X
ClientSampleId :
VX0805WBS01


Tgt Ion: 43 Resp: 328421
Ion Ratio Lower Upper
43 100
58 52.5 30.9 92.8



#60
Dibromochloromethane
Concen: 20.793 ug/l
RT: 9.524 min Scan# 1384
Delta R.T. -0.001 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59

Tgt Ion: 129 Resp: 47683
Ion Ratio Lower Upper
129 100
127 77.7 38.4 115.2









#61
1,2-Dibromoethane
Concen: 23.121 ug/l
RT: 9.616 min Scan# 1399
Delta R.T. -0.000 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59




Instrument :
MSVOA_X
ClientSampleId :
VX0805WBS01


Tgt Ion:107 Resp: 54182
Ion Ratio Lower Upper
107 100
109 93.9 75.8 113.6



#62
4-Bromofluorobenzene
Concen: 57.521 ug/l
RT: 11.085 min Scan# 1640
Delta R.T. -0.000 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59

Tgt Ion: 95 Resp: 146671
Ion Ratio Lower Upper
95 100
174 82.0 0.0 154.6
176 78.5 0.0 155.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59




Instrument :
MSVOA_X
ClientSampleId :
VX0805WBS01


Tgt Ion:117 Resp: 300596
Ion Ratio Lower Upper
117 100
82 56.4 47.4 71.2
119 32.2 26.6 39.8



#64
Tetrachloroethene
Concen: 21.228 ug/l
RT: 9.274 min Scan# 1343
Delta R.T. -0.001 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59

Tgt Ion:164 Resp: 50784
Ion Ratio Lower Upper
164 100
166 130.8 103.4 155.2
129 102.8 70.8 106.2
131 93.3 72.5 108.7









#65
Chlorobenzene
Concen: 21.558 ug/l
RT: 10.085 min Scan# 1476
Delta R.T. 0.000 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59




Instrument :
MSVOA_X
ClientSampleId :
VX0805WBS01


Tgt Ion:112 Resp: 132633
Ion Ratio Lower Upper
112 100
114 32.6 25.8 38.6



#66
1,1,1,2-Tetrachloroethane
Concen: 21.220 ug/l
RT: 10.164 min Scan# 1489
Delta R.T. -0.001 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59

Tgt Ion:131 Resp: 48815
Ion Ratio Lower Upper
131 100
133 95.5 47.7 143.1
119 65.1 32.1 96.5









#67
Ethyl Benzene
Concen: 21.875 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. -0.000 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59




Instrument :
MSVOA_X
ClientSampleId :
VX0805WBS01


Tgt Ion: 91 Resp: 233415
Ion Ratio Lower Upper
91 100
106 30.4 24.6 37.0



#68
m/p-Xylenes
Concen: 42.886 ug/l
RT: 10.305 min Scan# 1512
Delta R.T. -0.000 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59

Tgt Ion: 106 Resp: 175994
Ion Ratio Lower Upper
106 100
91 203.9 162.8 244.2









#69
o-Xylene
Concen: 21.809 ug/l
RT: 10.646 min Scan# 1568
Delta R.T. 0.000 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59




Instrument :
MSVOA_X
ClientSampled :
VX0805WBS01


Tgt Ion:106 Resp: 88732
Ion Ratio Lower Upper
106 100
91 211.2 108.1 324.2



#70
Styrene
Concen: 22.377 ug/l
RT: 10.658 min Scan# 1570
Delta R.T. -0.001 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59

Tgt Ion:104 Resp: 144251
Ion Ratio Lower Upper
104 100
78 50.4 40.0 60.0
103 54.4 43.4 65.0









#71
Bromoform
Concen: 20.291 ug/l
RT: 10.805 min Scan# 1594
Delta R.T. -0.000 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59




Instrument :
MSVOA_X
ClientSampleId :
VX0805WBS01


Tgt Ion:173 Resp: 33430
Ion Ratio Lower Upper
173 100
175 47.2 23.9 71.8
254 0.0 0.1 0.1#



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. -0.000 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59

Tgt Ion:152 Resp: 135225
Ion Ratio Lower Upper
152 100
115 70.6 41.1 123.3
150 166.3 0.0 349.0









#73
Isopropylbenzene
Concen: 22.558 ug/l
RT: 10.963 min Scan# 1620
Delta R.T. 0.000 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59




Instrument :
MSVOA_X
ClientSampleId :
VX0805WBS01


Tgt Ion:105 Resp: 230176
Ion Ratio Lower Upper
105 100
120 27.3 13.2 39.6



#74
N-ethyl acetate
Concen: 22.580 ug/l
RT: 10.847 min Scan# 1601
Delta R.T. -0.001 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59

Tgt Ion: 43 Resp: 105114
Ion Ratio Lower Upper
43 100
70 40.8 39.0 58.6
55 24.0 21.5 32.3
61 22.6 22.1 33.1









#75
1,1,2,2-Tetrachloroethane
Concen: 23.349 ug/l
RT: 11.213 min Scan# 1661
Delta R.T. -0.000 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59




Instrument :
MSVOA_X
ClientSampleId :
VX0805WBS01


Tgt Ion: 83 Resp: 77609
Ion Ratio Lower Upper
83 100
131 10.7 5.1 15.2
85 64.2 31.8 95.4



#76
1,2,3-Trichloropropane
Concen: 29.494 ug/l
RT: 11.243 min Scan# 1666
Delta R.T. -0.001 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59

Tgt Ion: 75 Resp: 87766
Ion Ratio Lower Upper
75 100
77 32.5 19.1 57.5









#77
Bromobenzene
Concen: 23.231 ug/l
RT: 11.201 min Scan# 1659
Delta R.T. -0.000 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59




Instrument :
MSVOA_X
ClientSampleId :
VX0805WBS01


Tgt Ion:156 Resp: 57183
Ion Ratio Lower Upper
156 100
77 156.5 82.8 248.6
158 94.7 49.1 147.4



#78
n-propylbenzene
Concen: 22.487 ug/l
RT: 11.304 min Scan# 1676
Delta R.T. -0.007 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59

Tgt Ion: 91 Resp: 259687
Ion Ratio Lower Upper
91 100
120 23.2 11.5 34.5









#79
2-Chlorotoluene
Concen: 22.572 ug/l
RT: 11.365 min Scan# 1686
Delta R.T. -0.001 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59




Instrument :
MSVOA_X
ClientSampleId :
VX0805WBS01


Tgt Ion: 91 Resp: 155762
Ion Ratio Lower Upper
91 100
126 34.9 16.6 49.8



#80
1,3,5-Trimethylbenzene
Concen: 22.835 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. -0.000 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59

Tgt Ion: 105 Resp: 189756
Ion Ratio Lower Upper
105 100
120 48.4 23.8 71.5









#81
trans-1,4-Dichloro-2-butene
Concen: 19.881 ug/l
RT: 11.024 min Scan# 1630
Delta R.T. -0.000 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59




Instrument :
MSVOA_X
ClientSampleId :
VX0805WBS01


Tgt Ion: 75 Resp: 22002
Ion Ratio Lower Upper
75 100
53 114.1 74.5 111.7#
89 48.5 37.9 56.9



#82
4-Chlorotoluene
Concen: 22.589 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. -0.000 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59

Tgt Ion: 91 Resp: 176347
Ion Ratio Lower Upper
91 100
126 30.4 14.9 44.5









#83
tert-Butylbenzene
Concen: 22.932 ug/l
RT: 11.719 min Scan# 1744
Delta R.T. -0.000 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59




Instrument :
MSVOA_X
ClientSampleId :
VX0805WBS01


Tgt Ion:	119	Resp:	186920
Ion	Ratio	Lower	Upper
119	100		
91	58.9	27.9	83.7
134	22.9	11.3	33.9



#84
1,2,4-Trimethylbenzene
Concen: 23.226 ug/l
RT: 11.756 min Scan# 1750
Delta R.T. -0.000 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59

Tgt Ion:	105	Resp:	191977
Ion	Ratio	Lower	Upper
105	100		
120	44.3	22.4	67.3









#85
 sec-Butylbenzene
 Concen: 22.259 ug/l
 RT: 11.896 min Scan# 1773
 Delta R.T. -0.000 min
 Lab File: VX023547.D
 Acq: 04 Aug 2021 17:59




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0805WBS01


Tgt Ion:105 Resp: 235250
 Ion Ratio Lower Upper
 105 100
 134 20.2 10.0 29.8



#86
 p-Isopropyltoluene
 Concen: 22.589 ug/l
 RT: 12.012 min Scan# 1792
 Delta R.T. -0.000 min
 Lab File: VX023547.D
 Acq: 04 Aug 2021 17:59

Tgt Ion:119 Resp: 192793
 Ion Ratio Lower Upper
 119 100
 134 26.5 12.9 38.7
 91 24.7 11.7 35.0









#87
1,3-Dichlorobenzene
Concen: 21.990 ug/l
RT: 11.975 min Scan# 1786
Delta R.T. 0.000 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59




Instrument :
MSVOA_X
ClientSampled :
VX0805WBS01


Tgt Ion:146 Resp: 100138
Ion Ratio Lower Upper
146 100
111 39.5 20.2 60.5
148 62.9 31.9 95.7



#88
1,4-Dichlorobenzene
Concen: 21.654 ug/l
RT: 12.042 min Scan# 1797
Delta R.T. 0.000 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59

Tgt Ion:146 Resp: 100055
Ion Ratio Lower Upper
146 100
111 40.1 19.6 58.8
148 63.6 31.6 95.0









#89
n-Butylbenzene
Concen: 21.395 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. -0.000 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59




Instrument :
MSVOA_X
ClientSampleId :
VX0805WBS01


Tgt Ion: 91 Resp: 165300
Ion Ratio Lower Upper
91 100
92 52.3 27.4 82.0
134 26.2 13.2 39.5



#90
Hexachloroethane
Concen: 20.281 ug/l
RT: 12.542 min Scan# 1879
Delta R.T. 0.000 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59

Tgt Ion: 117 Resp: 29208
Ion Ratio Lower Upper
117 100
201 76.4 40.4 121.2









#91
1,2-Dichlorobenzene
Concen: 22.607 ug/l
RT: 12.341 min Scan# 1846
Delta R.T. -0.000 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59




Instrument :
MSVOA_X
ClientSampleId :
VX0805WBS01


Tgt Ion:146 Resp: 99680
Ion Ratio Lower Upper
146 100
111 40.9 20.8 62.2
148 63.4 31.9 95.5



#92
1,2-Dibromo-3-Chloropropane
Concen: 20.628 ug/l
RT: 12.944 min Scan# 1945
Delta R.T. -0.001 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59

Tgt Ion: 75 Resp: 15586
Ion Ratio Lower Upper
75 100
155 82.0 45.3 135.9
157 106.2 57.7 173.1









#93
 1,2,4-Trichlorobenzene
 Concen: 21.467 ug/l
 RT: 13.591 min Scan# 2051
 Delta R.T. -0.000 min
 Lab File: VX023547.D
 Acq: 04 Aug 2021 17:59




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0805WBS01


Tgt Ion:	180	Resp:	59469
Ion Ratio	Lower	Upper	
180	100		
182	92.5	47.7	143.1
145	31.7	15.0	45.0



#94
 Hexachlorobutadiene
 Concen: 20.983 ug/l
 RT: 13.725 min Scan# 2073
 Delta R.T. -0.006 min
 Lab File: VX023547.D
 Acq: 04 Aug 2021 17:59


Tgt Ion:	225	Resp:	26086
Ion Ratio	Lower	Upper	
225	100		
223	62.2	31.1	93.2
227	62.7	31.6	95.0








#95
Naphthalene
Concen: 21.909 ug/l
RT: 13.780 min Scan# 2082
Delta R.T. -0.000 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59

Instrument :
MSVOA_X
ClientSampleId :
VX0805WBS01



Tgt Ion:	128	Resp:	208676
Ion Ratio	Lower	Upper	
128	100		
127	13.1	10.3	15.5
129	10.6	8.8	13.2



#96
1,2,3-Trichlorobenzene
Concen: 21.117 ug/l
RT: 13.962 min Scan# 2112
Delta R.T. -0.001 min
Lab File: VX023547.D
Acq: 04 Aug 2021 17:59

Tgt Ion:	180	Resp:	59075
Ion Ratio	Lower	Upper	
180	100		
182	95.4	47.6	142.9
145	33.2	16.1	48.3

