

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080521\
 Data File : VX023549.D
 Acq On : 04 Aug 2021 18:46
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Aug 05 06:09:16 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	167240	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	275184	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	241488	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	90122	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	105906	50.766	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	101.540%	
35) Dibromofluoromethane	5.397	113	86657	50.365	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	100.740%	
50) Toluene-d8	8.653	98	328729	52.002	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	104.000%	
62) 4-Bromofluorobenzene	11.085	95	103454	48.128	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	96.260%	
Target Compounds						
					Qvalue	
20) Methylene Chloride	2.794	84	2548	1.280	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080521\
Data File : VX023549.D
Acq On : 04 Aug 2021 18:46
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Aug 05 06:09:16 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

