

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080521\
 Data File : VX023553.D
 Acq On : 04 Aug 2021 20:21
 Operator : JC/MD
 Sample : M3251-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SAMPLE-MANA

Quant Time: Aug 05 06:10:02 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	162444	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	269278	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	232717	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	78582	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	105018	51.826	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.660%
35) Dibromofluoromethane	5.397	113	86661	51.472	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.940%
50) Toluene-d8	8.653	98	322692	52.167	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	104.340%
62) 4-Bromofluorobenzene	11.085	95	95295	45.305	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.620%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080521\
Data File : VX023553.D
Acq On : 04 Aug 2021 20:21
Operator : JC/MD
Sample : M3251-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANA

Quant Time: Aug 05 06:10:02 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

