

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080521\
 Data File : VX023555.D
 Acq On : 04 Aug 2021 21:08
 Operator : JC/MD
 Sample : M3278-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-4

Quant Time: Aug 05 06:10:23 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	157201	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	263819	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	243039	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	97561	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	103629	52.846	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 105.700%	
35) Dibromofluoromethane	5.391	113	84587	51.280	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.560%	
50) Toluene-d8	8.653	98	318732	52.593	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 105.180%	
62) 4-Bromofluorobenzene	11.085	95	108196	52.503	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 105.000%	
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	5095	5.297	ug/l	89
17) Carbon Disulfide	2.514	76	5729	1.504	ug/l	100
30) Chloroform	5.117	83	1465	0.470	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080521\
Data File : VX023555.D
Acq On : 04 Aug 2021 21:08
Operator : JC/MD
Sample : M3278-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-4

Quant Time: Aug 05 06:10:23 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

