

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX080521\  
 Data File : VX023556.D  
 Acq On : 04 Aug 2021 21:32  
 Operator : JC/MD  
 Sample : M3278-03  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 14 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 MW-5

Quant Time: Aug 05 06:10:32 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X072221W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Jul 23 06:05:38 2021  
 Response via : Initial Calibration

| Compound                    | R.T. QIon |       | Response | Conc     | Units      | Dev(Min) |
|-----------------------------|-----------|-------|----------|----------|------------|----------|
| -----                       |           |       |          |          |            |          |
| Internal Standards          |           |       |          |          |            |          |
| 1) Pentafluorobenzene       | 5.562     | 168   | 152724   | 50.000   | ug/l       | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.769     | 114   | 262332   | 50.000   | ug/l       | 0.00     |
| 63) Chlorobenzene-d5        | 10.061    | 117   | 240359   | 50.000   | ug/l       | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.024    | 152   | 95660    | 50.000   | ug/l       | 0.00     |
| System Monitoring Compounds |           |       |          |          |            |          |
| 33) 1,2-Dichloroethane-d4   | 5.964     | 65    | 103522   | 54.339   | ug/l       | 0.00     |
| Spiked Amount               | 50.000    | Range | 78 - 117 | Recovery | = 108.680% |          |
| 35) Dibromofluoromethane    | 5.391     | 113   | 82123    | 50.068   | ug/l       | 0.00     |
| Spiked Amount               | 50.000    | Range | 75 - 124 | Recovery | = 100.140% |          |
| 50) Toluene-d8              | 8.653     | 98    | 317967   | 52.764   | ug/l       | 0.00     |
| Spiked Amount               | 50.000    | Range | 92 - 112 | Recovery | = 105.520% |          |
| 62) 4-Bromofluorobenzene    | 11.085    | 95    | 104417   | 50.956   | ug/l       | 0.00     |
| Spiked Amount               | 50.000    | Range | 83 - 123 | Recovery | = 101.920% |          |
| Target Compounds            |           |       |          |          |            |          |
|                             |           |       |          |          | Qvalue     |          |
| 16) Acetone                 | 2.386     | 43    | 2388     | 2.555    | ug/l #     | 87       |
| 20) Methylene Chloride      | 2.794     | 84    | 700      | 0.385    | ug/l #     | 71       |
| 30) Chloroform              | 5.105     | 83    | 4599     | 1.518    | ug/l       | 92       |
| -----                       |           |       |          |          |            |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX080521\  
Data File : VX023556.D  
Acq On : 04 Aug 2021 21:32  
Operator : JC/MD  
Sample : M3278-03  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 14 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
MW-5

Quant Time: Aug 05 06:10:32 2021  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X072221W.M  
Quant Title : SW846 8260  
QLast Update : Fri Jul 23 06:05:38 2021  
Response via : Initial Calibration

