

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080521\
 Data File : VX023557.D
 Acq On : 04 Aug 2021 21:55
 Operator : JC/MD
 Sample : M3276-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PRE-MAEC-01

Quant Time: Aug 05 06:10:41 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	159508	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	270004	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	253558	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	99430	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	104571	52.556	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 105.120%			
35) Dibromofluoromethane	5.397	113	85809	50.829	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 101.660%			
50) Toluene-d8	8.653	98	336421	54.240	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 108.480%			
62) 4-Bromofluorobenzene	11.085	95	111265	52.755	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 105.520%			
Target Compounds						Qvalue
11) Tert butyl alcohol	2.977	59	2726	5.997	ug/l	99
16) Acetone	2.386	43	21443	21.969	ug/l	90
20) Methylene Chloride	2.794	84	15423	8.124	ug/l	91
25) 2-Butanone	4.581	43	15791	10.421	ug/l	91
27) cis-1,2-Dichloroethene	4.507	96	2296	1.187	ug/l	89
43) Isopropyl Acetate	6.355	43	22623	4.741	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080521\
Data File : VX023557.D
Acq On : 04 Aug 2021 21:55
Operator : JC/MD
Sample : M3276-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PRE-MAEC-01

Quant Time: Aug 05 06:10:41 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

