

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080521\
 Data File : VX023558.D
 Acq On : 04 Aug 2021 22:19
 Operator : JC/MD
 Sample : M3276-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PRE-MAEC-02

Quant Time: Aug 05 06:10:52 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	154461	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	259746	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	248891	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	100229	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	102467	53.181	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 106.360%		
35) Dibromofluoromethane	5.397	113	83647	51.505	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 103.020%		
50) Toluene-d8	8.653	98	326680	54.750	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 109.500%		
62) 4-Bromofluorobenzene	11.085	95	110261	54.344	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 108.680%		
Target Compounds						
						Qvalue
11) Tert butyl alcohol	2.977	59	2600	5.907	ug/l	# 95
16) Acetone	2.386	43	20184	21.355	ug/l	90
20) Methylene Chloride	2.794	84	13277	7.222	ug/l	92
25) 2-Butanone	4.574	43	15600	10.632	ug/l	93
27) cis-1,2-Dichloroethene	4.501	96	3584	1.914	ug/l	90
43) Isopropyl Acetate	6.349	43	21459	4.674	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080521\
Data File : VX023558.D
Acq On : 04 Aug 2021 22:19
Operator : JC/MD
Sample : M3276-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PRE-MAEC-02

Quant Time: Aug 05 06:10:52 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

