

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080521\
 Data File : VX023569.D
 Acq On : 05 Aug 2021 13:52
 Operator : JC/MD
 Sample : M3279-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SPLP-AMENDED-COM-3

Quant Time: Aug 06 03:22:59 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	152661	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	259712	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	235512	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	92812	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	101494	53.297	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 106.600%			
35) Dibromofluoromethane	5.391	113	69721	42.936	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 85.880%			
50) Toluene-d8	8.653	98	309871	51.939	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 103.880%			
62) 4-Bromofluorobenzene	11.085	95	103321	50.930	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 101.860%			
Target Compounds						
16) Acetone	2.386	43	18821	20.148	ug/l #	85
17) Carbon Disulfide	2.520	76	3583	0.969	ug/l #	94
20) Methylene Chloride	2.794	84	4432	2.439	ug/l #	82
25) 2-Butanone	4.574	43	4988	3.439	ug/l	97
95) Naphthalene	13.774	128	2101	0.321	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080521\
Data File : VX023569.D
Acq On : 05 Aug 2021 13:52
Operator : JC/MD
Sample : M3279-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SPLP-AMENDED-COM-3

Quant Time: Aug 06 03:22:59 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

