

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080521\
 Data File : VX023570.D
 Acq On : 05 Aug 2021 14:15
 Operator : JC/MD
 Sample : M3297-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 125-27

Quant Time: Aug 06 03:23:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	149772	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	254233	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	230571	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	91551	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	99738	53.385	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 106.780%	
35) Dibromofluoromethane	5.391	113	81101	51.020	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.040%	
50) Toluene-d8	8.653	98	305105	52.243	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 104.480%	
62) 4-Bromofluorobenzene	11.085	95	98859	49.781	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 99.560%	
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	1893	2.066	ug/l	97
17) Carbon Disulfide	2.514	76	1158	0.319	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080521\
Data File : VX023570.D
Acq On : 05 Aug 2021 14:15
Operator : JC/MD
Sample : M3297-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
125-27

Quant Time: Aug 06 03:23:09 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

