

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080521\
 Data File : VX023571.D
 Acq On : 05 Aug 2021 14:39
 Operator : JC/MD
 Sample : M3297-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 123-24

Quant Time: Aug 06 03:23:19 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	145219	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	242652	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	216920	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	85772	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	96084	53.042	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	106.080%
35) Dibromofluoromethane	5.397	113	77403	51.018	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.040%
50) Toluene-d8	8.653	98	293260	52.611	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	105.220%
62) 4-Bromofluorobenzene	11.085	95	94404	49.806	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.620%
Target Compounds						
						Qvalue
11) Tert butyl alcohol	2.977	59	23810	57.538	ug/l #	73
16) Acetone	2.386	43	2618	2.946	ug/l #	87
19) Methyl tert-butyl Ether	3.123	73	2806494	565.676	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080521\
Data File : VX023571.D
Acq On : 05 Aug 2021 14:39
Operator : JC/MD
Sample : M3297-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
123-24

Quant Time: Aug 06 03:23:19 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

