

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080521\
 Data File : VX023572.D
 Acq On : 05 Aug 2021 15:03
 Operator : JC/MD
 Sample : M3297-03DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 123-24DL

Quant Time: Aug 06 03:23:30 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	152349	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	257614	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	234830	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	95441	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	103413	54.416	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 108.840%	
35) Dibromofluoromethane	5.397	113	83619	51.914	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.820%	
50) Toluene-d8	8.653	98	311654	52.664	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 105.320%	
62) 4-Bromofluorobenzene	11.085	95	104523	51.942	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 103.880%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.983	59	2996	6.901	ug/l #	73
16) Acetone	2.392	43	1427	1.531	ug/l #	71
19) Methyl tert-butyl Ether	3.123	73	290897	55.889	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080521\
Data File : VX023572.D
Acq On : 05 Aug 2021 15:03
Operator : JC/MD
Sample : M3297-03DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
123-24DL

Quant Time: Aug 06 03:23:30 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

