

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080521\
 Data File : VX023573.D
 Acq On : 05 Aug 2021 15:26
 Operator : JC/MD
 Sample : M3301-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MH-P17

Quant Time: Aug 06 03:23:41 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	148232	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	251175	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	225061	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	93078	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	99255	53.679	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 107.360%	
35) Dibromofluoromethane	5.397	113	80617	51.333	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.660%	
50) Toluene-d8	8.653	98	300902	52.150	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 104.300%	
62) 4-Bromofluorobenzene	11.085	95	99599	50.764	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 101.520%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.983	59	3162	7.486	ug/l #	93
16) Acetone	2.386	43	4321	4.764	ug/l #	72
19) Methyl tert-butyl Ether	3.123	73	1522	0.301	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080521\
Data File : VX023573.D
Acq On : 05 Aug 2021 15:26
Operator : JC/MD
Sample : M3301-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MH-P17

Quant Time: Aug 06 03:23:41 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

