

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080521\
 Data File : VX023574.D
 Acq On : 05 Aug 2021 15:50
 Operator : JC/MD
 Sample : M3297-05 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 121

Quant Time: Aug 06 03:23:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	152741	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	257833	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	238601	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	98238	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	103894	54.529	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery	= 109.060%		
35) Dibromofluoromethane	5.397	113	83078	51.534	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery	= 103.060%		
50) Toluene-d8	8.653	98	314077	53.028	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery	= 106.060%		
62) 4-Bromofluorobenzene	11.085	95	104373	51.824	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery	= 103.640%		
Target Compounds						Qvalue
16) Acetone	2.392	43	12148	12.998	ug/l #	85
20) Methylene Chloride	2.794	84	2967	1.632	ug/l	92
25) 2-Butanone	4.574	43	6372	4.391	ug/l #	88
51) 4-Methyl-2-Pentanone	8.580	43	2511	0.901	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080521\
Data File : VX023574.D
Acq On : 05 Aug 2021 15:50
Operator : JC/MD
Sample : M3297-05 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
121

Quant Time: Aug 06 03:23:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

