

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080521\
 Data File : VX023576.D
 Acq On : 05 Aug 2021 16:37
 Operator : JC/MD
 Sample : M3298-06 200X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 1940

Quant Time: Aug 06 03:24:16 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	145381	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	243382	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	222499	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	85731	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	98309	54.210	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 108.420%	
35) Dibromofluoromethane	5.397	113	77879	51.178	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.360%	
50) Toluene-d8	8.653	98	296921	53.108	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 106.220%	
62) 4-Bromofluorobenzene	11.085	95	95290	50.123	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 100.240%	
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	3934	4.422	ug/l	91
20) Methylene Chloride	2.794	84	3437	1.986	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080521\
Data File : VX023576.D
Acq On : 05 Aug 2021 16:37
Operator : JC/MD
Sample : M3298-06 200X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
1940

Quant Time: Aug 06 03:24:16 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

