

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080521\
 Data File : VX023577.D
 Acq On : 05 Aug 2021 17:00
 Operator : JC/MD
 Sample : M3298-08 200X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ANTIFREEZE

Quant Time: Aug 06 03:24:27 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	155086	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	264240	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	248244	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	106690	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	105152	54.354	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 108.700%	
35) Dibromofluoromethane	5.397	113	83100	50.298	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.600%	
50) Toluene-d8	8.653	98	325890	53.688	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 107.380%	
62) 4-Bromofluorobenzene	11.085	95	114449	55.449	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 110.900%	
Target Compounds						
						Qvalue
16) Acetone	2.386	43	4538	4.782	ug/l	90
20) Methylene Chloride	2.794	84	3321	1.799	ug/l	93
52) Toluene	8.720	92	1484	0.338	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080521\
Data File : VX023577.D
Acq On : 05 Aug 2021 17:00
Operator : JC/MD
Sample : M3298-08 200X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ANTIFREEZE

Quant Time: Aug 06 03:24:27 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

