

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080521\
 Data File : VX023580.D
 Acq On : 05 Aug 2021 18:11
 Operator : JC/MD
 Sample : M3298-07 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 1964-1968

Quant Time: Aug 06 03:25:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	155585	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	263315	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	239670	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	99474	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	102301	52.711	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	105.420%
35) Dibromofluoromethane	5.391	113	83779	50.887	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.780%
50) Toluene-d8	8.653	98	323032	53.405	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	106.800%
62) 4-Bromofluorobenzene	11.085	95	107005	52.024	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	104.040%
Target Compounds						
						Qvalue
20) Methylene Chloride	2.794	84	2561	1.383	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080521\
Data File : VX023580.D
Acq On : 05 Aug 2021 18:11
Operator : JC/MD
Sample : M3298-07 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
1964-1968

Quant Time: Aug 06 03:25:01 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

