

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX080521\  
 Data File : VX023545.D  
 Acq On : 04 Aug 2021 16:40  
 Operator : JC/MD  
 Sample : VSTDCCC050  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VSTDCCC050

Manual Integrations  
 APPROVED

MMDadoda  
 8/6/2021 5:40:59 PM

Quant Time: Aug 05 06:08:10 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X072221W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Jul 23 06:05:38 2021  
 Response via : Initial Calibration

| Compound                     | R.T.           | QIon | Response   | Conc     | Units  | Dev(Min) |
|------------------------------|----------------|------|------------|----------|--------|----------|
| -----                        |                |      |            |          |        |          |
| Internal Standards           |                |      |            |          |        |          |
| 1) Pentafluorobenzene        | 5.562          | 168  | 212693     | 50.000   | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene      | 6.769          | 114  | 344599     | 50.000   | ug/l   | 0.00     |
| 63) Chlorobenzene-d5         | 10.055         | 117  | 325772     | 50.000   | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4   | 12.024         | 152  | 153231     | 50.000   | ug/l   | 0.00     |
|                              |                |      |            |          |        |          |
| System Monitoring Compounds  |                |      |            |          |        |          |
| 33) 1,2-Dichloroethane-d4    | 5.964          | 65   | 129129     | 48.670   | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 78 - 117 |      | Recovery = | 97.340%  |        |          |
| 35) Dibromofluoromethane     | 5.397          | 113  | 101140     | 46.942   | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 75 - 124 |      | Recovery = | 93.880%  |        |          |
| 50) Toluene-d8               | 8.653          | 98   | 363964     | 45.978   | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 92 - 112 |      | Recovery = | 91.960%# |        |          |
| 62) 4-Bromofluorobenzene     | 11.085         | 95   | 140540     | 52.211   | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 83 - 123 |      | Recovery = | 104.420% |        |          |
|                              |                |      |            |          |        |          |
| Target Compounds             |                |      |            |          | Qvalue |          |
| 2) Dichlorodifluoromethane   | 1.166          | 85   | 93422      | 47.743   | ug/l   | 100      |
| 3) Chloromethane             | 1.294          | 50   | 109947     | 48.234   | ug/l   | 98       |
| 4) Vinyl Chloride            | 1.374          | 62   | 108929     | 47.208   | ug/l   | 100      |
| 5) Bromomethane              | 1.599          | 94   | 65177      | 47.579   | ug/l   | 98       |
| 6) Chloroethane              | 1.678          | 64   | 72458      | 51.167   | ug/l   | 97       |
| 7) Trichlorofluoromethane    | 1.880          | 101  | 180891     | 48.660   | ug/l   | 96       |
| 8) Diethyl Ether             | 2.136          | 74   | 71240      | 51.920   | ug/l   | 89       |
| 9) 1,1,2-Trichlorotrifluo... | 2.325          | 101  | 103229     | 49.014   | ug/l   | 97       |
| 10) Methyl Iodide            | 2.453          | 142  | 145492     | 55.138   | ug/l   | 96       |
| 11) Tert butyl alcohol       | 2.983          | 59   | 154036     | 254.147  | ug/l   | 99       |
| 12) 1,1-Dichloroethene       | 2.319          | 96   | 97717      | 46.757   | ug/l   | 97       |
| 13) Acrolein                 | 2.239          | 56   | 74891      | 243.099  | ug/l   | 98       |
| 14) Allyl chloride           | 2.666          | 41   | 188436     | 49.594   | ug/l   | 94       |
| 15) Acrylonitrile            | 3.068          | 53   | 358717     | 262.098  | ug/l   | 98       |
| 16) Acetone                  | 2.386          | 43   | 382337     | 293.770  | ug/l   | 92       |
| 17) Carbon Disulfide         | 2.508          | 76   | 217628     | 42.231   | ug/l   | 99       |
| 18) Methyl Acetate           | 2.709          | 43   | 165267     | 53.417   | ug/l   | 93       |
| 19) Methyl tert-butyl Ether  | 3.123          | 73   | 392943     | 54.076   | ug/l   | 99       |
| 20) Methylene Chloride       | 2.794          | 84   | 124536     | 49.193   | ug/l   | 92       |
| 21) trans-1,2-Dichloroethene | 3.093          | 96   | 106106     | 47.794   | ug/l   | 93       |
| 22) Diisopropyl ether        | 3.769          | 45   | 393584     | 52.476   | ug/l   | 97       |
| 23) Vinyl Acetate            | 3.727          | 43   | 1789575    | 270.289  | ug/l   | 94       |
| 24) 1,1-Dichloroethane       | 3.611          | 63   | 207765     | 50.399   | ug/l   | 97       |
| 25) 2-Butanone               | 4.568          | 43   | 556692     | 275.519  | ug/l   | 92       |
| 26) 2,2-Dichloropropane      | 4.483          | 77   | 169497     | 50.271   | ug/l   | 99       |
| 27) cis-1,2-Dichloroethene   | 4.495          | 96   | 130679     | 50.685   | ug/l   | 97       |
| 28) Bromochloromethane       | 4.903          | 49   | 92819      | 48.245   | ug/l   | 95       |
| 29) Tetrahydrofuran          | 5.019          | 42   | 341300     | 257.702  | ug/l   | 89       |
| 30) Chloroform               | 5.099          | 83   | 216743     | 51.368   | ug/l   | 100      |
| 31) Cyclohexane              | 5.477          | 56   | 178423     | 45.397   | ug/l   | 95       |
| 32) 1,1,1-Trichloroethane    | 5.385          | 97   | 182598     | 50.636   | ug/l   | 97       |
| 36) 1,1-Dichloropropene      | 5.696          | 75   | 151754     | 48.682   | ug/l   | 100      |
| 37) Ethyl Acetate            | 4.727          | 43   | 200266     | 50.953   | ug/l   | 95       |
| 38) Carbon Tetrachloride     | 5.684          | 117  | 155678     | 49.107   | ug/l   | 100      |
| 39) Methylcyclohexane        | 7.385          | 83   | 189825     | 48.218   | ug/l   | 99       |
| 40) Benzene                  | 6.043          | 78   | 459124     | 49.675   | ug/l   | 98       |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX080521\  
 Data File : VX023545.D  
 Acq On : 04 Aug 2021 16:40  
 Operator : JC/MD  
 Sample : VSTDCCC050  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VSTDCCC050

Manual Integrations  
 APPROVED

MMDadoda  
 8/6/2021 5:40:59 PM

Quant Time: Aug 05 06:08:10 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X072221W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Jul 23 06:05:38 2021  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc     | Units  | Dev(Min) |
|-------------------------------|--------|------|----------|----------|--------|----------|
| 41) Methacrylonitrile         | 4.928  | 41   | 109546   | 52.786   | ug/l   | 92       |
| 42) 1,2-Dichloroethane        | 6.092  | 62   | 183633   | 53.735   | ug/l   | 95       |
| 43) Isopropyl Acetate         | 6.348  | 43   | 323428   | 53.105   | ug/l   | 93       |
| 44) Trichloroethene           | 7.129  | 130  | 126026   | 49.633   | ug/l   | 98       |
| 45) 1,2-Dichloropropane       | 7.433  | 63   | 126978   | 51.918   | ug/l   | 100      |
| 46) Dibromomethane            | 7.586  | 93   | 89731    | 54.089   | ug/l   | 98       |
| 47) Bromodichloromethane      | 7.824  | 83   | 162539   | 54.382   | ug/l   | 100      |
| 48) Methyl methacrylate       | 7.696  | 41   | 157615   | 55.879   | ug/l   | 87       |
| 49) 1,4-Dioxane               | 7.665  | 88   | 64476    | 1073.488 | ug/l # | 96       |
| 51) 4-Methyl-2-Pentanone      | 8.580  | 43   | 1058756  | 284.218  | ug/l   | 91       |
| 52) Toluene                   | 8.720  | 92   | 295442   | 51.596   | ug/l   | 98       |
| 53) t-1,3-Dichloropropene     | 8.982  | 75   | 190972   | 50.365   | ug/l   | 97       |
| 54) cis-1,3-Dichloropropene   | 8.372  | 75   | 201685   | 54.131   | ug/l   | 92       |
| 55) 1,1,2-Trichloroethane     | 9.153  | 97   | 131931   | 56.898   | ug/l   | 99       |
| 56) Ethyl methacrylate        | 9.122  | 69   | 213191   | 57.359   | ug/l # | 82       |
| 57) 1,3-Dichloropropane       | 9.311  | 76   | 215938   | 54.807   | ug/l   | 99       |
| 58) 2-Chloroethyl Vinyl ether | 8.244  | 63   | 520020   | 264.111  | ug/l   | 97       |
| 59) 2-Hexanone                | 9.433  | 43   | 824685   | 291.532  | ug/l   | 88       |
| 60) Dibromochloromethane      | 9.525  | 129  | 133256   | 51.779   | ug/l   | 99       |
| 61) 1,2-Dibromoethane         | 9.616  | 107  | 139011   | 56.192   | ug/l   | 99       |
| 64) Tetrachloroethene         | 9.275  | 164  | 123458   | 47.617   | ug/l   | 94       |
| 65) Chlorobenzene             | 10.079 | 112  | 332459   | 49.861   | ug/l   | 99       |
| 66) 1,1,1,2-Tetrachloroethane | 10.165 | 131  | 125697   | 49.078   | ug/l   | 99       |
| 67) Ethyl Benzene             | 10.195 | 91   | 581711   | 50.303   | ug/l   | 99       |
| 68) m/p-Xylenes               | 10.305 | 106  | 447283   | 100.570  | ug/l   | 99       |
| 69) o-Xylene                  | 10.646 | 106  | 224748   | 50.970   | ug/l   | 98       |
| 70) Styrene                   | 10.659 | 104  | 375482   | 53.746   | ug/l   | 99       |
| 71) Bromoform                 | 10.805 | 173  | 96940    | 48.394   | ug/l # | 99       |
| 73) Isopropylbenzene          | 10.963 | 105  | 579833   | 50.148   | ug/l   | 99       |
| 74) N-amyl acetate            | 10.848 | 43   | 282787   | 53.609   | ug/l   | 89       |
| 75) 1,1,2,2-Tetrachloroethane | 11.213 | 83   | 192859   | 51.204   | ug/l   | 99       |
| 76) 1,2,3-Trichloropropane    | 11.244 | 75   | 184485m  | 54.712   | ug/l   |          |
| 77) Bromobenzene              | 11.201 | 156  | 147092   | 52.736   | ug/l   | 97       |
| 78) n-propylbenzene           | 11.305 | 91   | 667741   | 51.027   | ug/l   | 99       |
| 79) 2-Chlorotoluene           | 11.366 | 91   | 402066   | 51.418   | ug/l   | 98       |
| 80) 1,3,5-Trimethylbenzene    | 11.457 | 105  | 485882   | 51.600   | ug/l   | 99       |
| 81) trans-1,4-Dichloro-2-b... | 11.018 | 75   | 65852    | 52.512   | ug/l   | 92       |
| 82) 4-Chlorotoluene           | 11.457 | 91   | 462668   | 52.300   | ug/l   | 100      |
| 83) tert-Butylbenzene         | 11.719 | 119  | 472301   | 51.134   | ug/l   | 96       |
| 84) 1,2,4-Trimethylbenzene    | 11.756 | 105  | 486521   | 51.944   | ug/l   | 100      |
| 85) sec-Butylbenzene          | 11.896 | 105  | 611372   | 51.049   | ug/l   | 100      |
| 86) p-Isopropyltoluene        | 12.012 | 119  | 505811   | 52.301   | ug/l   | 98       |
| 87) 1,3-Dichlorobenzene       | 11.969 | 146  | 265519   | 51.456   | ug/l   | 99       |
| 88) 1,4-Dichlorobenzene       | 12.042 | 146  | 264254   | 50.469   | ug/l   | 100      |
| 89) n-Butylbenzene            | 12.335 | 91   | 453590   | 51.810   | ug/l   | 97       |
| 90) Hexachloroethane          | 12.542 | 117  | 82062    | 47.477   | ug/l   | 95       |
| 91) 1,2-Dichlorobenzene       | 12.335 | 146  | 254211   | 50.879   | ug/l   | 99       |
| 92) 1,2-Dibromo-3-Chloropr... | 12.945 | 75   | 42086    | 46.225   | ug/l   | 93       |
| 93) 1,2,4-Trichlorobenzene    | 13.591 | 180  | 163776   | 52.172   | ug/l   | 99       |
| 94) Hexachlorobutadiene       | 13.731 | 225  | 69016    | 48.991   | ug/l   | 99       |
| 95) Naphthalene               | 13.780 | 128  | 576770   | 53.439   | ug/l   | 100      |
| 96) 1,2,3-Trichlorobenzene    | 13.963 | 180  | 165886   | 52.331   | ug/l   | 99       |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX080521\  
Data File : VX023545.D  
Acq On : 04 Aug 2021 16:40  
Operator : JC/MD  
Sample : VSTDCCC050  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 2 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050

Manual Integrations  
APPROVED

MMDadoda  
8/6/2021 5:40:59 PM


Quant Time: Aug 05 06:08:10 2021  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X072221W.M  
Quant Title : SW846 8260  
QLast Update : Fri Jul 23 06:05:38 2021  
Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|----------|------|------|----------|------|-------|----------|
|----------|------|------|----------|------|-------|----------|


-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed







#1  
Pentafluorobenzene  
Concen: 50.000 ug/l  
RT: 5.562 min Scan# 734  
Delta R.T. -0.000 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40




Tgt Ion:168 Resp: 212693  
Ion Ratio Lower Upper  
168 100  
99 54.6 45.8 68.6




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050

Manual Integrations  
APPROVED



MMDadoda  
8/6/2021 5:40:59 PM




#2  
Dichlorodifluoromethane  
Concen: 47.743 ug/l  
RT: 1.166 min Scan# 13  
Delta R.T. 0.000 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40



Tgt Ion: 85 Resp: 93422  
Ion Ratio Lower Upper  
85 100  
87 33.0 16.4 49.2









#3  
Chloromethane  
Concen: 48.234 ug/l  
RT: 1.294 min Scan# 34  
Delta R.T. 0.000 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40

Tgt Ion: 50 Resp: 109947  
Ion Ratio Lower Upper  
50 100  
52 31.8 26.4 39.6

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050




Manual Integrations  
APPROVED


MMDadoda  
8/6/2021 5:40:59 PM



#4  
Vinyl Chloride  
Concen: 47.208 ug/l  
RT: 1.374 min Scan# 47  
Delta R.T. -0.000 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40

Tgt Ion: 62 Resp: 108929  
Ion Ratio Lower Upper  
62 100  
64 32.0 25.7 38.5









#5  
Bromomethane  
Concen: 47.579 ug/l  
RT: 1.599 min Scan# 84  
Delta R.T. 0.000 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40

Tgt Ion: 94 Resp: 65177  
Ion Ratio Lower Upper  
94 100  
96 94.0 76.4 114.6

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050




Manual Integrations  
APPROVED


MMDadoda  
8/6/2021 5:40:59 PM



#6  
Chloroethane  
Concen: 51.167 ug/l  
RT: 1.678 min Scan# 97  
Delta R.T. 0.000 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40

Tgt Ion: 64 Resp: 72458  
Ion Ratio Lower Upper  
64 100  
66 30.0 25.4 38.0









#7  
Trichlorofluoromethane  
Concen: 48.660 ug/l  
RT: 1.880 min Scan# 130  
Delta R.T. -0.006 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40

Tgt Ion:101 Resp: 180891  
Ion Ratio Lower Upper  
101 100  
103 63.1 53.2 79.8

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050




Manual Integrations  
APPROVED

MMDadoda  
8/6/2021 5:40:59 PM




#8  
Diethyl Ether  
Concen: 51.920 ug/l  
RT: 2.136 min Scan# 172  
Delta R.T. -0.000 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40

Tgt Ion: 74 Resp: 71240  
Ion Ratio Lower Upper  
74 100  
45 98.9 44.5 133.5











#9  
1,1,2-Trichlorotrifluoroethane  
Concen: 49.014 ug/l  
RT: 2.325 min Scan# 203  
Delta R.T. -0.006 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40

Tgt Ion:101 Resp: 103229  
Ion Ratio Lower Upper  
101 100  
85 44.1 34.9 52.3  
151 81.0 61.4 92.2

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050




Manual Integrations  
APPROVED


MMDadoda  
8/6/2021 5:40:59 PM



#10  
Methyl Iodide  
Concen: 55.138 ug/l  
RT: 2.453 min Scan# 224  
Delta R.T. -0.000 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40

Tgt Ion:142 Resp: 145492  
Ion Ratio Lower Upper  
142 100  
127 44.7 33.3 49.9  
141 14.3 11.4 17.0






#11

Tert butyl alcohol  
Concen: 254.147 ug/l  
RT: 2.983 min Scan# 311  
Delta R.T. 0.000 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050



Tgt Ion: 59 Resp: 154036  
Ion Ratio Lower Upper  
59 100  
57 10.2 7.8 11.8


Manual Integrations  
APPROVED

MMDadoda  
8/6/2021 5:40:59 PM






#12

1,1-Dichloroethene  
Concen: 46.757 ug/l  
RT: 2.319 min Scan# 202  
Delta R.T. -0.000 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40



Tgt Ion: 96 Resp: 97717  
Ion Ratio Lower Upper  
96 100  
61 158.4 124.6 187.0  
98 66.5 50.6 76.0









#13  
Acrolein  
Concen: 243.099 ug/l  
RT: 2.239 min Scan# 189  
Delta R.T. 0.000 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40

Tgt Ion: 56 Resp: 74891  
Ion Ratio Lower Upper  
56 100  
55 72.4 56.4 84.6

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050




Manual Integrations  
APPROVED


MMDadoda  
8/6/2021 5:40:59 PM



#14  
Allyl chloride  
Concen: 49.594 ug/l  
RT: 2.666 min Scan# 259  
Delta R.T. 0.000 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40


Tgt Ion: 41 Resp: 188436  
Ion Ratio Lower Upper  
41 100  
39 67.4 51.6 77.4  
76 31.9 30.5 45.7





#15  
Acrylonitrile  
Concen: 262.098 ug/l  
RT: 3.068 min Scan# 325  
Delta R.T. -0.007 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050



Tgt Ion: 53 Resp: 358717  
Ion Ratio Lower Upper  
53 100  
52 83.3 65.3 97.9  
51 36.6 28.2 42.2




Manual Integrations  
APPROVED


MMDadoda  
8/6/2021 5:40:59 PM



#16  
Acetone  
Concen: 293.770 ug/l  
RT: 2.386 min Scan# 213  
Delta R.T. -0.000 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40

Tgt Ion: 43 Resp: 382337  
Ion Ratio Lower Upper  
43 100  
58 30.4 27.8 41.8









#17  
Carbon Disulfide  
Concen: 42.231 ug/l  
RT: 2.508 min Scan# 233  
Delta R.T. -0.006 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40

Tgt Ion: 76 Resp: 217628  
Ion Ratio Lower Upper  
76 100  
78 9.3 7.2 10.8

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050




Manual Integrations  
APPROVED


MMDadoda  
8/6/2021 5:40:59 PM



#18  
Methyl Acetate  
Concen: 53.417 ug/l  
RT: 2.709 min Scan# 266  
Delta R.T. -0.000 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40

Tgt Ion: 43 Resp: 165267  
Ion Ratio Lower Upper  
43 100  
74 23.5 21.9 32.9









#19  
Methyl tert-butyl Ether  
Concen: 54.076 ug/l  
RT: 3.123 min Scan# 334  
Delta R.T. 0.000 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050

Tgt Ion: 73 Resp: 392943  
Ion Ratio Lower Upper  
73 100  
57 21.8 17.1 25.7




Manual Integrations  
APPROVED


MMDadoda  
8/6/2021 5:40:59 PM



#20  
Methylene Chloride  
Concen: 49.193 ug/l  
RT: 2.794 min Scan# 280  
Delta R.T. 0.000 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40

Tgt Ion: 84 Resp: 124536  
Ion Ratio Lower Upper  
84 100  
49 122.9 88.6 132.8  
51 38.0 27.0 40.6  
86 65.5 50.3 75.5









#21  
trans-1,2-Dichloroethene  
Concen: 47.794 ug/l  
RT: 3.093 min Scan# 329  
Delta R.T. -0.000 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40

Tgt Ion: 96 Resp: 106106  
Ion Ratio Lower Upper  
96 100  
61 144.1 107.7 161.5  
98 67.5 50.4 75.6

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050




Manual Integrations  
APPROVED


MMDadoda  
8/6/2021 5:40:59 PM



#22  
Diisopropyl ether  
Concen: 52.476 ug/l  
RT: 3.769 min Scan# 440  
Delta R.T. 0.000 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40

Tgt Ion: 45 Resp: 393584  
Ion Ratio Lower Upper  
45 100  
43 74.5 58.0 87.0  
87 28.7 25.4 38.0  
59 11.3 10.2 15.2









#23  
 Vinyl Acetate  
 Concen: 270.289 ug/l  
 RT: 3.727 min Scan# 433  
 Delta R.T. -0.006 min  
 Lab File: VX023545.D  
 Acq: 04 Aug 2021 16:40

Tgt Ion: 43 Resp: 1789575  
 Ion Ratio Lower Upper  
 43 100  
 86 10.1 9.9 14.9

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VSTDCCC050




Manual Integrations  
 APPROVED

MMDadoda  
 8/6/2021 5:40:59 PM




#24  
 1,1-Dichloroethane  
 Concen: 50.399 ug/l  
 RT: 3.611 min Scan# 414  
 Delta R.T. -0.006 min  
 Lab File: VX023545.D  
 Acq: 04 Aug 2021 16:40

Tgt Ion: 63 Resp: 207765  
 Ion Ratio Lower Upper  
 63 100  
 98 6.2 3.8 11.2  
 100 4.4 2.5 7.5








#25  
2-Butanone  
Concen: 275.519 ug/l  
RT: 4.568 min Scan# 571  
Delta R.T. 0.000 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050




Tgt Ion: 43 Resp: 556692  
Ion Ratio Lower Upper  
43 100  
72 25.2 23.7 35.5

Manual Integrations  
APPROVED



MMDadoda  
8/6/2021 5:40:59 PM




#26  
2,2-Dichloropropane  
Concen: 50.271 ug/l  
RT: 4.483 min Scan# 557  
Delta R.T. -0.000 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40



Tgt Ion: 77 Resp: 169497  
Ion Ratio Lower Upper  
77 100  
97 23.0 11.7 35.1





#27  
 cis-1,2-Dichloroethene  
 Concen: 50.685 ug/l  
 RT: 4.495 min Scan# 559  
 Delta R.T. -0.000 min  
 Lab File: VX023545.D  
 Acq: 04 Aug 2021 16:40





Tgt Ion: 96 Resp: 130679

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 96  | 100   |       |       |
| 61  | 147.1 | 0.0   | 285.6 |
| 98  | 66.1  | 0.0   | 129.6 |

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VSTDCCC050

Manual Integrations  
 APPROVED




MMDadoda  
 8/6/2021 5:40:59 PM




#28  
 Bromochloromethane  
 Concen: 48.245 ug/l  
 RT: 4.903 min Scan# 626  
 Delta R.T. -0.007 min  
 Lab File: VX023545.D  
 Acq: 04 Aug 2021 16:40

Tgt Ion: 49 Resp: 92819

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 49  | 100   |       |       |
| 129 | 2.1   | 0.0   | 4.0   |
| 130 | 80.6  | 68.2  | 102.2 |





#29

Tetrahydrofuran

Concen: 257.702 ug/l

RT: 5.019 min Scan# 645

Delta R.T. 0.006 min

Lab File: VX023545.D

Acq: 04 Aug 2021 16:40

Tgt Ion: 42 Resp: 341300

Ion Ratio Lower Upper

42 100

72 44.1 41.7 62.5

71 40.8 38.4 57.6

Instrument :

MSVOA\_X

ClientSampleId :





VSTDCCC050

Manual Integrations

APPROVED

MMDadoda

8/6/2021 5:40:59 PM



#30

Chloroform

Concen: 51.368 ug/l

RT: 5.099 min Scan# 658

Delta R.T. -0.006 min

Lab File: VX023545.D




Acq: 04 Aug 2021 16:40


Tgt Ion: 83 Resp: 216743

Ion Ratio Lower Upper

83 100


85 64.8 52.0 78.0





#31  
Cyclohexane  
Concen: 45.397 ug/l  
RT: 5.477 min Scan# 720  
Delta R.T. -0.000 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050




Tgt Ion: 56 Resp: 178423  
Ion Ratio Lower Upper  
56 100  
69 30.2 25.3 37.9  
84 84.1 71.4 107.0

Manual Integrations  
APPROVED



MMDadoda  
8/6/2021 5:40:59 PM




#32  
1,1,1-Trichloroethane  
Concen: 50.636 ug/l  
RT: 5.385 min Scan# 705  
Delta R.T. -0.006 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40



Tgt Ion: 97 Resp: 182598  
Ion Ratio Lower Upper  
97 100  
99 63.5 51.5 77.3  
61 46.4 34.2 51.4





#33

1,2-Dichloroethane-d4

Concen: 48.670 ug/l

RT: 5.964 min Scan# 800

Delta R.T. 0.000 min

Lab File: VX023545.D





Acq: 04 Aug 2021 16:40

Tgt Ion: 65 Resp: 129129

Ion Ratio Lower Upper

65 100

67 52.7 0.0 105.4



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.769 min Scan# 932

Delta R.T. -0.000 min

Lab File: VX023545.D

Acq: 04 Aug 2021 16:40




Tgt Ion: 114 Resp: 344599

Ion Ratio Lower Upper

114 100

63 20.2 0.0 40.8

88 16.1 0.0 32.8



Instrument :

MSVOA\_X


ClientSampleId :

VSTDCCC050


Manual Integrations

APPROVED



MMDadoda  
8/6/2021 5:40:59 PM



#35  
Dibromofluoromethane  
Concen: 46.942 ug/l  
RT: 5.397 min Scan# 707  
Delta R.T. 0.000 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40




Tgt Ion: 113 Resp: 101140  
Ion Ratio Lower Upper  
113 100  
111 102.1 81.9 122.9  
192 20.2 16.7 25.1




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050

Manual Integrations  
APPROVED



MMDadoda  
8/6/2021 5:40:59 PM




#36  
1,1-Dichloropropene  
Concen: 48.682 ug/l  
RT: 5.696 min Scan# 756  
Delta R.T. -0.000 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40




Tgt Ion: 75 Resp: 151754  
Ion Ratio Lower Upper  
75 100  
110 36.5 18.1 54.4  
77 31.3 25.3 37.9





#37  
Ethyl Acetate  
Concen: 50.953 ug/l  
RT: 4.727 min Scan# 597  
Delta R.T. -0.000 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050




Tgt Ion: 43 Resp: 200266  
Ion Ratio Lower Upper  
43 100  
61 14.0 12.8 19.2  
70 10.8 10.5 15.7

Manual Integrations  
APPROVED



MMDadoda  
8/6/2021 5:40:59 PM




#38  
Carbon Tetrachloride  
Concen: 49.107 ug/l  
RT: 5.684 min Scan# 754  
Delta R.T. -0.000 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40



Tgt Ion: 117 Resp: 155678  
Ion Ratio Lower Upper  
117 100  
119 95.3 76.2 114.2  
121 31.3 24.8 37.2





#39

Methylcyclohexane

Concen: 48.218 ug/l

RT: 7.385 min Scan# 1033

Delta R.T. -0.000 min

Lab File: VX023545.D

Acq: 04 Aug 2021 16:40

Tgt Ion: 83 Resp: 189825

Ion Ratio Lower Upper

83 100

55 74.9 58.9 88.3

98 48.4 38.1 57.1

Instrument :

MSVOA\_X

ClientSampleId :





VSTDCCC050

Manual Integrations

APPROVED

MMDadoda

8/6/2021 5:40:59 PM



#40

Benzene

Concen: 49.675 ug/l

RT: 6.043 min Scan# 813

Delta R.T. 0.000 min

Lab File: VX023545.D




Acq: 04 Aug 2021 16:40

Tgt Ion: 78 Resp: 459124


Ion Ratio Lower Upper

78 100

77 24.4 18.9 28.3











#41  
Methacrylonitrile  
Concen: 52.786 ug/l  
RT: 4.928 min Scan# 630  
Delta R.T. -0.000 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40

Tgt Ion: 41 Resp: 109546  
Ion Ratio Lower Upper  
41 100  
39 57.5 42.5 63.7  
67 70.7 65.0 97.6  
52 30.6 25.3 37.9

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050




Manual Integrations  
APPROVED


MMDadoda  
8/6/2021 5:40:59 PM




#42  
1,2-Dichloroethane  
Concen: 53.735 ug/l  
RT: 6.092 min Scan# 821  
Delta R.T. -0.006 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40

Tgt Ion: 62 Resp: 183633  
Ion Ratio Lower Upper  
62 100  
98 8.3 0.0 20.2







#43  
Isopropyl Acetate  
Concen: 53.105 ug/l  
RT: 6.348 min Scan# 863  
Delta R.T. 0.000 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40




Tgt Ion: 43 Resp: 323428  
Ion Ratio Lower Upper  
43 100  
61 19.3 18.2 27.4  
87 13.3 12.5 18.7




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050

Manual Integrations  
APPROVED



MMDadoda  
8/6/2021 5:40:59 PM




#44  
Trichloroethene  
Concen: 49.633 ug/l  
RT: 7.129 min Scan# 991  
Delta R.T. -0.000 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40



Tgt Ion:130 Resp: 126026  
Ion Ratio Lower Upper  
130 100  
95 97.1 0.0 198.0





#45





1,2-Dichloropropane  
Concen: 51.918 ug/l  
RT: 7.433 min Scan# 1041  
Delta R.T. -0.007 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40

Tgt Ion: 63 Resp: 126978  
Ion Ratio Lower Upper  
63 100  
65 31.9 25.7 38.5

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050

Manual Integrations  
APPROVED




MMDadoda  
8/6/2021 5:40:59 PM




#46

Dibromomethane  
Concen: 54.089 ug/l  
RT: 7.586 min Scan# 1066  
Delta R.T. -0.000 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40

Tgt Ion: 93 Resp: 89731  
Ion Ratio Lower Upper  
93 100  
95 82.9 66.8 100.2  
174 101.5 83.1 124.7









#47  
Bromodichloromethane  
Concen: 54.382 ug/l  
RT: 7.824 min Scan# 1105  
Delta R.T. -0.006 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40

Tgt Ion: 83 Resp: 162539  
Ion Ratio Lower Upper  
83 100  
85 64.2 51.4 77.0  
127 8.7 6.8 10.2

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050




Manual Integrations  
APPROVED


MMDadoda  
8/6/2021 5:40:59 PM



#48  
Methyl methacrylate  
Concen: 55.879 ug/l  
RT: 7.696 min Scan# 1084  
Delta R.T. -0.006 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40

Tgt Ion: 41 Resp: 157615  
Ion Ratio Lower Upper  
41 100  
69 79.7 76.0 114.0  
39 61.6 44.6 67.0









#49  
1,4-Dioxane  
Concen: 1073.488 ug/l  
RT: 7.665 min Scan# 1079  
Delta R.T. 0.000 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40

Tgt Ion: 88 Resp: 64476  
Ion Ratio Lower Upper  
88 100  
43 35.1 23.2 34.8#  
58 68.3 55.0 82.4

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050




Manual Integrations  
APPROVED


MMDadoda  
8/6/2021 5:40:59 PM



#50  
Toluene-d8  
Concen: 45.978 ug/l  
RT: 8.653 min Scan# 1241  
Delta R.T. -0.000 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40

Tgt Ion: 98 Resp: 363964  
Ion Ratio Lower Upper  
98 100  
100 64.4 51.9 77.9









#51  
4-Methyl-2-Pentanone  
Concen: 284.218 ug/l  
RT: 8.580 min Scan# 1229  
Delta R.T. -0.000 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40

Tgt Ion: 43 Resp: 1058756  
Ion Ratio Lower Upper  
43 100  
58 38.3 35.4 53.0

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050




Manual Integrations  
APPROVED


MMDadoda  
8/6/2021 5:40:59 PM



#52  
Toluene  
Concen: 51.596 ug/l  
RT: 8.720 min Scan# 1252  
Delta R.T. -0.000 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40

Tgt Ion: 92 Resp: 295442  
Ion Ratio Lower Upper  
92 100  
91 174.4 137.1 205.7









#53  
t-1,3-Dichloropropene  
Concen: 50.365 ug/l  
RT: 8.982 min Scan# 1295  
Delta R.T. -0.000 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40

Tgt Ion: 75 Resp: 190972  
Ion Ratio Lower Upper  
75 100  
77 32.8 24.9 37.3

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050




Manual Integrations  
APPROVED


MMDadoda  
8/6/2021 5:40:59 PM



#54  
cis-1,3-Dichloropropene  
Concen: 54.131 ug/l  
RT: 8.372 min Scan# 1195  
Delta R.T. 0.000 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40

Tgt Ion: 75 Resp: 201685  
Ion Ratio Lower Upper  
75 100  
77 31.6 25.7 38.5  
39 52.4 35.1 52.7





#55  
1,1,2-Trichloroethane  
Concen: 56.898 ug/l  
RT: 9.153 min Scan# 1323  
Delta R.T. -0.006 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40





Tgt Ion: 97 Resp: 131931

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 97  | 100   |       |       |
| 83  | 85.2  | 69.3  | 103.9 |
| 85  | 56.7  | 44.3  | 66.5  |
| 99  | 62.5  | 50.3  | 75.5  |

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050

Manual Integrations  
APPROVED




MMDadoda  
8/6/2021 5:40:59 PM




#56  
Ethyl methacrylate  
Concen: 57.359 ug/l  
RT: 9.122 min Scan# 1318  
Delta R.T. 0.000 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40

Tgt Ion: 69 Resp: 213191

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 69  | 100   |       |       |
| 41  | 70.6  | 46.8  | 70.2# |
| 39  | 43.0  | 24.7  | 37.1# |














#57  
1,3-Dichloropropane  
Concen: 54.807 ug/l  
RT: 9.311 min Scan# 1349  
Delta R.T. 0.000 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40

Tgt Ion: 76 Resp: 215938  
Ion Ratio Lower Upper  
76 100  
78 32.3 26.2 39.2



#58  
2-Chloroethyl Vinyl ether  
Concen: 264.111 ug/l  
RT: 8.244 min Scan# 1174  
Delta R.T. -0.006 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40


Tgt Ion: 63 Resp: 520020  
Ion Ratio Lower Upper  
63 100  
106 27.8 23.5 35.3



Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050

Manual Integrations  
APPROVED

MMDadoda  
8/6/2021 5:40:59 PM







#59  
2-Hexanone  
Concen: 291.532 ug/l  
RT: 9.433 min Scan# 1369  
Delta R.T. 0.000 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40

Tgt Ion: 43 Resp: 824685  
Ion Ratio Lower Upper  
43 100  
58 52.7 30.9 92.8

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050




Manual Integrations  
APPROVED


MMDadoda  
8/6/2021 5:40:59 PM



#60  
Dibromochloromethane  
Concen: 51.779 ug/l  
RT: 9.525 min Scan# 1384  
Delta R.T. -0.000 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40

Tgt Ion: 129 Resp: 133256  
Ion Ratio Lower Upper  
129 100  
127 77.4 38.4 115.2









#61  
1,2-Dibromoethane  
Concen: 56.192 ug/l  
RT: 9.616 min Scan# 1399  
Delta R.T. 0.000 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050

Tgt Ion:107 Resp: 139011  
Ion Ratio Lower Upper  
107 100  
109 93.8 75.8 113.6




Manual Integrations  
APPROVED


MMDadoda  
8/6/2021 5:40:59 PM



#62  
4-Bromofluorobenzene  
Concen: 52.211 ug/l  
RT: 11.085 min Scan# 1640  
Delta R.T. 0.000 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40


Tgt Ion: 95 Resp: 140540  
Ion Ratio Lower Upper  
95 100  
174 81.4 0.0 154.6  
176 76.8 0.0 155.8





#63  
Chlorobenzene-d5  
Concen: 50.000 ug/l  
RT: 10.055 min Scan# 1471  
Delta R.T. -0.000 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050




Tgt Ion:117 Resp: 325772  
Ion Ratio Lower Upper  
117 100  
82 56.8 47.4 71.2  
119 32.1 26.6 39.8

Manual Integrations  
APPROVED



MMDadoda  
8/6/2021 5:40:59 PM




#64  
Tetrachloroethene  
Concen: 47.617 ug/l  
RT: 9.275 min Scan# 1343  
Delta R.T. -0.000 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40



Tgt Ion:164 Resp: 123458  
Ion Ratio Lower Upper  
164 100  
166 122.7 103.4 155.2  
129 97.3 70.8 106.2  
131 93.5 72.5 108.7









#65  
Chlorobenzene  
Concen: 49.861 ug/l  
RT: 10.079 min Scan# 1475  
Delta R.T. -0.006 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050

Tgt Ion:112 Resp: 332459  
Ion Ratio Lower Upper  
112 100  
114 33.0 25.8 38.6




Manual Integrations  
APPROVED


MMDadoda  
8/6/2021 5:40:59 PM



#66  
1,1,1,2-Tetrachloroethane  
Concen: 49.078 ug/l  
RT: 10.165 min Scan# 1489  
Delta R.T. -0.000 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40

Tgt Ion:131 Resp: 125697  
Ion Ratio Lower Upper  
131 100  
133 94.2 47.7 143.1  
119 65.6 32.1 96.5







#67  
Ethyl Benzene  
Concen: 50.303 ug/l  
RT: 10.195 min Scan# 1494  
Delta R.T. 0.000 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40

Tgt Ion: 91 Resp: 581711  
Ion Ratio Lower Upper  
91 100  
106 31.4 24.6 37.0

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050


Manual Integrations  
APPROVED


MMDadoda  
8/6/2021 5:40:59 PM



#68  
m/p-Xylenes  
Concen: 100.570 ug/l  
RT: 10.305 min Scan# 1512  
Delta R.T. -0.000 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40

Tgt Ion:106 Resp: 447283  
Ion Ratio Lower Upper  
106 100  
91 202.3 162.8 244.2









#69  
o-Xylene  
Concen: 50.970 ug/l  
RT: 10.646 min Scan# 1568  
Delta R.T. 0.000 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050

Tgt Ion:106 Resp: 224748  
Ion Ratio Lower Upper  
106 100  
91 212.7 108.1 324.2




Manual Integrations  
APPROVED


MMDadoda  
8/6/2021 5:40:59 PM




#70  
Styrene  
Concen: 53.746 ug/l  
RT: 10.659 min Scan# 1570  
Delta R.T. -0.001 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40

Tgt Ion:104 Resp: 375482  
Ion Ratio Lower Upper  
104 100  
78 51.0 40.0 60.0  
103 55.0 43.4 65.0









#71  
Bromoform  
Concen: 48.394 ug/l  
RT: 10.805 min Scan# 1594  
Delta R.T. -0.000 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40





Tgt Ion:173 Resp: 96940  
Ion Ratio Lower Upper  
173 100  
175 48.3 23.9 71.8  
254 0.0 0.1 0.1#



#72  
1,4-Dichlorobenzene-d4  
Concen: 50.000 ug/l  
RT: 12.024 min Scan# 1794  
Delta R.T. 0.000 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40



Tgt Ion:152 Resp: 153231  
Ion Ratio Lower Upper  
152 100  
115 85.7 41.1 123.3  
150 174.0 0.0 349.0




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050

Manual Integrations  
APPROVED

MMDadoda  
8/6/2021 5:40:59 PM









#73  
Isopropylbenzene  
Concen: 50.148 ug/l  
RT: 10.963 min Scan# 1620  
Delta R.T. 0.000 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40

Tgt Ion: 105 Resp: 579833  
Ion Ratio Lower Upper  
105 100  
120 27.1 13.2 39.6

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050




Manual Integrations  
APPROVED


MMDadoda  
8/6/2021 5:40:59 PM



#74  
N-ethyl acetate  
Concen: 53.609 ug/l  
RT: 10.848 min Scan# 1601  
Delta R.T. -0.001 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40

Tgt Ion: 43 Resp: 282787  
Ion Ratio Lower Upper  
43 100  
70 39.9 39.0 58.6  
55 22.9 21.5 32.3  
61 23.0 22.1 33.1









#75  
1,1,2,2-Tetrachloroethane  
Concen: 51.204 ug/l  
RT: 11.213 min Scan# 1661  
Delta R.T. 0.000 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40

Tgt Ion: 83 Resp: 192859  
Ion Ratio Lower Upper  
83 100  
131 10.3 5.1 15.2  
85 64.1 31.8 95.4

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050




Manual Integrations  
APPROVED


MMDadoda  
8/6/2021 5:40:59 PM



#76  
1,2,3-Trichloropropane  
Concen: 54.712 ug/l m  
RT: 11.244 min Scan# 1666  
Delta R.T. -0.000 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40

Tgt Ion: 75 Resp: 184485  
Ion Ratio Lower Upper  
75 100  
77 39.8 19.1 57.5









#77  
Bromobenzene  
Concen: 52.736 ug/l  
RT: 11.201 min Scan# 1659  
Delta R.T. 0.000 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40

Tgt Ion: 156 Resp: 147092  
Ion Ratio Lower Upper  
156 100  
77 158.8 82.8 248.6  
158 97.6 49.1 147.4

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050




Manual Integrations  
APPROVED


MMDadoda  
8/6/2021 5:40:59 PM



#78  
n-propylbenzene  
Concen: 51.027 ug/l  
RT: 11.305 min Scan# 1676  
Delta R.T. -0.006 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40

Tgt Ion: 91 Resp: 667741  
Ion Ratio Lower Upper  
91 100  
120 23.6 11.5 34.5









#79  
2-Chlorotoluene  
Concen: 51.418 ug/l  
RT: 11.366 min Scan# 1686  
Delta R.T. -0.000 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050

Tgt Ion: 91 Resp: 402066  
Ion Ratio Lower Upper  
91 100  
126 34.6 16.6 49.8




Manual Integrations  
APPROVED


MMDadoda  
8/6/2021 5:40:59 PM



#80  
1,3,5-Trimethylbenzene  
Concen: 51.600 ug/l  
RT: 11.457 min Scan# 1701  
Delta R.T. 0.000 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40


Tgt Ion: 105 Resp: 485882  
Ion Ratio Lower Upper  
105 100  
120 48.3 23.8 71.5





#81  
trans-1,4-Dichloro-2-butene  
Concen: 52.512 ug/l  
RT: 11.018 min Scan# 1629  
Delta R.T. -0.006 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050




Tgt Ion: 75 Resp: 65852  
Ion Ratio Lower Upper  
75 100  
53 104.2 74.5 111.7  
89 46.6 37.9 56.9

Manual Integrations  
APPROVED



MMDadoda  
8/6/2021 5:40:59 PM




#82  
4-Chlorotoluene  
Concen: 52.300 ug/l  
RT: 11.457 min Scan# 1701  
Delta R.T. 0.000 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40




Tgt Ion: 91 Resp: 462668  
Ion Ratio Lower Upper  
91 100  
126 29.9 14.9 44.5





#83  
tert-Butylbenzene  
Concen: 51.134 ug/l  
RT: 11.719 min Scan# 1744  
Delta R.T. 0.000 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050




Tgt Ion:119 Resp: 472301  
Ion Ratio Lower Upper  
119 100  
91 59.2 27.9 83.7  
134 23.5 11.3 33.9

Manual Integrations  
APPROVED



MMDadoda  
8/6/2021 5:40:59 PM




#84  
1,2,4-Trimethylbenzene  
Concen: 51.944 ug/l  
RT: 11.756 min Scan# 1750  
Delta R.T. -0.000 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40



Tgt Ion:105 Resp: 486521  
Ion Ratio Lower Upper  
105 100  
120 44.9 22.4 67.3









#85  
 sec-Butylbenzene  
 Concen: 51.049 ug/l  
 RT: 11.896 min Scan# 1773  
 Delta R.T. 0.000 min  
 Lab File: VX023545.D  
 Acq: 04 Aug 2021 16:40

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VSTDCCC050

Tgt Ion:105 Resp: 611372  
 Ion Ratio Lower Upper  
 105 100  
 134 20.0 10.0 29.8




Manual Integrations  
 APPROVED


MMDadoda  
 8/6/2021 5:40:59 PM




#86  
 p-Isopropyltoluene  
 Concen: 52.301 ug/l  
 RT: 12.012 min Scan# 1792  
 Delta R.T. -0.000 min  
 Lab File: VX023545.D  
 Acq: 04 Aug 2021 16:40

Tgt Ion:119 Resp: 505811  
 Ion Ratio Lower Upper  
 119 100  
 134 26.5 12.9 38.7  
 91 24.8 11.7 35.0









#87  
1,3-Dichlorobenzene  
Concen: 51.456 ug/l  
RT: 11.969 min Scan# 1785  
Delta R.T. -0.006 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40





Tgt Ion:146 Resp: 265519  
Ion Ratio Lower Upper  
146 100  
111 39.8 20.2 60.5  
148 63.5 31.9 95.7



#88  
1,4-Dichlorobenzene  
Concen: 50.469 ug/l  
RT: 12.042 min Scan# 1797  
Delta R.T. 0.000 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40



Tgt Ion:146 Resp: 264254  
Ion Ratio Lower Upper  
146 100  
111 38.9 19.6 58.8  
148 63.2 31.6 95.0




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050

Manual Integrations  
APPROVED


MMDadoda  
8/6/2021 5:40:59 PM





#89  
n-Butylbenzene  
Concen: 51.810 ug/l  
RT: 12.335 min Scan# 1845  
Delta R.T. 0.000 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050




Tgt Ion: 91 Resp: 453590  
Ion Ratio Lower Upper  
91 100  
92 52.2 27.4 82.0  
134 25.9 13.2 39.5

Manual Integrations  
APPROVED



MMDadoda  
8/6/2021 5:40:59 PM




#90  
Hexachloroethane  
Concen: 47.477 ug/l  
RT: 12.542 min Scan# 1879  
Delta R.T. 0.000 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40



Tgt Ion: 117 Resp: 82062  
Ion Ratio Lower Upper  
117 100  
201 76.5 40.4 121.2









#91  
1,2-Dichlorobenzene  
Concen: 50.879 ug/l  
RT: 12.335 min Scan# 1845  
Delta R.T. -0.006 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40

Tgt Ion: 146 Resp: 254211  
Ion Ratio Lower Upper  
146 100  
111 41.7 20.8 62.2  
148 64.5 31.9 95.5

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050




Manual Integrations  
APPROVED


MMDadoda  
8/6/2021 5:40:59 PM



#92  
1,2-Dibromo-3-Chloropropane  
Concen: 46.225 ug/l  
RT: 12.945 min Scan# 1945  
Delta R.T. -0.000 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40


Tgt Ion: 75 Resp: 42086  
Ion Ratio Lower Upper  
75 100  
155 83.6 45.3 135.9  
157 107.7 57.7 173.1





#93  
1,2,4-Trichlorobenzene  
Concen: 52.172 ug/l  
RT: 13.591 min Scan# 2051  
Delta R.T. -0.000 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050




Tgt Ion:180 Resp: 163776  
Ion Ratio Lower Upper  
180 100  
182 94.3 47.7 143.1  
145 31.0 15.0 45.0

Manual Integrations  
APPROVED



MMDadoda  
8/6/2021 5:40:59 PM




#94  
Hexachlorobutadiene  
Concen: 48.991 ug/l  
RT: 13.731 min Scan# 2074  
Delta R.T. 0.000 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40




Tgt Ion:225 Resp: 69016  
Ion Ratio Lower Upper  
225 100  
223 64.4 31.1 93.2  
227 63.3 31.6 95.0





#95  
Naphthalene  
Concen: 53.439 ug/l  
RT: 13.780 min Scan# 2082  
Delta R.T. -0.000 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC050




Tgt Ion:128 Resp: 576770  
Ion Ratio Lower Upper  
128 100  
127 12.7 10.3 15.5  
129 10.8 8.8 13.2

Manual Integrations  
APPROVED

MMDadoda  
8/6/2021 5:40:59 PM



#96  
1,2,3-Trichlorobenzene  
Concen: 52.331 ug/l  
RT: 13.963 min Scan# 2112  
Delta R.T. -0.000 min  
Lab File: VX023545.D  
Acq: 04 Aug 2021 16:40



Tgt Ion:180 Resp: 165886  
Ion Ratio Lower Upper  
180 100  
182 95.4 47.6 142.9  
145 33.4 16.1 48.3

