

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX080521\  
 Data File : VX023547.D  
 Acq On : 04 Aug 2021 17:59  
 Operator : JC/MD  
 Sample : VX0805WBS01  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 5 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX0805WBS01

Manual Integrations  
 APPROVED

MMDadoda  
 8/6/2021 5:41:01 PM

Quant Time: Aug 05 06:08:39 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X072221W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Jul 23 06:05:38 2021  
 Response via : Initial Calibration

| Compound                     | R.T.           | QIon | Response            | Conc    | Units  | Dev(Min) |
|------------------------------|----------------|------|---------------------|---------|--------|----------|
| -----                        |                |      |                     |         |        |          |
| Internal Standards           |                |      |                     |         |        |          |
| 1) Pentafluorobenzene        | 5.562          | 168  | 200800              | 50.000  | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene      | 6.769          | 114  | 326433              | 50.000  | ug/l   | 0.00     |
| 63) Chlorobenzene-d5         | 10.055         | 117  | 300596              | 50.000  | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4   | 12.024         | 152  | 135225              | 50.000  | ug/l   | 0.00     |
|                              |                |      |                     |         |        |          |
| System Monitoring Compounds  |                |      |                     |         |        |          |
| 33) 1,2-Dichloroethane-d4    | 5.964          | 65   | 138626              | 55.344  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 78 - 117 |      | Recovery = 110.680% |         |        |          |
| 35) Dibromofluoromethane     | 5.397          | 113  | 110675              | 54.226  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 75 - 124 |      | Recovery = 108.460% |         |        |          |
| 50) Toluene-d8               | 8.652          | 98   | 403359              | 53.791  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 92 - 112 |      | Recovery = 107.580% |         |        |          |
| 62) 4-Bromofluorobenzene     | 11.085         | 95   | 146671              | 57.521  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 83 - 123 |      | Recovery = 115.040% |         |        |          |
|                              |                |      |                     |         |        |          |
| Target Compounds             |                |      |                     |         | Qvalue |          |
| 2) Dichlorodifluoromethane   | 1.172          | 85   | 40309               | 21.820  | ug/l   | 91       |
| 3) Chloromethane             | 1.294          | 50   | 45814               | 21.289  | ug/l   | 100      |
| 4) Vinyl Chloride            | 1.373          | 62   | 46817               | 21.492  | ug/l   | 99       |
| 5) Bromomethane              | 1.599          | 94   | 33592               | 25.974  | ug/l   | 98       |
| 6) Chloroethane              | 1.684          | 64   | 30436               | 22.766  | ug/l   | 99       |
| 7) Trichlorofluoromethane    | 1.885          | 101  | 75184               | 21.422  | ug/l   | 99       |
| 8) Diethyl Ether             | 2.141          | 74   | 27923               | 21.556  | ug/l   | 92       |
| 9) 1,1,2-Trichlorotrifluo... | 2.330          | 101  | 43522               | 21.889  | ug/l   | 96       |
| 10) Methyl Iodide            | 2.452          | 142  | 52265               | 20.980  | ug/l   | 97       |
| 11) Tert butyl alcohol       | 2.977          | 59   | 63572               | 111.101 | ug/l   | 97       |
| 12) 1,1-Dichloroethene       | 2.318          | 96   | 41568               | 21.068  | ug/l   | 95       |
| 13) Acrolein                 | 2.239          | 56   | 24073               | 83.670  | ug/l   | 95       |
| 14) Allyl chloride           | 2.666          | 41   | 75125               | 20.943  | ug/l   | 93       |
| 15) Acrylonitrile            | 3.068          | 53   | 146556              | 113.424 | ug/l   | 98       |
| 16) Acetone                  | 2.385          | 43   | 132932              | 108.188 | ug/l   | 95       |
| 17) Carbon Disulfide         | 2.513          | 76   | 88526               | 18.196  | ug/l   | 100      |
| 18) Methyl Acetate           | 2.708          | 43   | 67958               | 23.266  | ug/l   | 92       |
| 19) Methyl tert-butyl Ether  | 3.123          | 73   | 150158              | 21.888  | ug/l   | 98       |
| 20) Methylene Chloride       | 2.794          | 84   | 50459               | 21.112  | ug/l # | 88       |
| 21) trans-1,2-Dichloroethene | 3.099          | 96   | 44539               | 21.250  | ug/l   | 100      |
| 22) Diisopropyl ether        | 3.769          | 45   | 157410              | 22.230  | ug/l   | 91       |
| 23) Vinyl Acetate            | 3.733          | 43   | 695761              | 111.308 | ug/l   | 95       |
| 24) 1,1-Dichloroethane       | 3.617          | 63   | 85413               | 21.946  | ug/l   | 97       |
| 25) 2-Butanone               | 4.568          | 43   | 210967              | 110.596 | ug/l   | 93       |
| 26) 2,2-Dichloropropane      | 4.489          | 77   | 67557               | 21.224  | ug/l   | 99       |
| 27) cis-1,2-Dichloroethene   | 4.501          | 96   | 52428               | 21.539  | ug/l   | 96       |
| 28) Bromochloromethane       | 4.915          | 49   | 39911               | 21.973  | ug/l   | 92       |
| 29) Tetrahydrofuran          | 5.019          | 42   | 139467              | 111.543 | ug/l   | 89       |
| 30) Chloroform               | 5.104          | 83   | 88141               | 22.127  | ug/l   | 99       |
| 31) Cyclohexane              | 5.476          | 56   | 72982               | 19.669  | ug/l   | 96       |
| 32) 1,1,1-Trichloroethane    | 5.397          | 97   | 75510               | 22.180  | ug/l   | 98       |
| 36) 1,1-Dichloropropene      | 5.696          | 75   | 62598               | 21.199  | ug/l   | 100      |
| 37) Ethyl Acetate            | 4.726          | 43   | 79844               | 21.445  | ug/l # | 95       |
| 38) Carbon Tetrachloride     | 5.690          | 117  | 63455               | 21.130  | ug/l   | 96       |
| 39) Methylcyclohexane        | 7.384          | 83   | 77708               | 20.837  | ug/l   | 95       |
| 40) Benzene                  | 6.043          | 78   | 189696              | 21.666  | ug/l   | 98       |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX080521\  
 Data File : VX023547.D  
 Acq On : 04 Aug 2021 17:59  
 Operator : JC/MD  
 Sample : VX0805WBS01  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 5 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX0805WBS01

Manual Integrations  
 APPROVED

MMDadoda  
 8/6/2021 5:41:01 PM

Quant Time: Aug 05 06:08:39 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X072221W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Jul 23 06:05:38 2021  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc    | Units  | Dev(Min) |
|-------------------------------|--------|------|----------|---------|--------|----------|
| 41) Methacrylonitrile         | 4.928  | 41   | 42437    | 21.587  | ug/l   | 92       |
| 42) 1,2-Dichloroethane        | 6.098  | 62   | 73704    | 22.768  | ug/l   | 95       |
| 43) Isopropyl Acetate         | 6.348  | 43   | 126725   | 21.965  | ug/l   | 93       |
| 44) Trichloroethene           | 7.134  | 130  | 51676    | 21.484  | ug/l   | 91       |
| 45) 1,2-Dichloropropane       | 7.439  | 63   | 50097    | 21.623  | ug/l   | 99       |
| 46) Dibromomethane            | 7.586  | 93   | 35168    | 22.379  | ug/l   | 99       |
| 47) Bromodichloromethane      | 7.829  | 83   | 61840    | 21.842  | ug/l   | 98       |
| 48) Methyl methacrylate       | 7.701  | 41   | 60633    | 22.692  | ug/l   | 88       |
| 49) 1,4-Dioxane               | 7.665  | 88   | 26839    | 471.721 | ug/l   | 98       |
| 51) 4-Methyl-2-Pentanone      | 8.579  | 43   | 423382   | 119.980 | ug/l   | 90       |
| 52) Toluene                   | 8.720  | 92   | 122570   | 22.597  | ug/l   | 99       |
| 53) t-1,3-Dichloropropene     | 8.982  | 75   | 69977    | 20.596  | ug/l   | 96       |
| 54) cis-1,3-Dichloropropene   | 8.372  | 75   | 77015    | 21.821  | ug/l   | 92       |
| 55) 1,1,2-Trichloroethane     | 9.158  | 97   | 51745    | 23.558  | ug/l   | 99       |
| 56) Ethyl methacrylate        | 9.122  | 69   | 80532    | 22.873  | ug/l # | 83       |
| 57) 1,3-Dichloropropane       | 9.311  | 76   | 85405    | 22.883  | ug/l   | 99       |
| 58) 2-Chloroethyl Vinyl ether | 8.244  | 63   | 186865   | 100.188 | ug/l   | 97       |
| 59) 2-Hexanone                | 9.433  | 43   | 328421   | 122.560 | ug/l   | 88       |
| 60) Dibromochloromethane      | 9.524  | 129  | 47683    | 20.793  | ug/l   | 99       |
| 61) 1,2-Dibromoethane         | 9.616  | 107  | 54182    | 23.121  | ug/l   | 99       |
| 64) Tetrachloroethene         | 9.274  | 164  | 50784    | 21.228  | ug/l   | 94       |
| 65) Chlorobenzene             | 10.085 | 112  | 132633   | 21.558  | ug/l   | 99       |
| 66) 1,1,1,2-Tetrachloroethane | 10.164 | 131  | 48815    | 21.220  | ug/l   | 100      |
| 67) Ethyl Benzene             | 10.195 | 91   | 233415   | 21.875  | ug/l   | 99       |
| 68) m/p-Xylenes               | 10.305 | 106  | 175994   | 42.886  | ug/l   | 100      |
| 69) o-Xylene                  | 10.646 | 106  | 88732    | 21.809  | ug/l   | 97       |
| 70) Styrene                   | 10.658 | 104  | 144251   | 22.377  | ug/l   | 100      |
| 71) Bromoform                 | 10.805 | 173  | 33430    | 20.291  | ug/l # | 99       |
| 73) Isopropylbenzene          | 10.963 | 105  | 230176   | 22.558  | ug/l   | 98       |
| 74) N-amyl acetate            | 10.847 | 43   | 105114   | 22.580  | ug/l   | 90       |
| 75) 1,1,2,2-Tetrachloroethane | 11.213 | 83   | 77609    | 23.349  | ug/l   | 99       |
| 76) 1,2,3-Trichloropropane    | 11.243 | 75   | 71740m   | 24.108  | ug/l   |          |
| 77) Bromobenzene              | 11.201 | 156  | 57183    | 23.231  | ug/l   | 94       |
| 78) n-propylbenzene           | 11.304 | 91   | 259687   | 22.487  | ug/l   | 100      |
| 79) 2-Chlorotoluene           | 11.365 | 91   | 155762   | 22.572  | ug/l   | 97       |
| 80) 1,3,5-Trimethylbenzene    | 11.457 | 105  | 189756   | 22.835  | ug/l   | 99       |
| 81) trans-1,4-Dichloro-2-b... | 11.024 | 75   | 22002    | 19.881  | ug/l # | 85       |
| 82) 4-Chlorotoluene           | 11.457 | 91   | 176347   | 22.589  | ug/l   | 99       |
| 83) tert-Butylbenzene         | 11.719 | 119  | 186920   | 22.932  | ug/l   | 97       |
| 84) 1,2,4-Trimethylbenzene    | 11.756 | 105  | 191977   | 23.226  | ug/l   | 99       |
| 85) sec-Butylbenzene          | 11.896 | 105  | 235250   | 22.259  | ug/l   | 99       |
| 86) p-Isopropyltoluene        | 12.012 | 119  | 192793   | 22.589  | ug/l   | 98       |
| 87) 1,3-Dichlorobenzene       | 11.975 | 146  | 100138   | 21.990  | ug/l   | 99       |
| 88) 1,4-Dichlorobenzene       | 12.042 | 146  | 100055   | 21.654  | ug/l   | 99       |
| 89) n-Butylbenzene            | 12.335 | 91   | 165300   | 21.395  | ug/l   | 98       |
| 90) Hexachloroethane          | 12.542 | 117  | 29208    | 20.281  | ug/l   | 95       |
| 91) 1,2-Dichlorobenzene       | 12.341 | 146  | 99680    | 22.607  | ug/l   | 99       |
| 92) 1,2-Dibromo-3-Chloropr... | 12.944 | 75   | 15586    | 20.628  | ug/l   | 91       |
| 93) 1,2,4-Trichlorobenzene    | 13.591 | 180  | 59469    | 21.467  | ug/l   | 97       |
| 94) Hexachlorobutadiene       | 13.725 | 225  | 26086    | 20.983  | ug/l   | 100      |
| 95) Naphthalene               | 13.780 | 128  | 208676   | 21.909  | ug/l   | 99       |
| 96) 1,2,3-Trichlorobenzene    | 13.962 | 180  | 59075    | 21.117  | ug/l   | 100      |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX080521\  
Data File : VX023547.D  
Acq On : 04 Aug 2021 17:59  
Operator : JC/MD  
Sample : VX0805WBS01  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 5 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0805WBS01

Manual Integrations  
APPROVED

MMDadoda  
8/6/2021 5:41:01 PM


Quant Time: Aug 05 06:08:39 2021  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X072221W.M  
Quant Title : SW846 8260  
QLast Update : Fri Jul 23 06:05:38 2021  
Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|----------|------|------|----------|------|-------|----------|
|----------|------|------|----------|------|-------|----------|


-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed







#1  
Pentafluorobenzene  
Concen: 50.000 ug/l  
RT: 5.562 min Scan# 734  
Delta R.T. -0.000 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59




Tgt Ion: 168 Resp: 200800  
Ion Ratio Lower Upper  
168 100  
99 57.6 45.8 68.6




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0805WBS01

Manual Integrations  
APPROVED



MMDadoda  
8/6/2021 5:41:01 PM




#2  
Dichlorodifluoromethane  
Concen: 21.820 ug/l  
RT: 1.172 min Scan# 14  
Delta R.T. 0.006 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59



Tgt Ion: 85 Resp: 40309  
Ion Ratio Lower Upper  
85 100  
87 28.0 16.4 49.2









#3  
Chloromethane  
Concen: 21.289 ug/l  
RT: 1.294 min Scan# 34  
Delta R.T. 0.000 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59

Tgt Ion: 50 Resp: 45814  
Ion Ratio Lower Upper  
50 100  
52 33.2 26.4 39.6

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0805WBS01




Manual Integrations  
APPROVED


MMDadoda  
8/6/2021 5:41:01 PM



#4  
Vinyl Chloride  
Concen: 21.492 ug/l  
RT: 1.373 min Scan# 47  
Delta R.T. -0.001 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59

Tgt Ion: 62 Resp: 46817  
Ion Ratio Lower Upper  
62 100  
64 32.6 25.7 38.5









#5  
Bromomethane  
Concen: 25.974 ug/l  
RT: 1.599 min Scan# 84  
Delta R.T. -0.000 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59

Tgt Ion: 94 Resp: 33592  
Ion Ratio Lower Upper  
94 100  
96 93.9 76.4 114.6

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0805WBS01




Manual Integrations  
APPROVED


MMDadoda  
8/6/2021 5:41:01 PM



#6  
Chloroethane  
Concen: 22.766 ug/l  
RT: 1.684 min Scan# 98  
Delta R.T. 0.006 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59

Tgt Ion: 64 Resp: 30436  
Ion Ratio Lower Upper  
64 100  
66 31.4 25.4 38.0









#7  
Trichlorofluoromethane  
Concen: 21.422 ug/l  
RT: 1.885 min Scan# 131  
Delta R.T. -0.001 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59

Tgt Ion:101 Resp: 75184  
Ion Ratio Lower Upper  
101 100  
103 67.0 53.2 79.8

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0805WBS01




Manual Integrations  
APPROVED

MMDadoda  
8/6/2021 5:41:01 PM




#8  
Diethyl Ether  
Concen: 21.556 ug/l  
RT: 2.141 min Scan# 173  
Delta R.T. 0.005 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59

Tgt Ion: 74 Resp: 27923  
Ion Ratio Lower Upper  
74 100  
45 96.5 44.5 133.5











#9  
1,1,2-Trichlorotrifluoroethane  
Concen: 21.889 ug/l  
RT: 2.330 min Scan# 204  
Delta R.T. -0.001 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59

Tgt Ion:101 Resp: 43522  
Ion Ratio Lower Upper  
101 100  
85 45.5 34.9 52.3  
151 80.3 61.4 92.2

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0805WBS01




Manual Integrations  
APPROVED


MMDadoda  
8/6/2021 5:41:01 PM



#10  
Methyl Iodide  
Concen: 20.980 ug/l  
RT: 2.452 min Scan# 224  
Delta R.T. -0.001 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59

Tgt Ion:142 Resp: 52265  
Ion Ratio Lower Upper  
142 100  
127 43.9 33.3 49.9  
141 14.3 11.4 17.0









#11  
Tert butyl alcohol  
Concen: 111.101 ug/l  
RT: 2.977 min Scan# 310  
Delta R.T. -0.006 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59

Tgt Ion: 59 Resp: 63572  
Ion Ratio Lower Upper  
59 100  
57 10.8 7.8 11.8

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0805WBS01




Manual Integrations  
APPROVED


MMDadoda  
8/6/2021 5:41:01 PM




#12  
1,1-Dichloroethene  
Concen: 21.068 ug/l  
RT: 2.318 min Scan# 202  
Delta R.T. -0.001 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59

Tgt Ion: 96 Resp: 41568  
Ion Ratio Lower Upper  
96 100  
61 164.0 124.6 187.0  
98 64.4 50.6 76.0







#13  
Acrolein  
Concen: 83.670 ug/l  
RT: 2.239 min Scan# 189  
Delta R.T. -0.000 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59




Tgt Ion: 56 Resp: 24073  
Ion Ratio Lower Upper  
56 100  
55 74.6 56.4 84.6




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0805WBS01

Manual Integrations  
APPROVED



MMDadoda  
8/6/2021 5:41:01 PM




#14  
Allyl chloride  
Concen: 20.943 ug/l  
RT: 2.666 min Scan# 259  
Delta R.T. -0.000 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59



Tgt Ion: 41 Resp: 75125  
Ion Ratio Lower Upper  
41 100  
39 69.5 51.6 77.4  
76 33.2 30.5 45.7









#15  
Acrylonitrile  
Concen: 113.424 ug/l  
RT: 3.068 min Scan# 325  
Delta R.T. -0.007 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59

Tgt Ion: 53 Resp: 146556  
Ion Ratio Lower Upper  
53 100  
52 83.2 65.3 97.9  
51 35.8 28.2 42.2

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0805WBS01




Manual Integrations  
APPROVED


MMDadoda  
8/6/2021 5:41:01 PM



#16  
Acetone  
Concen: 108.188 ug/l  
RT: 2.385 min Scan# 213  
Delta R.T. -0.001 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59

Tgt Ion: 43 Resp: 132932  
Ion Ratio Lower Upper  
43 100  
58 31.6 27.8 41.8









#17  
Carbon Disulfide  
Concen: 18.196 ug/l  
RT: 2.513 min Scan# 234  
Delta R.T. -0.001 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59

Tgt Ion: 76 Resp: 88526  
Ion Ratio Lower Upper  
76 100  
78 9.2 7.2 10.8

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0805WBS01




Manual Integrations  
APPROVED


MMDadoda  
8/6/2021 5:41:01 PM



#18  
Methyl Acetate  
Concen: 23.266 ug/l  
RT: 2.708 min Scan# 266  
Delta R.T. -0.001 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59

Tgt Ion: 43 Resp: 67958  
Ion Ratio Lower Upper  
43 100  
74 23.0 21.9 32.9





#19





Methyl tert-butyl Ether  
Concen: 21.888 ug/l  
RT: 3.123 min Scan# 334  
Delta R.T. -0.000 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59

Tgt Ion: 73 Resp: 150158  
Ion Ratio Lower Upper  
73 100  
57 22.1 17.1 25.7

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0805WBS01

Manual Integrations  
APPROVED




MMDadoda  
8/6/2021 5:41:01 PM




#20

Methylene Chloride  
Concen: 21.112 ug/l  
RT: 2.794 min Scan# 280  
Delta R.T. -0.000 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59

Tgt Ion: 84 Resp: 50459  
Ion Ratio Lower Upper  
84 100  
49 129.0 88.6 132.8  
51 41.0 27.0 40.6#  
86 64.6 50.3 75.5









#21  
trans-1,2-Dichloroethene  
Concen: 21.250 ug/l  
RT: 3.099 min Scan# 330  
Delta R.T. 0.006 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59

Tgt Ion: 96 Resp: 44539  
Ion Ratio Lower Upper  
96 100  
61 134.6 107.7 161.5  
98 63.8 50.4 75.6

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0805WBS01




Manual Integrations  
APPROVED


MMDadoda  
8/6/2021 5:41:01 PM



#22  
Diisopropyl ether  
Concen: 22.230 ug/l  
RT: 3.769 min Scan# 440  
Delta R.T. 0.000 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59

Tgt Ion: 45 Resp: 157410  
Ion Ratio Lower Upper  
45 100  
43 81.5 58.0 87.0  
87 28.8 25.4 38.0  
59 10.6 10.2 15.2









#23  
 Vinyl Acetate  
 Concen: 111.308 ug/l  
 RT: 3.733 min Scan# 434  
 Delta R.T. -0.000 min  
 Lab File: VX023547.D  
 Acq: 04 Aug 2021 17:59

Tgt Ion: 43 Resp: 695761  
 Ion Ratio Lower Upper  
 43 100  
 86 10.6 9.9 14.9

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX0805WBS01




Manual Integrations  
 APPROVED

MMDadoda  
 8/6/2021 5:41:01 PM




#24  
 1,1-Dichloroethane  
 Concen: 21.946 ug/l  
 RT: 3.617 min Scan# 415  
 Delta R.T. -0.000 min  
 Lab File: VX023547.D  
 Acq: 04 Aug 2021 17:59

Tgt Ion: 63 Resp: 85413  
 Ion Ratio Lower Upper  
 63 100  
 98 6.7 3.8 11.2  
 100 3.7 2.5 7.5











#25  
2-Butanone  
Concen: 110.596 ug/l  
RT: 4.568 min Scan# 571  
Delta R.T. -0.000 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59

Tgt Ion: 43 Resp: 210967  
Ion Ratio Lower Upper  
43 100  
72 26.1 23.7 35.5

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0805WBS01




Manual Integrations  
APPROVED


MMDadoda  
8/6/2021 5:41:01 PM



#26  
2,2-Dichloropropane  
Concen: 21.224 ug/l  
RT: 4.489 min Scan# 558  
Delta R.T. 0.006 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59

Tgt Ion: 77 Resp: 67557  
Ion Ratio Lower Upper  
77 100  
97 23.9 11.7 35.1





#27  
cis-1,2-Dichloroethene  
Concen: 21.539 ug/l  
RT: 4.501 min Scan# 560  
Delta R.T. 0.006 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59





Tgt Ion: 96 Resp: 52428

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 96  | 100   |       |       |
| 61  | 149.0 | 0.0   | 285.6 |
| 98  | 66.0  | 0.0   | 129.6 |

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0805WBS01

Manual Integrations  
APPROVED




MMDadoda  
8/6/2021 5:41:01 PM




#28  
Bromochloromethane  
Concen: 21.973 ug/l  
RT: 4.915 min Scan# 628  
Delta R.T. 0.005 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59

Tgt Ion: 49 Resp: 39911

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 49  | 100   |       |       |
| 129 | 1.9   | 0.0   | 4.0   |
| 130 | 77.4  | 68.2  | 102.2 |





#29

Tetrahydrofuran

Concen: 111.543 ug/l

RT: 5.019 min Scan# 645

Delta R.T. 0.006 min

Lab File: VX023547.D

Acq: 04 Aug 2021 17:59

Tgt Ion: 42 Resp: 139467

Ion Ratio Lower Upper

42 100

72 44.4 41.7 62.5

71 40.9 38.4 57.6

Instrument :

MSVOA\_X

ClientSampleId :





VX0805WBS01

Manual Integrations

APPROVED

MMDadoda

8/6/2021 5:41:01 PM



#30

Chloroform

Concen: 22.127 ug/l

RT: 5.104 min Scan# 659

Delta R.T. -0.001 min

Lab File: VX023547.D




Acq: 04 Aug 2021 17:59


Tgt Ion: 83 Resp: 88141

Ion Ratio Lower Upper


83 100

85 66.1 52.0 78.0








#31  
Cyclohexane  
Concen: 19.669 ug/l  
RT: 5.476 min Scan# 720  
Delta R.T. -0.001 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59




Tgt Ion: 56 Resp: 72982

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 56  | 100   |       |       |
| 69  | 29.9  | 25.3  | 37.9  |
| 84  | 85.1  | 71.4  | 107.0 |





#32  
1,1,1-Trichloroethane  
Concen: 22.180 ug/l  
RT: 5.397 min Scan# 707  
Delta R.T. 0.006 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59



Tgt Ion: 97 Resp: 75510


| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 97  | 100   |       |       |
| 99  | 64.4  | 51.5  | 77.3  |
| 61  | 46.3  | 34.2  | 51.4  |



Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0805WBS01

Manual Integrations  
APPROVED

MMDadoda  
8/6/2021 5:41:01 PM



#33





1,2-Dichloroethane-d4  
Concen: 55.344 ug/l  
RT: 5.964 min Scan# 800  
Delta R.T. -0.000 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59

Tgt Ion: 65 Resp: 138626  
Ion Ratio Lower Upper  
65 100  
67 51.9 0.0 105.4

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0805WBS01

Manual Integrations  
APPROVED




MMDadoda  
8/6/2021 5:41:01 PM




#34


1,4-Difluorobenzene  
Concen: 50.000 ug/l  
RT: 6.769 min Scan# 932  
Delta R.T. -0.000 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59

Tgt Ion: 114 Resp: 326433  
Ion Ratio Lower Upper  
114 100  
63 19.8 0.0 40.8  
88 15.8 0.0 32.8









#35  
Dibromofluoromethane  
Concen: 54.226 ug/l  
RT: 5.397 min Scan# 707  
Delta R.T. -0.000 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59





Tgt Ion: 113 Resp: 110675  
Ion Ratio Lower Upper  
113 100  
111 102.7 81.9 122.9  
192 19.6 16.7 25.1



#36  
1,1-Dichloropropene  
Concen: 21.199 ug/l  
RT: 5.696 min Scan# 756  
Delta R.T. -0.000 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59




Tgt Ion: 75 Resp: 62598  
Ion Ratio Lower Upper  
75 100  
110 36.6 18.1 54.4  
77 31.6 25.3 37.9



Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0805WBS01


Manual Integrations  
APPROVED

MMDadoda  
8/6/2021 5:41:01 PM



#37  
Ethyl Acetate  
Concen: 21.445 ug/l  
RT: 4.726 min Scan# 597  
Delta R.T. -0.001 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0805WBS01




Tgt Ion: 43 Resp: 79844  
Ion Ratio Lower Upper  
43 100  
61 14.2 12.8 19.2  
70 10.4 10.5 15.7#

Manual Integrations  
APPROVED



MMDadoda  
8/6/2021 5:41:01 PM




#38  
Carbon Tetrachloride  
Concen: 21.130 ug/l  
RT: 5.690 min Scan# 755  
Delta R.T. 0.006 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59



Tgt Ion: 117 Resp: 63455  
Ion Ratio Lower Upper  
117 100  
119 99.5 76.2 114.2  
121 30.3 24.8 37.2





#39

Methylcyclohexane

Concen: 20.837 ug/l

RT: 7.384 min Scan# 1033

Delta R.T. -0.001 min

Lab File: VX023547.D

Acq: 04 Aug 2021 17:59





Tgt Ion: 83 Resp: 77708

Ion Ratio Lower Upper

83 100

55 70.1 58.9 88.3

98 43.1 38.1 57.1



#40

Benzene

Concen: 21.666 ug/l

RT: 6.043 min Scan# 813

Delta R.T. 0.000 min

Lab File: VX023547.D




Acq: 04 Aug 2021 17:59

Tgt Ion: 78 Resp: 189696

Ion Ratio Lower Upper

78 100

77 22.8 18.9 28.3



Instrument :

MSVOA\_X

ClientSampleId :


VX0805WBS01

Manual Integrations


APPROVED

MMDadoda  
8/6/2021 5:41:01 PM









#41  
Methacrylonitrile  
Concen: 21.587 ug/l  
RT: 4.928 min Scan# 630  
Delta R.T. -0.000 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59





Tgt Ion: 41 Resp: 42437  
Ion Ratio Lower Upper  
41 100  
39 59.6 42.5 63.7  
67 73.0 65.0 97.6  
52 31.7 25.3 37.9



#42  
1,2-Dichloroethane  
Concen: 22.768 ug/l  
RT: 6.098 min Scan# 822  
Delta R.T. 0.000 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59




Tgt Ion: 62 Resp: 73704  
Ion Ratio Lower Upper  
62 100  
98 8.4 0.0 20.2




Instrument :  
MSVOA\_X  
ClientSampled :  
VX0805WBS01

Manual Integrations  
APPROVED



MMDadoda  
8/6/2021 5:41:01 PM



#43  
Isopropyl Acetate  
Concen: 21.965 ug/l  
RT: 6.348 min Scan# 863  
Delta R.T. -0.000 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59




Tgt Ion: 43 Resp: 126725  
Ion Ratio Lower Upper  
43 100  
61 19.0 18.2 27.4  
87 13.2 12.5 18.7




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0805WBS01

Manual Integrations  
APPROVED



MMDadoda  
8/6/2021 5:41:01 PM




#44  
Trichloroethene  
Concen: 21.484 ug/l  
RT: 7.134 min Scan# 992  
Delta R.T. 0.005 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59



Tgt Ion:130 Resp: 51676  
Ion Ratio Lower Upper  
130 100  
95 89.7 0.0 198.0





#45





1,2-Dichloropropane  
Concen: 21.623 ug/l  
RT: 7.439 min Scan# 1042  
Delta R.T. -0.001 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59

Tgt Ion: 63 Resp: 50097  
Ion Ratio Lower Upper  
63 100  
65 32.9 25.7 38.5

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0805WBS01

Manual Integrations  
APPROVED




MMDadoda  
8/6/2021 5:41:01 PM




#46

Dibromomethane  
Concen: 22.379 ug/l  
RT: 7.586 min Scan# 1066  
Delta R.T. -0.000 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59

Tgt Ion: 93 Resp: 35168  
Ion Ratio Lower Upper  
93 100  
95 81.9 66.8 100.2  
174 102.7 83.1 124.7









#47  
Bromodichloromethane  
Concen: 21.842 ug/l  
RT: 7.829 min Scan# 1106  
Delta R.T. -0.001 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59

Tgt Ion: 83 Resp: 61840  
Ion Ratio Lower Upper  
83 100  
85 65.5 51.4 77.0  
127 9.1 6.8 10.2

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0805WBS01




Manual Integrations  
APPROVED


MMDadoda  
8/6/2021 5:41:01 PM



#48  
Methyl methacrylate  
Concen: 22.692 ug/l  
RT: 7.701 min Scan# 1085  
Delta R.T. -0.001 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59

Tgt Ion: 41 Resp: 60633  
Ion Ratio Lower Upper  
41 100  
69 79.3 76.0 114.0  
39 60.3 44.6 67.0









#49  
1,4-Dioxane  
Concen: 471.721 ug/l  
RT: 7.665 min Scan# 1079  
Delta R.T. -0.000 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59

|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 88    | Resp: | 26839 |
| Ion      | Ratio | Lower | Upper |
| 88       | 100   |       |       |
| 43       | 31.7  | 23.2  | 34.8  |
| 58       | 69.5  | 55.0  | 82.4  |

Instrument :  
MSVOA\_X  
ClientSampled :  
VX0805WBS01




Manual Integrations  
APPROVED


MMDadoda  
8/6/2021 5:41:01 PM



#50  
Toluene-d8  
Concen: 53.791 ug/l  
RT: 8.652 min Scan# 1241  
Delta R.T. -0.001 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 98    | Resp: | 403359 |
| Ion      | Ratio | Lower | Upper  |
| 98       | 100   |       |        |
| 100      | 64.3  | 51.9  | 77.9   |









#51  
4-Methyl-2-Pentanone  
Concen: 119.980 ug/l  
RT: 8.579 min Scan# 1229  
Delta R.T. -0.001 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59

Tgt Ion: 43 Resp: 423382  
Ion Ratio Lower Upper  
43 100  
58 37.8 35.4 53.0

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0805WBS01




Manual Integrations  
APPROVED


MMDadoda  
8/6/2021 5:41:01 PM



#52  
Toluene  
Concen: 22.597 ug/l  
RT: 8.720 min Scan# 1252  
Delta R.T. -0.000 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59

Tgt Ion: 92 Resp: 122570  
Ion Ratio Lower Upper  
92 100  
91 172.3 137.1 205.7





#53





t-1,3-Dichloropropene  
Concen: 20.596 ug/l  
RT: 8.982 min Scan# 1295  
Delta R.T. -0.000 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59

Tgt Ion: 75 Resp: 69977  
Ion Ratio Lower Upper  
75 100  
77 33.0 24.9 37.3

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0805WBS01

Manual Integrations  
APPROVED




MMDadoda  
8/6/2021 5:41:01 PM




#54

cis-1,3-Dichloropropene  
Concen: 21.821 ug/l  
RT: 8.372 min Scan# 1195  
Delta R.T. 0.000 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59

Tgt Ion: 75 Resp: 77015  
Ion Ratio Lower Upper  
75 100  
77 29.9 25.7 38.5  
39 50.9 35.1 52.7









#55  
1,1,2-Trichloroethane  
Concen: 23.558 ug/l  
RT: 9.158 min Scan# 1324  
Delta R.T. -0.001 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59

|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 97    | Resp: | 51745 |
| Ion      | Ratio | Lower | Upper |
| 97       | 100   |       |       |
| 83       | 85.2  | 69.3  | 103.9 |
| 85       | 54.9  | 44.3  | 66.5  |
| 99       | 64.1  | 50.3  | 75.5  |

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0805WBS01




Manual Integrations  
APPROVED

MMDadoda  
8/6/2021 5:41:01 PM




#56  
Ethyl methacrylate  
Concen: 22.873 ug/l  
RT: 9.122 min Scan# 1318  
Delta R.T. -0.000 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59

|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 69    | Resp: | 80532 |
| Ion      | Ratio | Lower | Upper |
| 69       | 100   |       |       |
| 41       | 68.8  | 46.8  | 70.2  |
| 39       | 43.2  | 24.7  | 37.1  |











#57  
1,3-Dichloropropane  
Concen: 22.883 ug/l  
RT: 9.311 min Scan# 1349  
Delta R.T. -0.000 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59

Tgt Ion: 76 Resp: 85405  
Ion Ratio Lower Upper  
76 100  
78 32.2 26.2 39.2

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0805WBS01




Manual Integrations  
APPROVED


MMDadoda  
8/6/2021 5:41:01 PM



#58  
2-Chloroethyl Vinyl ether  
Concen: 100.188 ug/l  
RT: 8.244 min Scan# 1174  
Delta R.T. -0.006 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59

Tgt Ion: 63 Resp: 186865  
Ion Ratio Lower Upper  
63 100  
106 27.7 23.5 35.3









#59  
2-Hexanone  
Concen: 122.560 ug/l  
RT: 9.433 min Scan# 1369  
Delta R.T. -0.000 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59

Tgt Ion: 43 Resp: 328421  
Ion Ratio Lower Upper  
43 100  
58 52.5 30.9 92.8

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0805WBS01




Manual Integrations  
APPROVED


MMDadoda  
8/6/2021 5:41:01 PM



#60  
Dibromochloromethane  
Concen: 20.793 ug/l  
RT: 9.524 min Scan# 1384  
Delta R.T. -0.001 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59

Tgt Ion: 129 Resp: 47683  
Ion Ratio Lower Upper  
129 100  
127 77.7 38.4 115.2






#61  
1,2-Dibromoethane  
Concen: 23.121 ug/l  
RT: 9.616 min Scan# 1399  
Delta R.T. -0.000 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0805WBS01

Manual Integrations  
APPROVED

MMDadoda  
8/6/2021 5:41:01 PM







Tgt Ion:107 Resp: 54182  
Ion Ratio Lower Upper  
107 100  
109 93.9 75.8 113.6



#62  
4-Bromofluorobenzene  
Concen: 57.521 ug/l  
RT: 11.085 min Scan# 1640  
Delta R.T. -0.000 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59


Tgt Ion: 95 Resp: 146671  
Ion Ratio Lower Upper  
95 100  
174 82.0 0.0 154.6  
176 78.5 0.0 155.8





#63  
Chlorobenzene-d5  
Concen: 50.000 ug/l  
RT: 10.055 min Scan# 1471  
Delta R.T. -0.000 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0805WBS01




Tgt Ion:117 Resp: 300596  
Ion Ratio Lower Upper  
117 100  
82 56.4 47.4 71.2  
119 32.2 26.6 39.8

Manual Integrations  
APPROVED



MMDadoda  
8/6/2021 5:41:01 PM




#64  
Tetrachloroethene  
Concen: 21.228 ug/l  
RT: 9.274 min Scan# 1343  
Delta R.T. -0.001 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59



Tgt Ion:164 Resp: 50784  
Ion Ratio Lower Upper  
164 100  
166 130.8 103.4 155.2  
129 102.8 70.8 106.2  
131 93.3 72.5 108.7









#65  
Chlorobenzene  
Concen: 21.558 ug/l  
RT: 10.085 min Scan# 1476  
Delta R.T. 0.000 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59

Tgt Ion:112 Resp: 132633  
Ion Ratio Lower Upper  
112 100  
114 32.6 25.8 38.6

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0805WBS01




Manual Integrations  
APPROVED


MMDadoda  
8/6/2021 5:41:01 PM



#66  
1,1,1,2-Tetrachloroethane  
Concen: 21.220 ug/l  
RT: 10.164 min Scan# 1489  
Delta R.T. -0.001 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59


Tgt Ion:131 Resp: 48815  
Ion Ratio Lower Upper  
131 100  
133 95.5 47.7 143.1  
119 65.1 32.1 96.5





#67  
Ethyl Benzene  
Concen: 21.875 ug/l  
RT: 10.195 min Scan# 1494  
Delta R.T. -0.000 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0805WBS01




Tgt Ion: 91 Resp: 233415  
Ion Ratio Lower Upper  
91 100  
106 30.4 24.6 37.0

Manual Integrations  
APPROVED



MMDadoda  
8/6/2021 5:41:01 PM




#68  
m/p-Xylenes  
Concen: 42.886 ug/l  
RT: 10.305 min Scan# 1512  
Delta R.T. -0.000 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59



Tgt Ion: 106 Resp: 175994  
Ion Ratio Lower Upper  
106 100  
91 203.9 162.8 244.2









#69  
o-Xylene  
Concen: 21.809 ug/l  
RT: 10.646 min Scan# 1568  
Delta R.T. 0.000 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59

Tgt Ion:106 Resp: 88732  
Ion Ratio Lower Upper  
106 100  
91 211.2 108.1 324.2

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0805WBS01




Manual Integrations  
APPROVED


MMDadoda  
8/6/2021 5:41:01 PM




#70  
Styrene  
Concen: 22.377 ug/l  
RT: 10.658 min Scan# 1570  
Delta R.T. -0.001 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59

Tgt Ion:104 Resp: 144251  
Ion Ratio Lower Upper  
104 100  
78 50.4 40.0 60.0  
103 54.4 43.4 65.0









#71  
Bromoform  
Concen: 20.291 ug/l  
RT: 10.805 min Scan# 1594  
Delta R.T. -0.000 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59





Tgt Ion:173 Resp: 33430  
Ion Ratio Lower Upper  
173 100  
175 47.2 23.9 71.8  
254 0.0 0.1 0.1#



#72  
1,4-Dichlorobenzene-d4  
Concen: 50.000 ug/l  
RT: 12.024 min Scan# 1794  
Delta R.T. -0.000 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59



Tgt Ion:152 Resp: 135225  
Ion Ratio Lower Upper  
152 100  
115 70.6 41.1 123.3  
150 166.3 0.0 349.0




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0805WBS01

Manual Integrations  
APPROVED

MMDadoda  
8/6/2021 5:41:01 PM









#73  
Isopropylbenzene  
Concen: 22.558 ug/l  
RT: 10.963 min Scan# 1620  
Delta R.T. 0.000 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59

Tgt Ion: 105 Resp: 230176  
Ion Ratio Lower Upper  
105 100  
120 27.3 13.2 39.6

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0805WBS01




Manual Integrations  
APPROVED


MMDadoda  
8/6/2021 5:41:01 PM



#74  
N-ethyl acetate  
Concen: 22.580 ug/l  
RT: 10.847 min Scan# 1601  
Delta R.T. -0.001 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59

Tgt Ion: 43 Resp: 105114  
Ion Ratio Lower Upper  
43 100  
70 40.8 39.0 58.6  
55 24.0 21.5 32.3  
61 22.6 22.1 33.1









#75  
1,1,2,2-Tetrachloroethane  
Concen: 23.349 ug/l  
RT: 11.213 min Scan# 1661  
Delta R.T. -0.000 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59

Tgt Ion: 83 Resp: 77609  
Ion Ratio Lower Upper  
83 100  
131 10.7 5.1 15.2  
85 64.2 31.8 95.4

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0805WBS01




Manual Integrations  
APPROVED


MMDadoda  
8/6/2021 5:41:01 PM



#76  
1,2,3-Trichloropropane  
Concen: 24.108 ug/l m  
RT: 11.243 min Scan# 1666  
Delta R.T. -0.001 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59

Tgt Ion: 75 Resp: 71740  
Ion Ratio Lower Upper  
75 100  
77 39.8 19.1 57.5









#77  
Bromobenzene  
Concen: 23.231 ug/l  
RT: 11.201 min Scan# 1659  
Delta R.T. -0.000 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59

Tgt Ion: 156 Resp: 57183  
Ion Ratio Lower Upper  
156 100  
77 156.5 82.8 248.6  
158 94.7 49.1 147.4

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0805WBS01




Manual Integrations  
APPROVED


MMDadoda  
8/6/2021 5:41:01 PM



#78  
n-propylbenzene  
Concen: 22.487 ug/l  
RT: 11.304 min Scan# 1676  
Delta R.T. -0.007 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59

Tgt Ion: 91 Resp: 259687  
Ion Ratio Lower Upper  
91 100  
120 23.2 11.5 34.5









#79  
2-Chlorotoluene  
Concen: 22.572 ug/l  
RT: 11.365 min Scan# 1686  
Delta R.T. -0.001 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59

Tgt Ion: 91 Resp: 155762  
Ion Ratio Lower Upper  
91 100  
126 34.9 16.6 49.8

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0805WBS01




Manual Integrations  
APPROVED


MMDadoda  
8/6/2021 5:41:01 PM



#80  
1,3,5-Trimethylbenzene  
Concen: 22.835 ug/l  
RT: 11.457 min Scan# 1701  
Delta R.T. -0.000 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59


Tgt Ion: 105 Resp: 189756  
Ion Ratio Lower Upper  
105 100  
120 48.4 23.8 71.5





#81  
trans-1,4-Dichloro-2-butene  
Concen: 19.881 ug/l  
RT: 11.024 min Scan# 1630  
Delta R.T. -0.000 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0805WBS01




Tgt Ion: 75 Resp: 22002  
Ion Ratio Lower Upper  
75 100  
53 114.1 74.5 111.7#  
89 48.5 37.9 56.9

Manual Integrations  
APPROVED



MMDadoda  
8/6/2021 5:41:01 PM




#82  
4-Chlorotoluene  
Concen: 22.589 ug/l  
RT: 11.457 min Scan# 1701  
Delta R.T. -0.000 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59



Tgt Ion: 91 Resp: 176347  
Ion Ratio Lower Upper  
91 100  
126 30.4 14.9 44.5









#83  
tert-Butylbenzene  
Concen: 22.932 ug/l  
RT: 11.719 min Scan# 1744  
Delta R.T. -0.000 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59

Tgt Ion:119 Resp: 186920  
Ion Ratio Lower Upper  
119 100  
91 58.9 27.9 83.7  
134 22.9 11.3 33.9

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0805WBS01




Manual Integrations  
APPROVED


MMDadoda  
8/6/2021 5:41:01 PM



#84  
1,2,4-Trimethylbenzene  
Concen: 23.226 ug/l  
RT: 11.756 min Scan# 1750  
Delta R.T. -0.000 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59


Tgt Ion:105 Resp: 191977  
Ion Ratio Lower Upper  
105 100  
120 44.3 22.4 67.3





#85  
sec-Butylbenzene  
Concen: 22.259 ug/l  
RT: 11.896 min Scan# 1773  
Delta R.T. -0.000 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0805WBS01




Tgt Ion:105 Resp: 235250  
Ion Ratio Lower Upper  
105 100  
134 20.2 10.0 29.8

Manual Integrations  
APPROVED



MMDadoda  
8/6/2021 5:41:01 PM




#86  
p-Isopropyltoluene  
Concen: 22.589 ug/l  
RT: 12.012 min Scan# 1792  
Delta R.T. -0.000 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59



Tgt Ion:119 Resp: 192793  
Ion Ratio Lower Upper  
119 100  
134 26.5 12.9 38.7  
91 24.7 11.7 35.0









#87  
1,3-Dichlorobenzene  
Concen: 21.990 ug/l  
RT: 11.975 min Scan# 1786  
Delta R.T. 0.000 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59

Tgt Ion:146 Resp: 100138  
Ion Ratio Lower Upper  
146 100  
111 39.5 20.2 60.5  
148 62.9 31.9 95.7

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0805WBS01




Manual Integrations  
APPROVED

MMDadoda  
8/6/2021 5:41:01 PM




#88  
1,4-Dichlorobenzene  
Concen: 21.654 ug/l  
RT: 12.042 min Scan# 1797  
Delta R.T. 0.000 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59

Tgt Ion:146 Resp: 100055  
Ion Ratio Lower Upper  
146 100  
111 40.1 19.6 58.8  
148 63.6 31.6 95.0











#89  
n-Butylbenzene  
Concen: 21.395 ug/l  
RT: 12.335 min Scan# 1845  
Delta R.T. -0.000 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59

Tgt Ion: 91 Resp: 165300  
Ion Ratio Lower Upper  
91 100  
92 52.3 27.4 82.0  
134 26.2 13.2 39.5

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0805WBS01




Manual Integrations  
APPROVED


MMDadoda  
8/6/2021 5:41:01 PM



#90  
Hexachloroethane  
Concen: 20.281 ug/l  
RT: 12.542 min Scan# 1879  
Delta R.T. 0.000 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59

Tgt Ion: 117 Resp: 29208  
Ion Ratio Lower Upper  
117 100  
201 76.4 40.4 121.2









#91  
1,2-Dichlorobenzene  
Concen: 22.607 ug/l  
RT: 12.341 min Scan# 1846  
Delta R.T. -0.000 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59

Tgt Ion: 146 Resp: 99680  
Ion Ratio Lower Upper  
146 100  
111 40.9 20.8 62.2  
148 63.4 31.9 95.5

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0805WBS01




Manual Integrations  
APPROVED


MMDadoda  
8/6/2021 5:41:01 PM



#92  
1,2-Dibromo-3-Chloropropane  
Concen: 20.628 ug/l  
RT: 12.944 min Scan# 1945  
Delta R.T. -0.001 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59

Tgt Ion: 75 Resp: 15586  
Ion Ratio Lower Upper  
75 100  
155 82.0 45.3 135.9  
157 106.2 57.7 173.1









#93  
1,2,4-Trichlorobenzene  
Concen: 21.467 ug/l  
RT: 13.591 min Scan# 2051  
Delta R.T. -0.000 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59

|           |       |       |       |
|-----------|-------|-------|-------|
| Tgt Ion:  | 180   | Resp: | 59469 |
| Ion Ratio | Lower | Upper |       |
| 180       | 100   |       |       |
| 182       | 92.5  | 47.7  | 143.1 |
| 145       | 31.7  | 15.0  | 45.0  |

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0805WBS01




Manual Integrations  
APPROVED


MMDadoda  
8/6/2021 5:41:01 PM




#94  
Hexachlorobutadiene  
Concen: 20.983 ug/l  
RT: 13.725 min Scan# 2073  
Delta R.T. -0.006 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59

|           |       |       |       |
|-----------|-------|-------|-------|
| Tgt Ion:  | 225   | Resp: | 26086 |
| Ion Ratio | Lower | Upper |       |
| 225       | 100   |       |       |
| 223       | 62.2  | 31.1  | 93.2  |
| 227       | 62.7  | 31.6  | 95.0  |







#95  
Naphthalene  
Concen: 21.909 ug/l  
RT: 13.780 min Scan# 2082  
Delta R.T. -0.000 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59




Tgt Ion:128 Resp: 208676  
Ion Ratio Lower Upper  
128 100  
127 13.1 10.3 15.5  
129 10.6 8.8 13.2




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0805WBS01

Manual Integrations  
APPROVED

MMDadoda  
8/6/2021 5:41:01 PM



#96  
1,2,3-Trichlorobenzene  
Concen: 21.117 ug/l  
RT: 13.962 min Scan# 2112  
Delta R.T. -0.001 min  
Lab File: VX023547.D  
Acq: 04 Aug 2021 17:59



Tgt Ion:180 Resp: 59075  
Ion Ratio Lower Upper  
180 100  
182 95.4 47.6 142.9  
145 33.2 16.1 48.3

