

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080522\
 Data File : VX030432.D
 Acq On : 05 Aug 2022 16:09
 Operator : JC/MD
 Sample : N4003-19DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 F5D89DL

Quant Time: Aug 06 00:58:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML080422WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Aug 06 00:54:06 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.769	114	263688	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	226353	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	99634	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	87685	51.782	ug/L	0.00
Spiked Amount 50.000	Range 60	- 135	Recovery	=	103.560%	
7) Chloroethane-d5	1.666	69	68776	57.614	ug/L	0.00
Spiked Amount 50.000	Range 70	- 130	Recovery	=	115.220%	
11) 1,1-Dichloroethene-d2	2.306	63	140379	43.825	ug/L	0.00
Spiked Amount 50.000	Range 60	- 125	Recovery	=	87.660%	
21) 2-Butanone-d5	4.465	46	196250	125.081	ug/L	0.00
Spiked Amount 100.000	Range 40	- 130	Recovery	=	125.080%	
24) Chloroform-d	5.062	84	190197	54.563	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	=	109.120%	
26) 1,2-Dichloroethane-d4	5.964	65	135280	57.155	ug/L	0.01
Spiked Amount 50.000	Range 70	- 125	Recovery	=	114.320%	
32) Benzene-d6	5.983	84	373598	57.620	ug/L	0.01
Spiked Amount 50.000	Range 70	- 125	Recovery	=	115.240%	
36) 1,2-Dichloropropane-d6	7.318	67	127316	60.522	ug/L	0.00
Spiked Amount 50.000	Range 70	- 120	Recovery	=	121.040%#	
41) Toluene-d8	8.653	98	338910	55.453	ug/L	0.00
Spiked Amount 50.000	Range 80	- 120	Recovery	=	110.900%	
43) trans-1,3-Dichloroprop...	8.958	79	51906	49.599	ug/L	0.00
Spiked Amount 50.000	Range 60	- 125	Recovery	=	99.200%	
47) 2-Hexanone-d5	9.391	63	130331	121.548	ug/L	0.00
Spiked Amount 100.000	Range 45	- 130	Recovery	=	121.550%	
56) 1,1,2,2-Tetrachloroeth...	11.195	84	162339	58.744	ug/L	0.00
Spiked Amount 50.000	Range 65	- 120	Recovery	=	117.480%	
66) 1,2-Dichlorobenzene-d4	12.323	152	118591	61.366	ug/L	0.00
Spiked Amount 50.000	Range 80	- 120	Recovery	=	122.740%#	
Target Compounds						Qvalue
5) Vinyl chloride	1.374	62	58722	32.222	ug/L	98
17) trans-1,2-Dichloroethene	3.093	96	2773	1.622	ug/L	96
20) cis-1,2-Dichloroethene	4.495	96	42483	22.087	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080522\
Data File : VX030432.D
Acq On : 05 Aug 2022 16:09
Operator : JC/MD
Sample : N4003-19DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
F5D89DL

Quant Time: Aug 06 00:58:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML080422WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Aug 06 00:54:06 2022
Response via : Initial Calibration

