

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080525\
 Data File : VX047230.D
 Acq On : 05 Aug 2025 11:25
 Operator : JC/MD
 Sample : VX0805MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0805MBL01

Quant Time: Aug 06 05:39:19 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 22 03:59:51 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	281491	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	525600	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	505849	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	256242	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	186014	53.049	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	106.100%
35) Dibromofluoromethane	5.391	113	154418	47.583	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.160%
50) Toluene-d8	8.647	98	570424	54.276	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	108.560%
62) 4-Bromofluorobenzene	11.079	95	236668	52.256	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	104.520%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080525\
Data File : VX047230.D
Acq On : 05 Aug 2025 11:25
Operator : JC/MD
Sample : VX0805MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0805MBL01

Quant Time: Aug 06 05:39:19 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 22 03:59:51 2025
Response via : Initial Calibration

