



BFB

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080525\
Data File : VX047228.D
Acq On : 05 Aug 2025 08:46
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
Title : SW846 8260
Last Update : Tue Jul 22 03:59:51 2025



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	19.4	24043	PASS
75	95	30	60	51.9	64405	PASS
95	95	100	100	100.0	124160	PASS
96	95	5	9	6.3	7863	PASS
173	174	0.00	2	0.5	417	PASS
174	95	50	100	72.0	89400	PASS
175	174	5	9	7.6	6778	PASS
176	174	95	101	96.4	86195	PASS
177	176	5	9	6.8	5898	PASS