

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX080620\
 Data File : VX017799.D
 Acq On : 06 Aug 2020 11:22
 Operator : JC/SP
 Sample : L3570-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-01-200804

Quant Time: Aug 06 15:34:13 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 24 05:48:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	454560	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	681969	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	641283	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	307627	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	296220	50.73	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.46%	
35) Dibromofluoromethane	5.47	113	231157	50.68	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.36%	
50) Toluene-d8	8.70	98	813492	49.44	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.88%	
62) 4-Bromofluorobenzene	11.12	95	308282	46.82	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.64%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX080620\
Data File : VX017799.D
Acq On : 06 Aug 2020 11:22
Operator : JC/SP
Sample : L3570-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-01-200804

Quant Time: Aug 06 15:34:13 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 24 05:48:09 2020
Response via : Initial Calibration

