

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX080620\
 Data File : VX017804.D
 Acq On : 06 Aug 2020 13:24
 Operator : JC/SP
 Sample : L3570-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OUTFALL001-200804

Quant Time: Aug 06 15:42:53 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 24 05:48:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	483782	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	730856	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	690475	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	321559	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	322044	51.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.64%	
35) Dibromofluoromethane	5.47	113	247693	50.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.34%	
50) Toluene-d8	8.70	98	879823	49.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.80%	
62) 4-Bromofluorobenzene	11.12	95	320115	45.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.72%	
Target Compounds						
20) Methylene Chloride	2.84	84	9045	0.484	ug/l #	82
27) cis-1,2-Dichloroethene	4.57	96	18855	3.243	ug/l	87
44) Trichloroethene	7.19	130	2742	0.467	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX080620\
Data File : VX017804.D
Acq On : 06 Aug 2020 13:24
Operator : JC/SP
Sample : L3570-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OUTFALL001-200804

Quant Time: Aug 06 15:42:53 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 24 05:48:09 2020
Response via : Initial Calibration

