

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080621\
 Data File : VX023585.D
 Acq On : 06 Aug 2021 10:33
 Operator : JC/MD
 Sample : VX0806MBL01
 Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0806MBL01

Quant Time: Aug 07 04:41:10 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	157938	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	264507	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	236638	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	95575	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	106188	53.90	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 107.80%	
35) Dibromofluoromethane	5.391	113	85303	51.58	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.16%	
50) Toluene-d8	8.653	98	319543	52.59	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 105.18%	
62) 4-Bromofluorobenzene	11.085	95	107307	51.94	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 103.88%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080621\
Data File : VX023585.D
Acq On : 06 Aug 2021 10:33
Operator : JC/MD
Sample : VX0806MBL01
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0806MBL01

Quant Time: Aug 07 04:41:10 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

