

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080621\
 Data File : VX023590.D
 Acq On : 06 Aug 2021 12:47
 Operator : JC/MD
 Sample : M3298-04 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 1924

Quant Time: Aug 07 04:43:08 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	156336	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	268539	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	247731	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	101506	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	107613	55.18	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 110.36%	
35) Dibromofluoromethane	5.397	113	85910	51.17	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.34%	
50) Toluene-d8	8.653	98	328290	53.22	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 106.44%	
62) 4-Bromofluorobenzene	11.085	95	110620	52.74	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 105.48%	
Target Compounds						
						Qvalue
16) Acetone	2.386	43	2551	2.67	ug/l	90
20) Methylene Chloride	2.788	84	1998	1.07	ug/l	89
52) Toluene	8.720	92	40958	9.18	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080621\
Data File : VX023590.D
Acq On : 06 Aug 2021 12:47
Operator : JC/MD
Sample : M3298-04 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
1924

Quant Time: Aug 07 04:43:08 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

