

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080621\
 Data File : VX023593.D
 Acq On : 06 Aug 2021 13:58
 Operator : JC/MD
 Sample : M3298-08 100X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ANTIFREEZE

Quant Time: Aug 07 04:44:10 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	158197	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	267842	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	246596	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	104270	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	108383	54.92	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 109.84%			
35) Dibromofluoromethane	5.397	113	86259	51.51	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 103.02%			
50) Toluene-d8	8.653	98	324792	52.79	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 105.58%			
62) 4-Bromofluorobenzene	11.085	95	113146	54.08	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 108.16%			
Target Compounds						
16) Acetone	2.392	43	8059	8.33	ug/l	93
20) Methylene Chloride	2.800	84	2081	1.11	ug/l #	90
52) Toluene	8.720	92	3239	0.73	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080621\
Data File : VX023593.D
Acq On : 06 Aug 2021 13:58
Operator : JC/MD
Sample : M3298-08 100X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ANTIFREEZE

Quant Time: Aug 07 04:44:10 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

