

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080621\
 Data File : VX023594.D
 Acq On : 06 Aug 2021 14:21
 Operator : JC/MD
 Sample : M3316-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GROUND-WATER

Quant Time: Aug 07 04:44:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	161955	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	275146	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	249131	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	99332	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	109219	54.06	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	108.12%
35) Dibromofluoromethane	5.397	113	87119	50.64	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.28%
50) Toluene-d8	8.652	98	331115	52.39	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	104.78%
62) 4-Bromofluorobenzene	11.085	95	110264	51.30	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.60%
Target Compounds						
						Qvalue
16) Acetone	2.385	43	3249	3.28	ug/l #	84
20) Methylene Chloride	2.800	84	1100	0.57	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080621\
Data File : VX023594.D
Acq On : 06 Aug 2021 14:21
Operator : JC/MD
Sample : M3316-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GROUND-WATER

Quant Time: Aug 07 04:44:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

