

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080621\
 Data File : VX023595.D
 Acq On : 06 Aug 2021 14:45
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Aug 07 04:44:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	159048	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	266086	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	237215	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	88511	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	104489	52.67	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	105.34%	
35) Dibromofluoromethane	5.397	113	84685	50.90	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	101.80%	
50) Toluene-d8	8.653	98	318849	52.16	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	104.32%	
62) 4-Bromofluorobenzene	11.085	95	101218	48.70	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	97.40%	
Target Compounds						Qvalue
20) Methylene Chloride	2.800	84	3980	2.10	ug/l	# 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080621\
Data File : VX023595.D
Acq On : 06 Aug 2021 14:45
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Aug 07 04:44:45 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

