

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080621\
 Data File : VX023599.D
 Acq On : 06 Aug 2021 16:19
 Operator : JC/MD
 Sample : M3320-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SOIL-SAMPLE-M

Quant Time: Aug 07 04:45:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	149045	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	251343	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	236527	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	96919	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	100586	54.10	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 108.20%	
35) Dibromofluoromethane	5.397	113	82321	52.38	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.76%	
50) Toluene-d8	8.653	98	308189	53.38	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 106.76%	
62) 4-Bromofluorobenzene	11.085	95	104996	53.48	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 106.96%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080621\
Data File : VX023599.D
Acq On : 06 Aug 2021 16:19
Operator : JC/MD
Sample : M3320-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SOIL-SAMPLE-M

Quant Time: Aug 07 04:45:48 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

