

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080622\
 Data File : VX030494.D
 Acq On : 07 Aug 2022 03:39
 Operator : JC/MD
 Sample : VIBLK657
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VIBLK657

Quant Time: Aug 07 05:48:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML080422WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Aug 06 02:08:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.769	114	261665	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	228015	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	92653	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	70836	42.155	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	84.320%
7) Chloroethane-d5	1.666	69	57466	48.512	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	97.020%
11) 1,1-Dichloroethene-d2	2.306	63	104443	32.858	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	65.720%
21) 2-Butanone-d5	4.471	46	167511	107.589	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	107.590%
24) Chloroform-d	5.068	84	160092	46.282	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.560%
26) 1,2-Dichloroethane-d4	5.964	65	113039	48.128	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.260%
32) Benzene-d6	5.983	84	306270	46.892	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.780%
36) 1,2-Dichloropropane-d6	7.312	67	103738	48.954	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	97.900%
41) Toluene-d8	8.653	98	259442	42.141	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	84.280%
43) trans-1,3-Dichloroprop...	8.952	79	39945	37.891	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	75.780%
47) 2-Hexanone-d5	9.391	63	108173	100.148	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	100.150%
56) 1,1,2,2-Tetrachloroeth...	11.195	84	135133	48.543	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	97.080%
66) 1,2-Dichlorobenzene-d4	12.323	152	89153	49.609	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.220%
Target Compounds						
					Qvalue	
5) Vinyl chloride	1.374	62	2892	1.599	ug/L	# 61
20) cis-1,2-Dichloroethene	4.489	96	1700	0.891	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080622\
Data File : VX030494.D
Acq On : 07 Aug 2022 03:39
Operator : JC/MD
Sample : VIBLK657
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VIBLK657

Quant Time: Aug 07 05:48:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML080422WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Aug 06 02:08:42 2022
Response via : Initial Calibration

