

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080622\
 Data File : VX030479.D
 Acq On : 06 Aug 2022 21:24
 Operator : JC/MD
 Sample : VX0806WBL02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VBLK672

Quant Time: Aug 07 05:46:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML080422WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Aug 06 02:08:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.769	114	293345	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	253466	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	108608	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	74392	39.490	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	78.980%
7) Chloroethane-d5	1.666	69	62753	47.254	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	94.500%
11) 1,1-Dichloroethene-d2	2.306	63	113803	31.937	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	63.880%
21) 2-Butanone-d5	4.471	46	185489	106.270	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	106.270%
24) Chloroform-d	5.068	84	171097	44.121	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.240%
26) 1,2-Dichloroethane-d4	5.964	65	121648	46.200	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.400%
32) Benzene-d6	5.983	84	323866	44.607	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.220%
36) 1,2-Dichloropropane-d6	7.318	67	112932	47.942	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	95.880%
41) Toluene-d8	8.653	98	278336	40.670	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	81.340%
43) trans-1,3-Dichloroprop...	8.958	79	44552	38.018	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	76.040%
47) 2-Hexanone-d5	9.390	63	121395	101.104	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	101.100%
56) 1,1,2,2-Tetrachloroeth...	11.195	84	151566	48.979	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	97.960%
66) 1,2-Dichlorobenzene-d4	12.323	152	96584	45.849	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.700%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080622\
Data File : VX030479.D
Acq On : 06 Aug 2022 21:24
Operator : JC/MD
Sample : VX0806WBL02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VBLK672

Quant Time: Aug 07 05:46:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML080422WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Aug 06 02:08:42 2022
Response via : Initial Calibration

