

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080624\
 Data File : VX042872.D
 Acq On : 06 Aug 2024 15:33
 Operator : JC/MD
 Sample : P3446-02DL 4X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 002-35TH-AVE(JUNE)DL

Manual Integrations
 APPROVED

Reviewed By :John Carlone 08/07/2024
 Supervised By :Mahesh Dadoda 08/08/2024

Quant Time: Aug 07 05:00:27 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X080624W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Aug 07 04:53:59 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.897	128	42426	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.763	114	235044	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	208281	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.958	65	107231	27.371	ug/l	0.00
Spiked Amount 30.000	Range 91 - 110		Recovery =	91.233%		
60) 4-Bromofluorobenzene	11.079	95	94883	27.923	ug/l	0.00
Spiked Amount 30.000	Range 63 - 112		Recovery =	93.067%		
63) Toluene-d8	8.653	98	292884	31.131	ug/l	0.00
Spiked Amount 30.000	Range 91 - 112		Recovery =	103.767%		
Target Compounds						
					Qvalue	
15) Acetone	2.386	58	22899	41.866	ug/l	83
18) Methylene Chloride	2.794	84	3151	1.070	ug/l	94
30) 2-Butanone	4.586	43	14618m	6.553	ug/l	
58) 4-Methyl-2-Pentanone	8.580	43	3595	0.917	ug/l	97
62) Toluene	8.720	91	625568	61.322	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080624\
Data File : VX042872.D
Acq On : 06 Aug 2024 15:33
Operator : JC/MD
Sample : P3446-02DL 4X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
002-35TH-AVE(JUNE)DL

Quant Time: Aug 07 05:00:27 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X080624W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Aug 07 04:53:59 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : John Carlone 08/07/2024
Supervised By : Mahesh Dadoda 08/08/2024

