

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080625\
 Data File : VX047244.D
 Acq On : 06 Aug 2025 13:36
 Operator : JC/MD
 Sample : Q2763-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-2

Quant Time: Aug 07 01:17:22 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 22 03:59:51 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	313399	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	531979	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	476787	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	240511	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	193336	49.523	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.040%
35) Dibromofluoromethane	5.391	113	167100	50.874	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.740%
50) Toluene-d8	8.647	98	492543	46.304	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.600%
62) 4-Bromofluorobenzene	11.079	95	233771	50.997	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.000%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	94324	60.312	ug/l	100
18) Methyl Acetate	2.715	43	26780	6.444	ug/l	95
20) Methylene Chloride	2.800	84	10532	2.448	ug/l #	89
25) 2-Butanone	4.574	43	28931	12.705	ug/l	94
37) Ethyl Acetate	4.727	43	17544	3.401	ug/l	95
43) Isopropyl Acetate	6.348	43	12661	1.574	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080625\
Data File : VX047244.D
Acq On : 06 Aug 2025 13:36
Operator : JC/MD
Sample : Q2763-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-2

Quant Time: Aug 07 01:17:22 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 22 03:59:51 2025
Response via : Initial Calibration

