

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080625\
 Data File : VX047246.D
 Acq On : 06 Aug 2025 14:51
 Operator : JC/MD
 Sample : Q2780-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-3

Quant Time: Aug 07 01:18:28 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 22 03:59:51 2025
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 08/07/2025
 Supervised By : Mahesh Dadoda 08/07/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	274014	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	465715	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	407634	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	211138	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	169555	49.674	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 99.340%			
35) Dibromofluoromethane	5.391	113	144521	50.260	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 100.520%			
50) Toluene-d8	8.647	98	423681	45.498	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 91.000%#			
62) 4-Bromofluorobenzene	11.079	95	202423	50.441	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 100.880%			
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	61163	44.730	ug/l	98
18) Methyl Acetate	2.715	43	26924	7.409	ug/l	99
20) Methylene Chloride	2.800	84	21890	5.819	ug/l	98
25) 2-Butanone	4.587	43	12062m	6.058	ug/l	
37) Ethyl Acetate	4.733	43	17216m	3.812	ug/l	
43) Isopropyl Acetate	6.348	43	12231	1.737	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080625\
Data File : VX047246.D
Acq On : 06 Aug 2025 14:51
Operator : JC/MD
Sample : Q2780-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-3

Quant Time: Aug 07 01:18:28 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 22 03:59:51 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 08/07/2025
Supervised By :Mahesh Dadoda 08/07/2025

