

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX080718\
 Data File : VX003877.D
 Acq On : 07 Aug 2018 13:26
 Operator : JC/MD
 Sample : J4337-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE BLANK-WATERREF-4

Quant Time: Aug 08 09:32:07 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080618W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 07 07:45:48 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	317162	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	491541	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	472989	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	230063	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	218448	51.96	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.92%	
35) Dibromofluoromethane	5.51	113	184960	49.67	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.34%	
50) Toluene-d8	8.72	98	704357	48.64	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.28%	
62) 4-Bromofluorobenzene	11.14	95	248901	53.12	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.24%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX080718\
Data File : VX003877.D
Acq On : 07 Aug 2018 13:26
Operator : JC/MD
Sample : J4337-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGEBLANK-WATERREF-4

Quant Time: Aug 08 09:32:07 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X080618W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 07 07:45:48 2018
Response via : Initial Calibration

