

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX080718\
 Data File : VX003882.D
 Acq On : 07 Aug 2018 15:39
 Operator : JC/MD
 Sample : J4344-11DL 20X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FL-0530DL

Quant Time: Aug 08 02:02:40 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080618W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 07 07:45:48 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	296727	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	449642	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	428712	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	231529	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	195438	49.69	ug/l	0.00
Spiked Amount			Recovery	=	99.38%	
35) Dibromofluoromethane	5.50	113	162871	47.81	ug/l	0.00
Spiked Amount			Recovery	=	95.62%	
50) Toluene-d8	8.72	98	638904	48.23	ug/l	0.00
Spiked Amount			Recovery	=	96.46%	
62) 4-Bromofluorobenzene	11.14	95	233138	54.39	ug/l	0.00
Spiked Amount			Recovery	=	108.78%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
31) Cyclohexane	5.57	56	101976	17.99	ug/l	98
39) Methylcyclohexane	7.46	83	64893	10.93	ug/l	97
52) Toluene	8.79	92	4156	0.43	ug/l	99
67) Ethyl Benzene	10.26	91	724462	39.67	ug/l	98
68) m/p-Xylenes	10.36	106	515641	72.27	ug/l	95
69) o-Xylene	10.70	106	141149	20.96	ug/l	97
73) Isopropylbenzene	11.02	105	46652	2.72	ug/l	96
78) n-propylbenzene	11.36	91	94468	4.91	ug/l	97
80) 1,3,5-Trimethylbenzene	11.51	105	108691	7.78	ug/l	100
84) 1,2,4-Trimethylbenzene	11.81	105	439465	31.29	ug/l	100
85) sec-Butylbenzene	11.95	105	13381	0.83	ug/l	80
86) p-Isopropyltoluene	12.07	119	9705	0.67	ug/l	91
95) Naphthalene	13.83	128	121732	7.06	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX080718\
Data File : VX003882.D
Acq On : 07 Aug 2018 15:39
Operator : JC/MD
Sample : J4344-11DL 20X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FL-0530DL

Quant Time: Aug 08 02:02:40 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X080618W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 07 07:45:48 2018
Response via : Initial Calibration

