

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX080718\
 Data File : VX003890.D
 Acq On : 07 Aug 2018 19:39
 Operator : JC/MD
 Sample : J4344-10DL 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FL-0529DL

Manual Integrations
 APPROVED

sam
 8/21/2018 4:31:37 PM

Quant Time: Aug 08 09:37:56 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080618W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 07 07:45:48 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	341214	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	515239	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	477690	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	251318	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.07	65	219956	48.63	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.26%	
35) Dibromofluoromethane	5.51	113	184954	47.38	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.76%	
50) Toluene-d8	8.72	98	731426	48.19	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.38%	
62) 4-Bromofluorobenzene	11.14	95	265553	54.07	ug/l	0.00
Spiked Amount	50.000		Recovery	=	108.14%	

Target Compounds

						Qvalue
31) Cyclohexane	5.57	56	235614	36.14	ug/l	97
39) Methylcyclohexane	7.46	83	171141	25.15	ug/l	99
67) Ethyl Benzene	10.26	91	420448	20.66	ug/l	98
68) m/p-Xylenes	10.36	106	295712	37.19	ug/l	97
69) o-Xylene	10.70	106	14409	1.92	ug/l	93
73) Isopropylbenzene	11.02	105	135650	7.30	ug/l	96
78) n-propylbenzene	11.36	91	137197	6.57	ug/l	99
80) 1,3,5-Trimethylbenzene	11.51	105	321353	21.20	ug/l	98
84) 1,2,4-Trimethylbenzene	11.81	105	1016728	66.70	ug/l	100
85) sec-Butylbenzene	11.95	105	47604	2.70	ug/l	88
86) p-Isopropyltoluene	12.07	119	26082	1.65	ug/l	95
95) Naphthalene	13.83	128	241867	12.92	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX080718\
Data File : VX003890.D
Acq On : 07 Aug 2018 19:39
Operator : JC/MD
Sample : J4344-10DL 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FL-0529DL

Manual Integrations
APPROVED

sam
8/21/2018 4:31:37 PM

Quant Time: Aug 08 09:37:56 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X080618W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 07 07:45:48 2018
Response via : Initial Calibration

