

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX080718\
 Data File : VX003908.D
 Acq On : 08 Aug 2018 04:58
 Operator : JC/MD
 Sample : PB111813ZHE#04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB111813ZHE#04

Quant Time: Aug 08 08:21:04 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080618W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 07 07:45:48 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	313601	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	499767	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	487003	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	248479	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	219048	52.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.38%	
35) Dibromofluoromethane	5.51	113	185303	48.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.88%	
50) Toluene-d8	8.72	98	718165	48.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.56%	
62) 4-Bromofluorobenzene	11.14	95	246822	51.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.62%	
Target Compounds						
16) Acetone	2.45	43	18639	7.54	ug/l	98
18) Methyl Acetate	2.78	43	7681	1.43	ug/l	93
43) Isopropyl Acetate	6.47	43	199517	21.53	ug/l	99
74) N-amyl acetate	6.47	43	199517	20.87	ug/l #	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX080718\
Data File : VX003908.D
Acq On : 08 Aug 2018 04:58
Operator : JC/MD
Sample : PB111813ZHE#04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB111813ZHE#04

Quant Time: Aug 08 08:21:04 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X080618W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 07 07:45:48 2018
Response via : Initial Calibration

