

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX080718\
 Data File : VX003911.D
 Acq On : 08 Aug 2018 06:17
 Operator : JC/MD
 Sample : J4344-16
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FL-0539

Quant Time: Aug 08 08:36:10 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080618W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 07 07:45:48 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	349032	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	544263	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	510506	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	286447	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	233329	50.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.86%	
35) Dibromofluoromethane	5.51	113	199034	48.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.54%	
50) Toluene-d8	8.72	98	764885	47.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.40%	
62) 4-Bromofluorobenzene	11.14	95	277241	53.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.88%	
Target Compounds						
16) Acetone	2.45	43	9682	3.52	ug/l	98
68) m/p-Xylenes	10.36	106	4145	0.49	ug/l	92
84) 1,2,4-Trimethylbenzene	11.81	105	4323	0.25	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX080718\
Data File : VX003911.D
Acq On : 08 Aug 2018 06:17
Operator : JC/MD
Sample : J4344-16
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FL-0539

Quant Time: Aug 08 08:36:10 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X080618W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 07 07:45:48 2018
Response via : Initial Calibration

