

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX080718\
 Data File : VX003872.D
 Acq On : 07 Aug 2018 10:24
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 08 01:29:16 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080618W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 07 07:45:48 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	104	0.00
2 T	Dichlorodifluoromethane	0.507	0.524	-3.4	105	0.00
3 P	Chloromethane	0.689	0.640	7.1	100	0.00
4 C	Vinyl Chloride	0.688	0.667	3.1#	103	0.00
5 T	Bromomethane	0.438	0.387	11.6	102	0.00
6 T	Chloroethane	0.440	0.447	-1.6	101	0.00
7 T	Trichlorofluoromethane	0.908	0.894	1.5	105	0.00
8 T	Diethyl Ether	0.408	0.392	3.9	103	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.567	0.560	1.2	108	0.00
10 T	Methyl Iodide	0.706	0.712	-0.8	102	0.00
11 T	Tert butyl alcohol	0.159	0.141	11.3	93	0.00
12 CM	1,1-Dichloroethene	0.537	0.518	3.5#	104	0.00
13 T	Acrolein	0.088	0.075	14.8	96	0.00
14 T	Allyl chloride	1.035	1.012	2.2	102	0.00
15 T	Acrylonitrile	0.380	0.343	9.7	96	0.00
16 T	Acetone	0.394	0.378	4.1	104	0.00
17 T	Carbon Disulfide	1.637	1.492	8.9	104	0.00
18 T	Methyl Acetate	0.859	0.769	10.5	95	0.00
19 T	Methyl tert-butyl Ether	1.902	1.857	2.4	103	0.00
20 T	Methylene Chloride	0.956	0.599	37.3#	100	0.00
21 T	trans-1,2-Dichloroethene	0.609	0.570	6.4	103	0.00
22 T	Diisopropyl ether	1.902	1.833	3.6	104	0.00
23 T	Vinyl Acetate	1.582	1.555	1.7	102	0.00
24 P	1,1-Dichloroethane	1.130	1.079	4.5	102	0.00
25 T	2-Butanone	0.593	0.554	6.6	97	0.00
26 T	2,2-Dichloropropane	0.815	0.864	-6.0	114	0.00
27 T	cis-1,2-Dichloroethene	0.661	0.641	3.0	105	0.00
28 T	Bromochloromethane	0.483	0.477	1.2	103	0.00
29 T	Tetrahydrofuran	0.305	0.279	8.5	95	0.00
30 C	Chloroform	1.063	1.044	1.8#	104	0.00
31 T	Cyclohexane	0.955	0.961	-0.6	106	0.00
32 T	1,1,1-Trichloroethane	0.892	0.899	-0.8	104	0.00
33 S	1,2-Dichloroethane-d4	0.663	0.646	2.6	101	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	100	0.00
35 S	Dibromofluoromethane	0.379	0.382	-0.8	100	0.00
36 T	1,1-Dichloropropene	0.560	0.578	-3.2	107	0.00
37 T	Ethyl Acetate	0.616	0.596	3.2	98	0.00
38 T	Carbon Tetrachloride	0.533	0.561	-5.3	105	0.00
39 T	Methylcyclohexane	0.660	0.714	-8.2	110	0.00
40 TM	Benzene	1.682	1.725	-2.6	105	0.00
41 T	Methacrylonitrile	0.339	0.332	2.1	99	0.00
42 TM	1,2-Dichloroethane	0.569	0.581	-2.1	105	0.00
43 T	Isopropyl Acetate	0.927	0.938	-1.2	101	0.00
44 TM	Trichloroethene	0.460	0.465	-1.1	107	0.00
45 C	1,2-Dichloropropane	0.434	0.440	-1.4#	103	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX080718\
 Data File : VX003872.D
 Acq On : 07 Aug 2018 10:24
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 08 01:29:16 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080618W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 07 07:45:48 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.282	0.281	0.4	103	0.00
47 T	Bromodichloromethane	0.531	0.552	-4.0	103	0.00
48 T	Methyl methacrylate	0.471	0.487	-3.4	100	0.00
49 T	1,4-Dioxane	0.011	0.011	0.0	96	0.00
50 S	Toluene-d8	1.473	1.507	-2.3	101	0.00
51 T	4-Methyl-2-Pentanone	0.768	0.775	-0.9	98	0.00
52 CM	Toluene	1.063	1.105	-4.0#	104	0.00
53 T	t-1,3-Dichloropropene	0.600	0.655	-9.2	107	0.00
54 T	cis-1,3-Dichloropropene	0.647	0.705	-9.0	105	0.00
55 T	1,1,2-Trichloroethane	0.434	0.438	-0.9	102	0.00
56 T	Ethyl methacrylate	0.630	0.676	-7.3	103	0.00
57 T	1,3-Dichloropropane	0.727	0.741	-1.9	106	0.00
58 T	2-Chloroethyl Vinyl ether	0.310	0.340	-9.7	104	0.00
59 T	2-Hexanone	0.574	0.598	-4.2	101	0.00
60 T	Dibromochloromethane	0.405	0.445	-9.9	106	0.00
61 T	1,2-Dibromoethane	0.435	0.458	-5.3	106	0.00
62 S	4-Bromofluorobenzene	0.477	0.542	-13.6	118	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	103	0.00
64 T	Tetrachloroethene	0.492	0.488	0.8	108	0.00
65 PM	Chlorobenzene	1.301	1.315	-1.1	104	0.00
66 T	1,1,1,2-Tetrachloroethane	0.447	0.473	-5.8	105	0.00
67 C	Ethyl Benzene	2.130	2.293	-7.7#	109	0.00
68 T	m/p-Xylenes	0.832	0.911	-9.5	111	0.00
69 T	o-Xylene	0.786	0.870	-10.7	110	0.00
70 T	Styrene	1.286	1.485	-15.5	114	0.00
71 P	Bromoform	0.334	0.376	-12.6	110	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	110	0.00
73 T	Isopropylbenzene	3.699	3.901	-5.5	110	0.00
74 T	N-amyl acetate	1.923	1.742	9.4	101	0.00
75 P	1,1,2,2-Tetrachloroethane	1.268	1.179	7.0	110	0.00
76 T	1,2,3-Trichloropropane	1.104	1.095	0.8	114	0.00
77 T	Bromobenzene	0.987	0.959	2.8	115	0.00
78 T	n-propylbenzene	4.153	4.375	-5.3	117	0.00
79 T	2-Chlorotoluene	2.532	2.558	-1.0	117	0.00
80 T	1,3,5-Trimethylbenzene	3.016	3.228	-7.0	120	0.00
81 T	trans-1,4-Dichloro-2-butene	0.362	0.390	-7.7	114	0.00
82 T	4-Chlorotoluene	2.967	3.046	-2.7	118	0.00
83 T	tert-Butylbenzene	2.967	3.176	-7.0	117	0.00
84 T	1,2,4-Trimethylbenzene	3.033	3.358	-10.7	120	0.00
85 T	sec-Butylbenzene	3.502	3.791	-8.3	116	0.00
86 T	p-Isopropyltoluene	3.142	3.367	-7.2	115	0.00
87 T	1,3-Dichlorobenzene	1.824	1.864	-2.2	119	0.00
88 T	1,4-Dichlorobenzene	1.892	1.861	1.6	118	0.00
89 T	n-Butylbenzene	2.738	2.823	-3.1	112	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX080718\
Data File : VX003872.D
Acq On : 07 Aug 2018 10:24
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Aug 08 01:29:16 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080618W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 07 07:45:48 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.510	0.503	1.4	107	0.00
91 T	1,2-Dichlorobenzene	1.808	1.698	6.1	108	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.260	0.235	9.6	98	0.00
93 T	1,2,4-Trichlorobenzene	1.300	1.263	2.8	110	0.00
94 T	Hexachlorobutadiene	0.609	0.581	4.6	113	0.00
95 T	Naphthalene	3.724	3.705	0.5	103	0.00
96 T	1,2,3-Trichlorobenzene	1.298	1.257	3.2	108	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6