

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX080719\
 Data File : VX011358.D
 Acq On : 07 Aug 2019 13:08
 Operator : JC/SP
 Sample : PB122030TB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB122030TB

Quant Time: Aug 08 07:23:58 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 02 01:55:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	204444	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	298460	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	271445	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	123244	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	105350	51.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.34%	
35) Dibromofluoromethane	5.49	113	90609	46.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.00%	
50) Toluene-d8	8.71	98	355149	48.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.56%	
62) 4-Bromofluorobenzene	11.13	95	116602	44.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.80%	
Target Compounds						
16) Acetone	2.43	43	14529	14.826	ug/l	95
20) Methylene Chloride	2.85	84	14772	6.883	ug/l	97
43) Isopropyl Acetate	6.44	43	10307	2.466	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX080719\
Data File : VX011358.D
Acq On : 07 Aug 2019 13:08
Operator : JC/SP
Sample : PB122030TB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB122030TB

Quant Time: Aug 08 07:23:58 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 02 01:55:00 2019
Response via : Initial Calibration

