

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX080719\
 Data File : VX011364.D
 Acq On : 07 Aug 2019 15:33
 Operator : JC/SP
 Sample : K4190-04 50X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 30320

Quant Time: Aug 08 07:35:47 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 02 01:55:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	215534	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	313446	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	286603	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	150946	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	109977	50.67	ug/l	0.00
Spiked Amount			Recovery	=	101.34%	
35) Dibromofluoromethane	5.49	113	94772	46.31	ug/l	0.00
Spiked Amount			Recovery	=	92.62%	
50) Toluene-d8	8.71	98	371648	48.11	ug/l	0.00
Spiked Amount			Recovery	=	96.22%	
62) 4-Bromofluorobenzene	11.13	95	133186	48.29	ug/l	0.00
Spiked Amount			Recovery	=	96.58%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
17) Carbon Disulfide	2.56	76	11712	2.545	ug/l #	95
20) Methylene Chloride	2.85	84	142587	63.019	ug/l	97
31) Cyclohexane	5.57	56	26169	9.362	ug/l	96
39) Methylcyclohexane	7.46	83	59359	18.299	ug/l	98
40) Benzene	6.13	78	164878	20.190	ug/l	97
52) Toluene	8.78	92	165785	30.704	ug/l	99
67) Ethyl Benzene	10.25	91	44485	4.322	ug/l	100
68) m/p-Xylenes	10.35	106	104212	26.179	ug/l	99
69) o-Xylene	10.69	106	36660	9.373	ug/l	98
73) Isopropylbenzene	11.02	105	10732	1.040	ug/l	99
78) n-propylbenzene	11.36	91	19841	1.728	ug/l	97
80) 1,3,5-Trimethylbenzene	11.51	105	53898	6.237	ug/l	99
84) 1,2,4-Trimethylbenzene	11.80	105	113123	13.147	ug/l	99
85) sec-Butylbenzene	11.94	105	10543	1.035	ug/l	92
86) p-Isopropyltoluene	12.06	119	13670	1.459	ug/l #	78
89) n-Butylbenzene	12.38	91	18013	2.232	ug/l #	32
95) Naphthalene	13.83	128	50535	5.127	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX080719\
Data File : VX011364.D
Acq On : 07 Aug 2019 15:33
Operator : JC/SP
Sample : K4190-04 50X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
30320

Quant Time: Aug 08 07:35:47 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 02 01:55:00 2019
Response via : Initial Calibration

