

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\WX080720\
 Data File : VX017829.D
 Acq On : 07 Aug 2020 16:49
 Operator : JC/SP
 Sample : L3675-06
 Misc : 25.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GAYA2

Quant Time: Aug 08 01:39:53 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\SOMXTR080720WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Aug 08 00:10:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.83	114	292444	5.00	ug/L	0.00
28) Chlorobenzene-d5	10.10	117	284342	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	12.06	152	135760	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	90624	4.48	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	89.60%
7) Chloroethane-d5	1.69	69	79499	4.96	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	99.20%
11) 1,1-Dichloroethene-d2	2.34	63	140947	3.39	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	67.80%
20) 2-Butanone-d5	4.54	46	281617	52.47	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	104.94%
24) Chloroform-d	5.14	84	194216	4.73	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.60%
26) 1,2-Dichloroethane-d4	6.03	65	109783	4.87	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.40%
32) Benzene-d6	6.05	84	373959	4.79	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.80%
36) 1,2-Dichloropropane-d6	7.37	67	118490	4.91	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.20%
41) Toluene-d8	8.70	98	325486	4.38	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.60%
45) trans-1,3-Dichloropropene-	8.99	79	44845	4.29	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	85.80%
48) 2-Hexanone-d5	9.43	63	219713	50.59	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	101.18%
59) 1,1,2,2-Tetrachloroethane-	11.23	84	86582	4.72	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.40%
65) 1,2-Dichlorobenzene-d4	12.36	152	134400	5.14	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	102.80%

Target Compounds					Qvalue
13) Acetone	2.43	43	12001	3.061 ug/L	86
16) Methylene chloride	2.83	84	6051	0.195 ug/L #	81
25) Chloroform	5.18	83	29720	0.608 ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX080720\
Data File : VX017829.D
Acq On : 07 Aug 2020 16:49
Operator : JC/SP
Sample : L3675-06
Misc : 25.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GAYA2

Quant Time: Aug 08 01:39:53 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\SOMXTR080720WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Aug 08 00:10:20 2020
Response via : Initial Calibration

