

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080723\
 Data File : VX036836.D
 Acq On : 07 Aug 2023 14:52
 Operator : JC/MD
 Sample : 03899-02DL 40X
 Misc : 1.20g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 522DL

Quant Time: Aug 08 01:35:20 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	142101	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	255047	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	232447	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	106445	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	96633	45.959	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	91.920%		
35) Dibromofluoromethane	5.385	113	82394	47.000	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	94.000%		
50) Toluene-d8	8.647	98	304553	48.989	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	97.980%		
62) 4-Bromofluorobenzene	11.079	95	117602	50.140	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	100.280%		
Target Compounds					Qvalue	
16) Acetone	2.379	43	9221	10.569	ug/l	97
31) Cyclohexane	5.544	56	5580	1.917	ug/l #	1
39) Methylcyclohexane	7.385	83	4340	1.336	ug/l	88
52) Toluene	8.720	92	135258	28.395	ug/l	97
67) Ethyl Benzene	10.195	91	52863	5.648	ug/l	100
68) m/p-Xylenes	10.299	106	88390	25.309	ug/l	100
69) o-Xylene	10.640	106	33931	9.733	ug/l	99
73) Isopropylbenzene	10.963	105	8853	1.005	ug/l	99
78) n-propylbenzene	11.305	91	27040	2.654	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	32610	4.425	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	114240	15.500	ug/l	97
85) sec-Butylbenzene	11.890	105	13124	1.454	ug/l	85
86) p-Isopropyltoluene	12.012	119	7560	1.018	ug/l	96
89) n-Butylbenzene	12.329	91	26210	3.941	ug/l #	51

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080723\
Data File : VX036836.D
Acq On : 07 Aug 2023 14:52
Operator : JC/MD
Sample : 03899-02DL 40X
Misc : 1.20g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
522DL

Quant Time: Aug 08 01:35:20 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

