

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080723\
 Data File : VX036842.D
 Acq On : 07 Aug 2023 17:17
 Operator : JC/MD
 Sample : 03891-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-1

Quant Time: Aug 08 01:38:52 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : John Carlone 08/08/2023
 Supervised By : Mahesh Dadoda 08/08/2023

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	141187	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	258596	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	238979	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	109496	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	99420	47.591	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	95.180%		
35) Dibromofluoromethane	5.385	113	83804	47.149	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	94.300%		
50) Toluene-d8	8.653	98	308575	48.955	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	97.920%		
62) 4-Bromofluorobenzene	11.079	95	118763	49.940	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	99.880%		
Target Compounds						
16) Acetone	2.380	43	36680	42.312	ug/l #	88
20) Methylene Chloride	2.788	84	13796	7.156	ug/l #	94
25) 2-Butanone	4.562	43	17628	12.534	ug/l	91
37) Ethyl Acetate	4.727	43	7028	2.344	ug/l	98
43) Isopropyl Acetate	6.342	43	131443	25.465	ug/l #	94
51) 4-Methyl-2-Pentanone	8.574	43	37794m	13.231	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080723\
Data File : VX036842.D
Acq On : 07 Aug 2023 17:17
Operator : JC/MD
Sample : 03891-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-1

Quant Time: Aug 08 01:38:52 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : John Carlone 08/08/2023
Supervised By : Mahesh Dadoda 08/08/2023

