

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080723\
 Data File : VX036844.D
 Acq On : 07 Aug 2023 18:05
 Operator : JC/MD
 Sample : 03891-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-3

Quant Time: Aug 08 01:39:37 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 08/08/2023
 Supervised By : Mahesh Dadoda 08/08/2023

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	138878	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	255319	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	240501	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	111705	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	97407	47.402	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	94.800%		
35) Dibromofluoromethane	5.385	113	81922	46.681	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	93.360%		
50) Toluene-d8	8.647	98	307315	49.381	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	98.760%		
62) 4-Bromofluorobenzene	11.079	95	119150	50.746	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	101.500%		
Target Compounds						
					Qvalue	
16) Acetone	2.379	43	41042	48.131	ug/l	92
20) Methylene Chloride	2.794	84	11997	6.326	ug/l	91
25) 2-Butanone	4.568	43	16551	11.963	ug/l	99
37) Ethyl Acetate	4.720	43	4988	1.685	ug/l	100
43) Isopropyl Acetate	6.342	43	116909	22.940	ug/l	95
51) 4-Methyl-2-Pentanone	8.573	43	34675m	12.295	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080723\
Data File : VX036844.D
Acq On : 07 Aug 2023 18:05
Operator : JC/MD
Sample : 03891-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-3

Quant Time: Aug 08 01:39:37 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : John Carlone 08/08/2023
Supervised By : Mahesh Dadoda 08/08/2023

