

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080724\
 Data File : VX042887.D
 Acq On : 06 Aug 2024 23:46
 Operator : JC/MD
 Sample : VX0806MBS01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0807MBS01

Quant Time: Aug 07 06:26:40 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080724W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 07 02:55:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	255186	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	433694	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	380076	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	176204	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	177384	50.635	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 101.260%			
35) Dibromofluoromethane	5.385	113	140503	49.548	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 99.100%			
50) Toluene-d8	8.647	98	499607	49.629	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 99.260%			
62) 4-Bromofluorobenzene	11.079	95	177121	48.963	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 97.920%			
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.166	85	54680	19.879	ug/l	96
3) Chloromethane	1.300	50	56867	18.975	ug/l	97
4) Vinyl Chloride	1.374	62	54575	19.193	ug/l	100
5) Bromomethane	1.599	94	33058	19.209	ug/l	100
6) Chloroethane	1.666	64	30659	18.926	ug/l	96
7) Trichlorofluoromethane	1.874	101	83393	18.858	ug/l	98
8) Diethyl Ether	2.136	74	30730	18.893	ug/l	89
9) 1,1,2-Trichlorotrifluo...	2.319	101	52362	18.833	ug/l	93
10) Methyl Iodide	2.447	142	54246	18.869	ug/l	96
11) Tert butyl alcohol	2.995	59	67799	97.299	ug/l #	90
12) 1,1-Dichloroethene	2.306	96	52001	18.488	ug/l	89
13) Acrolein	2.239	56	71748	104.461	ug/l	97
14) Allyl chloride	2.660	41	82943	19.516	ug/l	94
15) Acrylonitrile	3.068	53	166942	93.867	ug/l	98
16) Acetone	2.392	43	140090	92.196	ug/l #	87
17) Carbon Disulfide	2.501	76	124457	18.183	ug/l	100
18) Methyl Acetate	2.709	43	87402	18.139	ug/l	98
19) Methyl tert-butyl Ether	3.123	73	176443	18.638	ug/l	97
20) Methylene Chloride	2.788	84	60813	18.641	ug/l	92
21) trans-1,2-Dichloroethene	3.087	96	52376	18.509	ug/l	98
22) Diisopropyl ether	3.763	45	181457	19.435	ug/l	92
23) Vinyl Acetate	3.727	43	1031353	97.329	ug/l	97
24) 1,1-Dichloroethane	3.611	63	102722	19.130	ug/l	98
25) 2-Butanone	4.568	43	229449	96.156	ug/l	97
26) 2,2-Dichloropropane	4.471	77	49805	15.087	ug/l	91
27) cis-1,2-Dichloroethene	4.489	96	65069	19.389	ug/l	87
28) Bromochloromethane	4.897	49	37754	18.324	ug/l	89
29) Tetrahydrofuran	5.013	42	150521	93.349	ug/l	95
30) Chloroform	5.092	83	108000	19.212	ug/l	99
31) Cyclohexane	5.464	56	83469	19.042	ug/l	97
32) 1,1,1-Trichloroethane	5.379	97	93510	18.896	ug/l	97
36) 1,1-Dichloropropene	5.690	75	67699	17.609	ug/l	95
37) Ethyl Acetate	4.727	43	90302	18.772	ug/l	98
38) Carbon Tetrachloride	5.672	117	78321	18.046	ug/l	97
39) Methylcyclohexane	7.379	83	84927	18.582	ug/l	100
40) Benzene	6.037	78	224736	18.834	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080724\
 Data File : VX042887.D
 Acq On : 06 Aug 2024 23:46
 Operator : JC/MD
 Sample : VX0806MBS01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0807MBS01

Quant Time: Aug 07 06:26:40 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080724W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 07 02:55:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.928	41	47344	19.109	ug/l	93
42) 1,2-Dichloroethane	6.086	62	82219	18.944	ug/l	99
43) Isopropyl Acetate	6.348	43	132922	18.684	ug/l	97
44) Trichloroethene	7.123	130	54839	18.291	ug/l	94
45) 1,2-Dichloropropane	7.434	63	54784	18.099	ug/l	96
46) Dibromomethane	7.586	93	40937	18.415	ug/l	91
47) Bromodichloromethane	7.824	83	76148	18.439	ug/l	98
48) Methyl methacrylate	7.696	41	63413	18.590	ug/l	91
49) 1,4-Dioxane	7.671	88	24107	318.352	ug/l #	97
51) 4-Methyl-2-Pentanone	8.574	43	462941	97.149	ug/l	99
52) Toluene	8.720	92	140421	19.401	ug/l	96
53) t-1,3-Dichloropropene	8.982	75	68363	17.908	ug/l	99
54) cis-1,3-Dichloropropene	8.366	75	77457	18.372	ug/l	90
55) 1,1,2-Trichloroethane	9.153	97	57049	19.389	ug/l	98
56) Ethyl methacrylate	9.116	69	87588	18.499	ug/l	93
57) 1,3-Dichloropropane	9.311	76	92251	19.139	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.244	63	223200	100.019	ug/l	100
59) 2-Hexanone	9.433	43	348539	94.392	ug/l	97
60) Dibromochloromethane	9.518	129	58915	18.227	ug/l	97
61) 1,2-Dibromoethane	9.610	107	59171	19.034	ug/l	98
64) Tetrachloroethene	9.275	164	52444	19.845	ug/l	93
65) Chlorobenzene	10.079	112	151907	18.834	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.165	131	53199	19.129	ug/l	100
67) Ethyl Benzene	10.195	91	260972	19.236	ug/l	99
68) m/p-Xylenes	10.299	106	198318	38.291	ug/l	99
69) o-Xylene	10.640	106	99399	19.159	ug/l	98
70) Styrene	10.652	104	162803	19.514	ug/l	97
71) Bromoform	10.799	173	41498	18.584	ug/l #	100
73) Isopropylbenzene	10.963	105	252325	19.807	ug/l	100
74) N-amyl acetate	10.841	43	112595	19.643	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.213	83	87939	18.843	ug/l	97
76) 1,2,3-Trichloropropane	11.238	75	90201	22.762	ug/l	94
77) Bromobenzene	11.201	156	62730	19.357	ug/l	89
78) n-propylbenzene	11.305	91	280649	19.833	ug/l	100
79) 2-Chlorotoluene	11.366	91	178337	19.663	ug/l	98
80) 1,3,5-Trimethylbenzene	11.451	105	210842	19.988	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	20671	16.961	ug/l	95
82) 4-Chlorotoluene	11.457	91	196365	19.262	ug/l	97
83) tert-Butylbenzene	11.713	119	210386	19.408	ug/l	98
84) 1,2,4-Trimethylbenzene	11.750	105	213099	20.067	ug/l	99
85) sec-Butylbenzene	11.890	105	257426	20.167	ug/l	100
86) p-Isopropyltoluene	12.006	119	211234	19.918	ug/l	98
87) 1,3-Dichlorobenzene	11.969	146	111575	19.025	ug/l	97
88) 1,4-Dichlorobenzene	12.042	146	110723	18.419	ug/l	97
89) n-Butylbenzene	12.335	91	169918	18.572	ug/l	97
90) Hexachloroethane	12.536	117	36554	18.843	ug/l	88
91) 1,2-Dichlorobenzene	12.335	146	113790	19.376	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	12.945	75	18861	18.393	ug/l	85
93) 1,2,4-Trichlorobenzene	13.585	180	65267	18.219	ug/l	95
94) Hexachlorobutadiene	13.725	225	29582	19.471	ug/l	97
95) Naphthalene	13.774	128	236274	19.060	ug/l	100
96) 1,2,3-Trichlorobenzene	13.963	180	69624	19.308	ug/l	98


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080724\
Data File : VX042887.D
Acq On : 06 Aug 2024 23:46
Operator : JC/MD
Sample : VX0806MBS01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0807MBS01

Quant Time: Aug 07 06:26:40 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080724W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 07 02:55:22 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX042887.D
Acq: 06 Aug 2024 23:46




Instrument :
MSVOA_X
ClientSampleId :
VX0807MBS01


Tgt Ion:168 Resp: 255186
Ion Ratio Lower Upper
168 100
99 57.7 56.6 85.0



#2
Dichlorodifluoromethane
Concen: 19.879 ug/l
RT: 1.166 min Scan# 13
Delta R.T. -0.000 min
Lab File: VX042887.D
Acq: 06 Aug 2024 23:46

Tgt Ion: 85 Resp: 54680
Ion Ratio Lower Upper
85 100
87 35.7 16.7 50.1





#3

Chloromethane

Concen: 18.975 ug/l

RT: 1.300 min Scan# 35

Delta R.T. -0.000 min

Lab File: VX042887.D

Acq: 06 Aug 2024 23:46

Instrument :

MSVOA_X

ClientSampleId :





VX0807MBS01

Tgt Ion: 50 Resp: 56867

Ion Ratio Lower Upper

50 100

52 33.6 25.5 38.3



#4

Vinyl Chloride

Concen: 19.193 ug/l

RT: 1.374 min Scan# 47

Delta R.T. -0.000 min

Lab File: VX042887.D




Acq: 06 Aug 2024 23:46


Tgt Ion: 62 Resp: 54575

Ion Ratio Lower Upper

62 100

64 31.9 25.4 38.0









#5
Bromomethane
Concen: 19.209 ug/l
RT: 1.599 min Scan# 84
Delta R.T. 0.006 min
Lab File: VX042887.D
Acq: 06 Aug 2024 23:46




Instrument :
MSVOA_X
ClientSampleId :
VX0807MBS01


Tgt Ion: 94 Resp: 33058
Ion Ratio Lower Upper
94 100
96 93.4 74.6 111.8



#6
Chloroethane
Concen: 18.926 ug/l
RT: 1.666 min Scan# 95
Delta R.T. 0.006 min
Lab File: VX042887.D
Acq: 06 Aug 2024 23:46

Tgt Ion: 64 Resp: 30659
Ion Ratio Lower Upper
64 100
66 33.4 25.1 37.7









#7
Trichlorofluoromethane
Concen: 18.858 ug/l
RT: 1.874 min Scan# 101
Delta R.T. 0.006 min
Lab File: VX042887.D
Acq: 06 Aug 2024 23:46




Instrument :
MSVOA_X
ClientSampleId :
VX0807MBS01


Tgt Ion:101 Resp: 83393
Ion Ratio Lower Upper
101 100
103 66.2 51.4 77.2



#8
Diethyl Ether
Concen: 18.893 ug/l
RT: 2.136 min Scan# 172
Delta R.T. -0.000 min
Lab File: VX042887.D
Acq: 06 Aug 2024 23:46

Tgt Ion: 74 Resp: 30730
Ion Ratio Lower Upper
74 100
45 95.3 53.3 159.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 18.833 ug/l
RT: 2.319 min Scan# 201
Delta R.T. 0.006 min
Lab File: VX042887.D
Acq: 06 Aug 2024 23:46




Instrument :
MSVOA_X
ClientSampleId :
VX0807MBS01


Tgt Ion:101 Resp: 52362
Ion Ratio Lower Upper
101 100
85 43.0 37.0 55.6
151 79.8 58.6 87.8



#10
Methyl Iodide
Concen: 18.869 ug/l
RT: 2.447 min Scan# 223
Delta R.T. 0.006 min
Lab File: VX042887.D
Acq: 06 Aug 2024 23:46

Tgt Ion:142 Resp: 54246
Ion Ratio Lower Upper
142 100
127 48.9 41.6 62.4
141 14.3 11.6 17.4





#11

Tert butyl alcohol

Concen: 97.299 ug/l

RT: 2.995 min Scan# 31

Delta R.T. -0.000 min

Lab File: VX042887.D

Acq: 06 Aug 2024 23:46

Instrument :

MSVOA_X

ClientSampleId :





VX0807MBS01

Tgt Ion: 59 Resp: 67799

Ion Ratio Lower Upper

59 100

57 10.9 6.0 9.0#



#12

1,1-Dichloroethene

Concen: 18.488 ug/l

RT: 2.306 min Scan# 200

Delta R.T. -0.000 min

Lab File: VX042887.D

Acq: 06 Aug 2024 23:46




Tgt Ion: 96 Resp: 52001


Ion Ratio Lower Upper

96 100

61 161.5 144.6 216.8

98 61.3 50.4 75.6









#13
Acrolein
Concen: 104.461 ug/l
RT: 2.239 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX042887.D
Acq: 06 Aug 2024 23:46




Instrument :
MSVOA_X
ClientSampleId :
VX0807MBS01


Tgt Ion: 56 Resp: 71748
Ion Ratio Lower Upper
56 100
55 68.8 57.1 85.7



#14
Allyl chloride
Concen: 19.516 ug/l
RT: 2.660 min Scan# 258
Delta R.T. -0.000 min
Lab File: VX042887.D
Acq: 06 Aug 2024 23:46

Tgt Ion: 41 Resp: 82943
Ion Ratio Lower Upper
41 100
39 65.6 55.8 83.6
76 35.9 24.3 36.5





#15

Acrylonitrile

Concen: 93.867 ug/l

RT: 3.068 min Scan# 31

Delta R.T. -0.000 min

Lab File: VX042887.D

Acq: 06 Aug 2024 23:46

Instrument :

MSVOA_X

ClientSampleId :

VX0807MBS01





Tgt Ion: 53 Resp: 166942

Ion Ratio Lower Upper

53 100

52 84.0 66.0 99.0

51 36.3 29.8 44.6



#16

Acetone

Concen: 92.196 ug/l

RT: 2.392 min Scan# 214

Delta R.T. -0.000 min

Lab File: VX042887.D




Acq: 06 Aug 2024 23:46


Tgt Ion: 43 Resp: 140090

Ion Ratio Lower Upper

43 100

58 34.1 21.9 32.9#





#17

Carbon Disulfide

Concen: 18.183 ug/l

RT: 2.501 min Scan# 21

Delta R.T. -0.000 min

Lab File: VX042887.D

Acq: 06 Aug 2024 23:46

Instrument :

MSVOA_X

ClientSampleId :





VX0807MBS01

Tgt Ion: 76 Resp: 124457

Ion Ratio Lower Upper

76 100

78 8.7 7.0 10.6



#18

Methyl Acetate

Concen: 18.139 ug/l

RT: 2.709 min Scan# 266

Delta R.T. -0.000 min

Lab File: VX042887.D




Acq: 06 Aug 2024 23:46


Tgt Ion: 43 Resp: 87402

Ion Ratio Lower Upper

43 100

74 23.3 17.8 26.8









#19
Methyl tert-butyl Ether
Concen: 18.638 ug/l
RT: 3.123 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX042887.D
Acq: 06 Aug 2024 23:46




Instrument :
MSVOA_X
ClientSampleId :
VX0807MBS01


Tgt Ion: 73 Resp: 176443
Ion Ratio Lower Upper
73 100
57 20.2 17.3 25.9



#20
Methylene Chloride
Concen: 18.641 ug/l
RT: 2.788 min Scan# 279
Delta R.T. 0.000 min
Lab File: VX042887.D
Acq: 06 Aug 2024 23:46

Tgt Ion: 84 Resp: 60813
Ion Ratio Lower Upper
84 100
49 119.4 103.6 155.4
51 37.2 33.8 50.6
86 66.4 49.4 74.0









#21
trans-1,2-Dichloroethene
Concen: 18.509 ug/l
RT: 3.087 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX042887.D
Acq: 06 Aug 2024 23:46




Instrument :
MSVOA_X
ClientSampleId :
VX0807MBS01


Tgt Ion: 96 Resp: 52376
Ion Ratio Lower Upper
96 100
61 148.0 119.2 178.8
98 65.7 48.7 73.1



#22
Diisopropyl ether
Concen: 19.435 ug/l
RT: 3.763 min Scan# 439
Delta R.T. -0.006 min
Lab File: VX042887.D
Acq: 06 Aug 2024 23:46

Tgt Ion: 45 Resp: 181457
Ion Ratio Lower Upper
45 100
43 96.8 69.7 104.5
87 27.4 20.8 31.2
59 11.9 9.0 13.4









#23
 Vinyl Acetate
 Concen: 97.329 ug/l
 RT: 3.727 min Scan# 41
 Delta R.T. -0.000 min
 Lab File: VX042887.D
 Acq: 06 Aug 2024 23:46




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0807MBS01


Tgt Ion: 43 Resp: 1031353
 Ion Ratio Lower Upper
 43 100
 86 10.8 7.8 11.8



#24
 1,1-Dichloroethane
 Concen: 19.130 ug/l
 RT: 3.611 min Scan# 414
 Delta R.T. 0.006 min
 Lab File: VX042887.D
 Acq: 06 Aug 2024 23:46

Tgt Ion: 63 Resp: 102722
 Ion Ratio Lower Upper
 63 100
 98 7.5 3.3 9.8
 100 3.9 2.1 6.5









#25
2-Butanone
Concen: 96.156 ug/l
RT: 4.568 min Scan# 571
Delta R.T. -0.000 min
Lab File: VX042887.D
Acq: 06 Aug 2024 23:46




Instrument :
MSVOA_X
ClientSampleId :
VX0807MBS01


Tgt Ion: 43 Resp: 229449
Ion Ratio Lower Upper
43 100
72 25.9 19.4 29.2



#26
2,2-Dichloropropane
Concen: 15.087 ug/l
RT: 4.471 min Scan# 555
Delta R.T. -0.006 min
Lab File: VX042887.D
Acq: 06 Aug 2024 23:46

Tgt Ion: 77 Resp: 49805
Ion Ratio Lower Upper
77 100
97 25.6 10.8 32.3





#27

cis-1,2-Dichloroethene

Concen: 19.389 ug/l

RT: 4.489 min Scan# 51

Delta R.T. -0.000 min

Lab File: VX042887.D

Acq: 06 Aug 2024 23:46

Instrument :

MSVOA_X

ClientSampleId :

VX0807MBS01





Tgt Ion: 96 Resp: 65069

Ion Ratio Lower Upper

96 100

61 134.2 0.0 312.2

98 62.0 0.0 129.0



#28

Bromochloromethane

Concen: 18.324 ug/l

RT: 4.897 min Scan# 625

Delta R.T. -0.006 min

Lab File: VX042887.D

Acq: 06 Aug 2024 23:46




Tgt Ion: 49 Resp: 37754


Ion Ratio Lower Upper

49 100

129 1.8 0.0 3.8

130 81.7 57.5 86.3





#29

Tetrahydrofuran

Concen: 93.349 ug/l

RT: 5.013 min Scan# 644

Delta R.T. -0.000 min

Lab File: VX042887.D

Acq: 06 Aug 2024 23:46

Instrument :

MSVOA_X

ClientSampleId :

VX0807MBS01





Tgt Ion: 42 Resp: 150521

Ion Ratio Lower Upper

42 100

72 45.8 34.2 51.4

71 43.0 32.2 48.2



#30

Chloroform

Concen: 19.212 ug/l

RT: 5.092 min Scan# 657

Delta R.T. -0.000 min

Lab File: VX042887.D




Acq: 06 Aug 2024 23:46


Tgt Ion: 83 Resp: 108000

Ion Ratio Lower Upper

83 100


85 63.5 50.3 75.5








#31
Cyclohexane
Concen: 19.042 ug/l
RT: 5.464 min Scan# 719
Delta R.T. -0.006 min
Lab File: VX042887.D
Acq: 06 Aug 2024 23:46

Instrument :
MSVOA_X
ClientSampleId :
VX0807MBS01







Tgt Ion:	56	Resp:	83469
Ion	Ratio	Lower	Upper
56	100		
69	27.9	23.4	35.2
84	85.0	65.9	98.9



#32
1,1,1-Trichloroethane
Concen: 18.896 ug/l
RT: 5.379 min Scan# 704
Delta R.T. -0.006 min
Lab File: VX042887.D
Acq: 06 Aug 2024 23:46

Tgt Ion:	97	Resp:	93510
Ion	Ratio	Lower	Upper
97	100		
99	65.0	51.1	76.7
61	42.6	36.4	54.6





#33

1,2-Dichloroethane-d4

Concen: 50.635 ug/l

RT: 5.958 min Scan# 798

Delta R.T. 0.006 min

Lab File: VX042887.D

Acq: 06 Aug 2024 23:46

Instrument :

MSVOA_X

ClientSampleId :





VX0807MBS01

Tgt Ion: 65 Resp: 177384

Ion Ratio Lower Upper

65 100

67 53.2 0.0 100.2



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.763 min Scan# 931

Delta R.T. -0.000 min

Lab File: VX042887.D

Acq: 06 Aug 2024 23:46




Tgt Ion: 114 Resp: 433694


Ion Ratio Lower Upper

114 100

63 20.0 0.0 45.0

88 15.7 0.0 36.6





#35

Dibromofluoromethane

Concen: 49.548 ug/l

RT: 5.385 min Scan# 705

Delta R.T. -0.000 min

Lab File: VX042887.D

Acq: 06 Aug 2024 23:46

Instrument :

MSVOA_X

ClientSampleId :

VX0807MBS01





Tgt Ion:113 Resp: 140503

Ion Ratio Lower Upper

113 100

111 103.5 82.9 124.3

192 19.3 13.3 19.9



#36

1,1-Dichloropropene

Concen: 17.609 ug/l

RT: 5.690 min Scan# 755

Delta R.T. 0.006 min

Lab File: VX042887.D

Acq: 06 Aug 2024 23:46




Tgt Ion: 75 Resp: 67699


Ion Ratio Lower Upper

75 100

110 37.5 17.0 50.9

77 32.5 24.7 37.1









#37
Ethyl Acetate
Concen: 18.772 ug/l
RT: 4.727 min Scan# 597
Delta R.T. 0.006 min
Lab File: VX042887.D
Acq: 06 Aug 2024 23:46




Instrument :
MSVOA_X
ClientSampleId :
VX0807MBS01


Tgt Ion:	43	Resp:	90302
Ion	Ratio	Lower	Upper
43	100		
61	14.1	10.6	16.0
70	10.0	8.3	12.5



#38
Carbon Tetrachloride
Concen: 18.046 ug/l
RT: 5.672 min Scan# 752
Delta R.T. -0.006 min
Lab File: VX042887.D
Acq: 06 Aug 2024 23:46

Tgt Ion:	117	Resp:	78321
Ion	Ratio	Lower	Upper
117	100		
119	94.1	78.2	117.2
121	29.1	24.4	36.6





#39

Methylcyclohexane

Concen: 18.582 ug/l

RT: 7.379 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX042887.D

Acq: 06 Aug 2024 23:46

Instrument :

MSVOA_X

ClientSampleId :

VX0807MBS01





Tgt Ion: 83 Resp: 84927

Ion Ratio Lower Upper

83 100

55 80.6 64.6 97.0

98 46.4 36.7 55.1



#40

Benzene

Concen: 18.834 ug/l

RT: 6.037 min Scan# 812

Delta R.T. -0.000 min

Lab File: VX042887.D




Acq: 06 Aug 2024 23:46


Tgt Ion: 78 Resp: 224736

Ion Ratio Lower Upper

78 100

77 23.4 19.3 28.9









#41
Methacrylonitrile
Concen: 19.109 ug/l
RT: 4.928 min Scan# 61
Delta R.T. -0.000 min
Lab File: VX042887.D
Acq: 06 Aug 2024 23:46




Instrument :
MSVOA_X
ClientSampleId :
VX0807MBS01


Tgt Ion:	41	Resp:	47344
Ion	Ratio	Lower	Upper
41	100		
39	52.1	45.2	67.8
67	73.5	52.6	78.8
52	30.3	22.7	34.1



#42
1,2-Dichloroethane
Concen: 18.944 ug/l
RT: 6.086 min Scan# 820
Delta R.T. -0.000 min
Lab File: VX042887.D
Acq: 06 Aug 2024 23:46

Tgt Ion:	62	Resp:	82219
Ion	Ratio	Lower	Upper
62	100		
98	8.2	0.0	16.0





#43

Isopropyl Acetate

Concen: 18.684 ug/l

RT: 6.348 min Scan# 8

Delta R.T. -0.000 min

Lab File: VX042887.D

Acq: 06 Aug 2024 23:46

Instrument :

MSVOA_X

ClientSampleId :

VX0807MBS01





Tgt Ion: 43 Resp: 132922

Ion Ratio Lower Upper

43 100

61 20.1 15.2 22.8

87 13.4 9.8 14.6



#44

Trichloroethene

Concen: 18.291 ug/l

RT: 7.123 min Scan# 990

Delta R.T. -0.000 min

Lab File: VX042887.D




Acq: 06 Aug 2024 23:46


Tgt Ion:130 Resp: 54839

Ion Ratio Lower Upper

130 100

95 96.1 0.0 204.4





#45

1,2-Dichloropropane

Concen: 18.099 ug/l

RT: 7.434 min Scan# 1041

Delta R.T. 0.006 min

Lab File: VX042887.D

Acq: 06 Aug 2024 23:46

Instrument :

MSVOA_X

ClientSampleId :





VX0807MBS01

Tgt Ion: 63 Resp: 54784

Ion Ratio Lower Upper

63 100

65 32.6 24.3 36.5



#46

Dibromomethane

Concen: 18.415 ug/l

RT: 7.586 min Scan# 1066

Delta R.T. 0.006 min

Lab File: VX042887.D

Acq: 06 Aug 2024 23:46




Tgt Ion: 93 Resp: 40937


Ion Ratio Lower Upper

93 100

95 82.5 66.6 100.0

174 101.4 69.4 104.0









#47
Bromodichloromethane
Concen: 18.439 ug/l
RT: 7.824 min Scan# 1105
Delta R.T. -0.000 min
Lab File: VX042887.D
Acq: 06 Aug 2024 23:46




Instrument :
MSVOA_X
ClientSampleId :
VX0807MBS01


Tgt Ion:	83	Resp:	76148
Ion	Ratio	Lower	Upper
83	100		
85	63.4	49.8	74.6
127	9.7	7.1	10.7



#48
Methyl methacrylate
Concen: 18.590 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. -0.000 min
Lab File: VX042887.D
Acq: 06 Aug 2024 23:46

Tgt Ion:	41	Resp:	63413
Ion	Ratio	Lower	Upper
41	100		
69	86.3	59.2	88.8
39	57.5	47.7	71.5





#49

1,4-Dioxane

Concen: 318.352 ug/l

RT: 7.671 min Scan# 1079

Delta R.T. 0.006 min

Lab File: VX042887.D

Acq: 06 Aug 2024 23:46

Instrument :

MSVOA_X

ClientSampleId :

VX0807MBS01





Tgt Ion: 88 Resp: 24107

Ion Ratio Lower Upper

88 100

43 37.2 0.0 0.0#

58 77.5 60.1 90.1



#50

Toluene-d8

Concen: 49.629 ug/l

RT: 8.647 min Scan# 1240

Delta R.T. -0.000 min

Lab File: VX042887.D




Acq: 06 Aug 2024 23:46


Tgt Ion: 98 Resp: 499607

Ion Ratio Lower Upper

98 100

100 63.2 53.1 79.7









#51
4-Methyl-2-Pentanone
Concen: 97.149 ug/l
RT: 8.574 min Scan# 1228
Delta R.T. -0.006 min
Lab File: VX042887.D
Acq: 06 Aug 2024 23:46




Instrument :
MSVOA_X
ClientSampleId :
VX0807MBS01


Tgt Ion: 43 Resp: 462941
Ion Ratio Lower Upper
43 100
58 38.4 30.4 45.6



#52
Toluene
Concen: 19.401 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. -0.000 min
Lab File: VX042887.D
Acq: 06 Aug 2024 23:46

Tgt Ion: 92 Resp: 140421
Ion Ratio Lower Upper
92 100
91 173.3 134.2 201.2





#53

t-1,3-Dichloropropene

Concen: 17.908 ug/l

RT: 8.982 min Scan# 1295

Delta R.T. -0.000 min

Lab File: VX042887.D

Acq: 06 Aug 2024 23:46

Instrument :

MSVOA_X

ClientSampleId :





VX0807MBS01

Tgt Ion: 75 Resp: 68363

Ion Ratio Lower Upper

75 100

77 31.6 25.7 38.5



#54

cis-1,3-Dichloropropene

Concen: 18.372 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX042887.D

Acq: 06 Aug 2024 23:46




Tgt Ion: 75 Resp: 77457


Ion Ratio Lower Upper

75 100

77 34.2 25.3 37.9

39 48.9 46.7 70.1









#55
1,1,2-Trichloroethane
Concen: 19.389 ug/l
RT: 9.153 min Scan# 1323
Delta R.T. -0.000 min
Lab File: VX042887.D
Acq: 06 Aug 2024 23:46




Instrument :
MSVOA_X
ClientSampleId :
VX0807MBS01


Tgt Ion:	97	Resp:	57049
Ion	Ratio	Lower	Upper
97	100		
83	83.7	65.8	98.8
85	53.6	41.6	62.4
99	61.7	47.8	71.8



#56
Ethyl methacrylate
Concen: 18.499 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. -0.000 min
Lab File: VX042887.D
Acq: 06 Aug 2024 23:46

Tgt Ion:	69	Resp:	87588
Ion	Ratio	Lower	Upper
69	100		
41	65.1	56.8	85.2
39	35.6	32.0	48.0





#57

1,3-Dichloropropane

Concen: 19.139 ug/l

RT: 9.311 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX042887.D

Acq: 06 Aug 2024 23:46

Instrument :

MSVOA_X

ClientSampleId :





VX0807MBS01

Tgt Ion: 76 Resp: 92251

Ion Ratio Lower Upper

76 100

78 32.5 26.1 39.1



#58

2-Chloroethyl Vinyl ether

Concen: 100.019 ug/l

RT: 8.244 min Scan# 1174

Delta R.T. -0.000 min

Lab File: VX042887.D




Acq: 06 Aug 2024 23:46


Tgt Ion: 63 Resp: 223200

Ion Ratio Lower Upper

63 100

106 28.1 22.5 33.7









#59
2-Hexanone
Concen: 94.392 ug/l
RT: 9.433 min Scan# 1369
Delta R.T. -0.000 min
Lab File: VX042887.D
Acq: 06 Aug 2024 23:46




Instrument :
MSVOA_X
ClientSampleId :
VX0807MBS01


Tgt Ion: 43 Resp: 348539
Ion Ratio Lower Upper
43 100
58 54.1 26.2 78.5



#60
Dibromochloromethane
Concen: 18.227 ug/l
RT: 9.518 min Scan# 1383
Delta R.T. -0.000 min
Lab File: VX042887.D
Acq: 06 Aug 2024 23:46

Tgt Ion: 129 Resp: 58915
Ion Ratio Lower Upper
129 100
127 80.1 38.8 116.4





#61

1,2-Dibromoethane

Concen: 19.034 ug/l

RT: 9.610 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX042887.D

Acq: 06 Aug 2024 23:46

Instrument :

MSVOA_X

ClientSampleId :





VX0807MBS01

Tgt Ion:107 Resp: 59171

Ion Ratio Lower Upper

107 100

109 96.2 75.2 112.8



#62

4-Bromofluorobenzene

Concen: 48.963 ug/l

RT: 11.079 min Scan# 1639

Delta R.T. -0.000 min

Lab File: VX042887.D

Acq: 06 Aug 2024 23:46




Tgt Ion: 95 Resp: 177121


Ion Ratio Lower Upper

95 100

174 77.9 0.0 135.6

176 76.0 0.0 131.0





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.055 min Scan# 1471

Delta R.T. -0.000 min

Lab File: VX042887.D

Acq: 06 Aug 2024 23:46

Instrument :

MSVOA_X

ClientSampleId :

VX0807MBS01





Tgt Ion:117 Resp: 380076

Ion Ratio Lower Upper

117 100

82 56.1 46.2 69.4

119 32.1 25.5 38.3



#64

Tetrachloroethene

Concen: 19.845 ug/l

RT: 9.275 min Scan# 1343

Delta R.T. -0.000 min

Lab File: VX042887.D

Acq: 06 Aug 2024 23:46

Tgt Ion:164 Resp: 52444




Ion Ratio Lower Upper


164 100

166 120.5 102.7 154.1

129 93.6 80.4 120.6

131 89.6 76.5 114.7









#65
Chlorobenzene
Concen: 18.834 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VX042887.D
Acq: 06 Aug 2024 23:46




Instrument :
MSVOA_X
ClientSampleId :
VX0807MBS01


Tgt Ion:112 Resp: 151907
Ion Ratio Lower Upper
112 100
114 32.5 25.7 38.5



#66
1,1,1,2-Tetrachloroethane
Concen: 19.129 ug/l
RT: 10.165 min Scan# 1489
Delta R.T. -0.000 min
Lab File: VX042887.D
Acq: 06 Aug 2024 23:46

Tgt Ion:131 Resp: 53199
Ion Ratio Lower Upper
131 100
133 93.6 47.1 141.4
119 64.3 32.2 96.6





#67

Ethyl Benzene

Concen: 19.236 ug/l

RT: 10.195 min Scan# 1494

Delta R.T. -0.000 min

Lab File: VX042887.D

Acq: 06 Aug 2024 23:46

Instrument :

MSVOA_X

ClientSampleId :





VX0807MBS01

Tgt Ion: 91 Resp: 260972

Ion Ratio Lower Upper

91 100

106 31.7 24.9 37.3



#68

m/p-Xylenes

Concen: 38.291 ug/l

RT: 10.299 min Scan# 1511

Delta R.T. -0.000 min

Lab File: VX042887.D




Acq: 06 Aug 2024 23:46


Tgt Ion: 106 Resp: 198318

Ion Ratio Lower Upper

106 100

91 205.6 165.5 248.3





#69

o-Xylene

Concen: 19.159 ug/l

RT: 10.640 min Scan# 1567

Delta R.T. -0.000 min

Lab File: VX042887.D

Acq: 06 Aug 2024 23:46

Instrument :

MSVOA_X

ClientSampleId :





VX0807MBS01

Tgt Ion:106 Resp: 99399

Ion Ratio Lower Upper

106 100

91 214.8 108.9 326.8



#70

Styrene

Concen: 19.514 ug/l

RT: 10.652 min Scan# 1569

Delta R.T. -0.000 min

Lab File: VX042887.D

Acq: 06 Aug 2024 23:46




Tgt Ion:104 Resp: 162803


Ion Ratio Lower Upper

104 100

78 51.3 43.6 65.4

103 56.1 43.6 65.4





#71

Bromoform

Concen: 18.584 ug/l

RT: 10.799 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX042887.D

Acq: 06 Aug 2024 23:46

Instrument :

MSVOA_X

ClientSampleId :

VX0807MBS01





Tgt Ion:173 Resp: 41498

Ion Ratio Lower Upper

173 100

175 48.5 24.1 72.4

254 0.0 0.0 0.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.024 min Scan# 1794

Delta R.T. 0.006 min

Lab File: VX042887.D

Acq: 06 Aug 2024 23:46




Tgt Ion:152 Resp: 176204


Ion Ratio Lower Upper

152 100

115 71.7 43.3 129.9

150 162.7 0.0 345.0









#73
Isopropylbenzene
Concen: 19.807 ug/l
RT: 10.963 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX042887.D
Acq: 06 Aug 2024 23:46




Instrument :
MSVOA_X
ClientSampleId :
VX0807MBS01


Tgt Ion:105 Resp: 252325
Ion Ratio Lower Upper
105 100
120 25.7 12.9 38.6



#74
N-ethyl acetate
Concen: 19.643 ug/l
RT: 10.841 min Scan# 1600
Delta R.T. -0.000 min
Lab File: VX042887.D
Acq: 06 Aug 2024 23:46

Tgt Ion: 43 Resp: 112595
Ion Ratio Lower Upper
43 100
70 42.4 32.9 49.3
55 24.3 19.3 28.9
61 24.3 18.8 28.2





#75

1,1,2,2-Tetrachloroethane

Concen: 18.843 ug/l

RT: 11.213 min Scan# 1661

Delta R.T. -0.000 min

Lab File: VX042887.D

Acq: 06 Aug 2024 23:46

Instrument :

MSVOA_X

ClientSampleId :

VX0807MBS01





Tgt Ion: 83 Resp: 87939

Ion Ratio Lower Upper

83 100

131 10.3 5.3 15.8

85 66.5 32.0 96.2



#76

1,2,3-Trichloropropane

Concen: 22.762 ug/l

RT: 11.238 min Scan# 1665

Delta R.T. -0.000 min

Lab File: VX042887.D




Acq: 06 Aug 2024 23:46


Tgt Ion: 75 Resp: 90201

Ion Ratio Lower Upper

75 100

77 34.4 19.1 57.5





#77

Bromobenzene

Concen: 19.357 ug/l

RT: 11.201 min Scan# 1659

Delta R.T. 0.006 min

Lab File: VX042887.D

Acq: 06 Aug 2024 23:46

Instrument :

MSVOA_X

ClientSampleId :

VX0807MBS01





Tgt Ion:156 Resp: 62730

Ion Ratio Lower Upper

156 100

77 154.3 88.4 265.3

158 98.3 48.0 144.0



#78

n-propylbenzene

Concen: 19.833 ug/l

RT: 11.305 min Scan# 1676

Delta R.T. -0.000 min

Lab File: VX042887.D




Acq: 06 Aug 2024 23:46


Tgt Ion: 91 Resp: 280649

Ion Ratio Lower Upper

91 100

120 23.1 11.5 34.4





#79

2-Chlorotoluene

Concen: 19.663 ug/l

RT: 11.366 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX042887.D

Acq: 06 Aug 2024 23:46

Instrument :

MSVOA_X

ClientSampleId :





VX0807MBS01

Tgt Ion: 91 Resp: 178337

Ion Ratio Lower Upper

91 100

126 33.6 16.3 48.8



#80

1,3,5-Trimethylbenzene

Concen: 19.988 ug/l

RT: 11.451 min Scan# 1700

Delta R.T. -0.000 min

Lab File: VX042887.D




Acq: 06 Aug 2024 23:46


Tgt Ion:105 Resp: 210842

Ion Ratio Lower Upper

105 100

120 48.0 24.1 72.3









#81
trans-1,4-Dichloro-2-butene
Concen: 16.961 ug/l
RT: 11.018 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX042887.D
Acq: 06 Aug 2024 23:46




Instrument :
MSVOA_X
ClientSampleId :
VX0807MBS01


Tgt Ion: 75 Resp: 20671
Ion Ratio Lower Upper
75 100
53 105.6 90.5 135.7
89 46.7 36.9 55.3



#82
4-Chlorotoluene
Concen: 19.262 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.006 min
Lab File: VX042887.D
Acq: 06 Aug 2024 23:46

Tgt Ion: 91 Resp: 196365
Ion Ratio Lower Upper
91 100
126 29.9 14.2 42.6





#83

tert-Butylbenzene

Concen: 19.408 ug/l

RT: 11.713 min Scan# 1743

Delta R.T. -0.000 min

Lab File: VX042887.D

Acq: 06 Aug 2024 23:46

Instrument :

MSVOA_X

ClientSampleId :

VX0807MBS01





Tgt Ion:119 Resp: 210386

Ion Ratio Lower Upper

119 100

91 60.7 31.2 93.6

134 22.7 11.7 35.0



#84

1,2,4-Trimethylbenzene

Concen: 20.067 ug/l

RT: 11.750 min Scan# 1749

Delta R.T. -0.000 min

Lab File: VX042887.D




Acq: 06 Aug 2024 23:46


Tgt Ion:105 Resp: 213099

Ion Ratio Lower Upper

105 100

120 42.7 21.2 63.6









#85
 sec-Butylbenzene
 Concen: 20.167 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. -0.000 min
 Lab File: VX042887.D
 Acq: 06 Aug 2024 23:46




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0807MBS01


Tgt Ion:105 Resp: 257426
 Ion Ratio Lower Upper
 105 100
 134 19.7 9.8 29.4



#86
 p-Isopropyltoluene
 Concen: 19.918 ug/l
 RT: 12.006 min Scan# 1791
 Delta R.T. -0.000 min
 Lab File: VX042887.D
 Acq: 06 Aug 2024 23:46

Tgt Ion:119 Resp: 211234
 Ion Ratio Lower Upper
 119 100
 134 26.2 12.8 38.3
 91 24.6 13.1 39.3









#87
1,3-Dichlorobenzene
Concen: 19.025 ug/l
RT: 11.969 min Scan# 1785
Delta R.T. -0.000 min
Lab File: VX042887.D
Acq: 06 Aug 2024 23:46




Instrument :
MSVOA_X
ClientSampleId :
VX0807MBS01


Tgt Ion:146 Resp: 111575
Ion Ratio Lower Upper
146 100
111 40.4 22.1 66.1
148 64.3 31.8 95.3



#88
1,4-Dichlorobenzene
Concen: 18.419 ug/l
RT: 12.042 min Scan# 1797
Delta R.T. -0.000 min
Lab File: VX042887.D
Acq: 06 Aug 2024 23:46

Tgt Ion:146 Resp: 110723
Ion Ratio Lower Upper
146 100
111 41.6 21.3 63.7
148 66.7 31.8 95.4





#89

n-Butylbenzene

Concen: 18.572 ug/l

RT: 12.335 min Scan# 1845

Delta R.T. 0.006 min

Lab File: VX042887.D

Acq: 06 Aug 2024 23:46

Instrument :

MSVOA_X

ClientSampleId :

VX0807MBS01





Tgt Ion: 91 Resp: 169918

Ion Ratio Lower Upper

91 100

92 51.8 27.4 82.0

134 25.7 12.6 37.8



#90

Hexachloroethane

Concen: 18.843 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. -0.000 min

Lab File: VX042887.D




Acq: 06 Aug 2024 23:46


Tgt Ion: 117 Resp: 36554

Ion Ratio Lower Upper

117 100

201 75.3 32.9 98.6









#91
1,2-Dichlorobenzene
Concen: 19.376 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. -0.000 min
Lab File: VX042887.D
Acq: 06 Aug 2024 23:46




Instrument :
MSVOA_X
ClientSampleId :
VX0807MBS01


Tgt Ion:146 Resp: 113790
Ion Ratio Lower Upper
146 100
111 42.0 22.4 67.0
148 63.9 31.6 95.0



#92
1,2-Dibromo-3-Chloropropane
Concen: 18.393 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. 0.006 min
Lab File: VX042887.D
Acq: 06 Aug 2024 23:46

Tgt Ion: 75 Resp: 18861
Ion Ratio Lower Upper
75 100
155 81.5 36.1 108.5
157 106.4 44.9 134.7





#93

1,2,4-Trichlorobenzene

Concen: 18.219 ug/l

RT: 13.585 min Scan# 2050

Delta R.T. -0.000 min

Lab File: VX042887.D

Acq: 06 Aug 2024 23:46

Instrument :

MSVOA_X

ClientSampleId :

VX0807MBS01





Tgt Ion:180 Resp: 65267

Ion Ratio Lower Upper

180 100

182 93.5 48.3 144.8

145 30.2 18.0 54.0



#94

Hexachlorobutadiene

Concen: 19.471 ug/l

RT: 13.725 min Scan# 2073

Delta R.T. -0.000 min

Lab File: VX042887.D

Acq: 06 Aug 2024 23:46




Tgt Ion:225 Resp: 29582


Ion Ratio Lower Upper

225 100

223 59.4 31.1 93.2

227 63.2 32.5 97.5





#95

Naphthalene

Concen: 19.060 ug/l

RT: 13.774 min Scan# 2081

Delta R.T. -0.000 min

Lab File: VX042887.D

Acq: 06 Aug 2024 23:46

Instrument :

MSVOA_X

ClientSampleId :

VX0807MBS01





Tgt Ion:128 Resp: 236274

Ion Ratio Lower Upper

128 100

127 12.9 10.3 15.5

129 10.6 8.8 13.2



#96

1,2,3-Trichlorobenzene

Concen: 19.308 ug/l

RT: 13.963 min Scan# 2112

Delta R.T. 0.006 min

Lab File: VX042887.D

Acq: 06 Aug 2024 23:46

Tgt Ion:180 Resp: 69624

Ion Ratio Lower Upper

180 100

182 93.9 46.9 140.6

145 31.7 17.8 53.5

